



March 10, 2023

Shanghai United Imaging Healthcare Co., Ltd.
% Xin GAO
Regulatory Affairs Manager
No. 2258 Chengbei Rd., Jiading Industrial District
Shanghai, Shanghai 201807
CHINA

Re: K223325/S001
Trade/Device Name: uMI Panorama
Regulation Number: 21 CFR 892.1200
Regulation Name: Emission computed tomography system
Regulatory Class: Class II
Product Code: KPS, JAK
Dated: February 17, 2023
Received: February 17, 2023

Dear Xin GAO:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

A handwritten signature in black ink, appearing to read 'D. Krainak', is written over a light blue watermark of the FDA logo.

Daniel M. Krainak, Ph.D.
Assistant Director
Magnetic Resonance and Nuclear Medicine Team
DHT8C: Division of Radiological Imaging
and Radiation Therapy Devices
OHT8: Office of Radiological Health
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K223325

Device Name
uMI Panorama

Indications for Use (Describe)

The uMI Panorama is a diagnostic imaging system that combines two existing imaging modalities PET and CT. The quantitative distribution information of PET radiopharmaceuticals within the patient body measured by PET can assist healthcare providers in assessing metabolic and physiological functions. CT provides diagnostic tomographic anatomical information as well as photon attenuation information for the scanned region. The accurate registration and fusion of PET and CT images provides anatomical reference for the findings in the PET images.

This system is intended to be operated by qualified healthcare professionals to assist in the detection, localization, diagnosis, staging, restaging, treatment planning and treatment response evaluation for diseases, inflammation, infection and disorders in, but not limited to oncology, cardiology and neurology. The system maintains independent functionality of the CT device, allowing for single modality CT diagnostic imaging.

This CT system can be used for low dose CT lung cancer screening for the early detection of lung nodules that may represent cancer. The screening must be performed within the established inclusion criteria of programs / protocols that have been approved and published by either a governmental body or professional medical society.

* Please refer to clinical literature, including the results of the National Lung Screening Trial (N Engl J Med 2011; 365:395-409) and subsequent literature, for further information.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510 (K) SUMMARY

K223325

1. Date of Preparation

October 29, 2022

2. Sponsor Identification

Shanghai United Imaging Healthcare Co.,Ltd.

No.2258 Chengbei Rd. Jiading District, 201807, Shanghai, China

Contact Person: Xin GAO

Position: Regulatory Affair Manager

Tel: +86-021-67076888-5386

Fax: +86-021-67076889

Email: xin.gao@united-imaging.com

3. Identification of Proposed Device

Device Name: uMI Panorama

Common Name: Positron Emission Tomography and Computed Tomography Systems

Model(s): uMI Panorama 28, uMI Panorama 35

Regulatory Information

Regulation Number: 21 CFR 892.1200

Regulation Name: Emission Computed Tomography System

Regulatory Class: II

Product Code: KPS, JAK

Review Panel: Radiology

4. Identification of Primary/Reference Device(s)

Predicate Device

510(k) Number: K172143

Device Name: uMI 780

Regulation Name: Emission Computed Tomography System

Regulatory Class: II

Product Code: KPS, JAK

Review Panel: Radiology

Reference Device#1

510(k) Number: K193241

Device Name: uMI 550

Regulation Name: Emission Computed Tomography System
Regulatory Class: II
Product Code: KPS, JAK
Review Panel: Radiology

Reference Device#2

510(k) Number: K210001
Device Name: HYPER AiR
Regulation Name: Emission Computed Tomography System
Regulatory Class: II
Product Code: KPS
Review Panel: Radiology

Reference Device#3

510(k) Number: K210418
Device Name: HYPER Focus
Regulation Name: Emission Computed Tomography System
Regulatory Class: II
Product Code: KPS
Review Panel: Radiology

5. Device Description:

The proposed device uMI Panorama combines a 280 or 350 mm axial field of view (FOV) PET and 160-slice CT system to provide high quality functional and anatomical images, fast PET/CT imaging and better patient experience. The system includes PET system, CT system, patient table, power distribution unit, control and reconstruction system (host, monitor, and reconstruction computer, system software, reconstruction software), vital signal module and other accessories.

The PET system features the following specification and technologies.

- 760mm patient bore size.
- Scalable LYSO detector configurations (96-ring and 120-ring) to have scalable Axial Field of Views (AFOV) of 280 and 350mm respectively, with corresponding imaging performances.
- 318 kg maximum table load capacity allows flexible positioning and access for all patients.
- uExcel Iterative (also named HYPER Iterative, has been cleared in K193241), uses a regularized iterative reconstruction algorithm, which allows for more iterations while keeping the image noise at an acceptable level by incorporating a noise penalty term into the objective function.
- uExcel DPR (also named HYPER AiR, has been cleared in K210001), involves pre-trained neural networks in the iteration reconstruction process to reduce noise and improve contrast of fluorodeoxyglucose (FDG) PET images.

- uExcel Focus (also named HYPER Focus, has been cleared in K210418), a respiratory motion correction feature can compromise respiratory motion effects and thus improve the measurement accuracy of SUV and lesion volume.

The CT system is UIH's 40mm detector uCT ATLAS Astound, which can also be used for standalone, diagnostic CT imaging.

The control and reconstruction system contains image acquisition and reconstruction, image display and post processing, data and patient management, CT dose display, networking, filming, etc.

6. Indications for Use

The uMI Panorama is a diagnostic imaging system that combines two existing imaging modalities PET and CT. The quantitative distribution information of PET radiopharmaceuticals within the patient body measured by PET can assist healthcare providers in assessing metabolic and physiological functions. CT provides diagnostic tomographic anatomical information as well as photon attenuation information for the scanned region. The accurate registration and fusion of PET and CT images provides anatomical reference for the findings in the PET images.

This system is intended to be operated by qualified healthcare professionals to assist in the detection, localization, diagnosis, staging, restaging, treatment planning and treatment response evaluation for diseases, inflammation, infection and disorders in, but not limited to oncology, cardiology and neurology. The system maintains independent functionality of the CT device, allowing for single modality CT diagnostic imaging.

This CT system can be used for low dose CT lung cancer screening for the early detection of lung nodules that may represent cancer. The screening must be performed within the established inclusion criteria of programs / protocols that have been approved and published by either a governmental body or professional medical society.

* Please refer to clinical literature, including the results of the National Lung Screening Trial (N Engl J Med 2011; 365:395-409) and subsequent literature, for further information.

7. Comparison of Technological Characteristics with the Predicate Device

uMI Panorama employs the same basic operating principles and fundamental technologies, and has the similar indications for use as the predicate device uMI 780. A comparison between the technological characteristics of proposed and predicate devices is provided as below.

| ITEM | Proposed Device uMI Panorama | Predicate Device uMI 780 (K172143) |
|--------------------------|---|--|
| Patient bore size | 760mm | 700mm |
| PET System | Scintillator material: LYSO Number of detector rings: • 96 (uMI Panorama 28) • 120 (uMI Panorama 35) Axial FOV: • 280mm (uMI Panorama 28) • 350mm (uMI Panorama 35) | Scintillator material: LYSO Number of detector rings: 112 Axial FOV: 300mm |
| CT System | uCT ATLAS Astound (K223028) | uCT 780 (K172135) |
| Maximum table load | 318kg | 250kg |
| Post-processing software | | |
| uExcel Iterative | Yes | No |
| uExcel DPR | Yes | No |
| uExcel Focus | Yes | No |

uMI Panorama’s technological characteristics do not raise new safety and effectiveness concerns.

8. Performance Data

The following performance data were provided in support of the substantial equivalence determination.

Non-Clinical Testing

Non-clinical testing including dosimetry and image performance tests were conducted for the uMI Panorama to verify that the proposed device met all design specifications as it is Substantially Equivalent (SE) to the predicate device.

UNITED IMAGING HEALTHCARE claims conformance to the following standards and guidance:

Electrical Safety and Electromagnetic Compatibility (EMC)

- ANSI/AAMI ES60601-1: 2005/ (R) 2012+A1:2012+C1:2009/(R)2012+A2:2010/(R)2012)[Including Amendment 2(2021)] Medical electrical equipment - Part 1: General requirements for basic safety and essential performance
- IEC60601-1-2:2014+A1:2020, Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic disturbances - Requirements and tests
- IEC 60601-1-3: 2008+AMD1:2013+A2:2021, Edition 2.2, Medical electrical equipment - Part 1-3: General requirements for basic safety and essential

- performance - Collateral Standard: Radiation protection in diagnostic X-ray equipment.
- IEC 60601-2-44:2009+A1:2012+A2:2016 Medical electrical equipment - Part 2-44: Particular requirements for the basic safety and essential performance of X-ray equipment for computed tomography
 - IEC 60825-1: 2014, Edition 3.0, Safety of laser products - Part 1: Equipment classification and requirements.
 - IEC 60601-1-6:2010+A1:2013+A2:2020, Edition 3.2, Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability.
 - IEC 62304:2006+AMD1:2015 CSV Consolidated version, Medical device software - Software life cycle processes
 - NEMA XR 25-2019, Computed Tomography Dose Check
 - NEMA XR 28-2018, Supplemental Requirements For User Information And System Function Related To Dose In CT
 - NEMA XR 29-2013, Standard Attributes on CT Equipment Related to Dose Optimization and Management
 - IEC 61223-3-5 2004 Edition 1.0, Evaluation And Routine Testing In Medical Imaging Departments - Part 3-5: Acceptance Tests - Imaging Performance Of Computed Tomography X-ray Equipment [Including: Technical Corrigendum 1 (2006)]
 - NEMA NU 2-2018, Performance Measurements of Positron Emission Tomographs

Software

- NEMA PS 3.1-3.20(2016): Digital Imaging and Communications in Medicine (DICOM)
- Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices
- Content of Premarket Submissions for Management of Cybersecurity in Medical Devices

Biocompatibility

- ISO 10993-5: 2009, Edition 3.0, Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity.
- ISO 10993-10: 2010, Edition 3.0, Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization.

Other Standards and Guidance

- ISO 14971: 2019, Edition 3.0, Medical Devices – Application of risk management to medical devices
- Code of Federal Regulations, Title 21, Part 820 - Quality System Regulation
- Code of Federal Regulations, Title 21, Subchapter J - Radiological Health
- Provision for Alternate Measure of the Computed Tomography Dose Index

(CTDI) to Assure Compliance with the Dose Information Requirements of the
Federal Performance Standard for Computed Tomography

Summary

The features described in this premarket submission are supported with the results of the testing mentioned above, the uMI Panorama was found to have a safety and effectiveness profile that is similar to the predicate device.

9. Conclusions

Based on the comparison and analysis above, the proposed device has similar intended use, performance, safety equivalence, and effectiveness as the predicate device. The differences above between the proposed device and predicate device do not affect the intended use, technology characteristics, safety, and effectiveness. And no issues are raised regarding to safety and effectiveness. The proposed device is determined to be Substantially Equivalent (SE) to the predicate device.