



January 5, 2023

Enztec Limited
% Nathan Wright, MS
Engineer & Regulatory Specialist
Empirical Technologies
4628 Northpark Drive
Colorado Springs, Colorado 80918

Re: K223671

Trade/Device Name: CONDUIT™ LLIF Straight Inserters
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: MAX
Dated: December 7, 2022
Received: December 7, 2022

Dear Mr. Wright:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Brent Showalter -S

Brent Showalter, Ph.D.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K223671

Device Name
CONDUIT™ LLIF Straight Inserters

Indications for Use (Describe)

The CONDUIT™ LLIF Straight Inserters are intended to be used with the EIT Cellular Titanium® LLIF Cages.

EIT Cellular Titanium® LLIF Cage

The EIT Cellular Titanium® LLIF Cages with a microscopic roughened surface and micro and nanoscale features are indicated for intervertebral body fusion procedures in skeletally mature patients with degenerative disc disease (DDD) at one or two contiguous levels from L2-S1. DDD is defined as discogenic back pain with degeneration of the disc confirmed by patient history and radiographic studies. These DDD patients may also have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level(s). These spinal implants are to be used with autogenous bone graft and/or allogeneic bone graft comprised of cancellous and/or corticocancellous bone graft. EIT Spine LLIF is to be used with supplemental fixation. Patients should have at least six (6) months of non-operative treatment prior to treatment with an intervertebral cage.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) SUMMARY

Submitter's Name:	Enztec Limited
Submitter's Address:	3/17 Print Place Middleton, Christchurch 8024
Submitter's Telephone:	+64 27 829 2440
Contact Person:	Nathan Wright MS Empirical Technologies 719-351-0248 nwright@empiricaltech.com
Date Summary was Prepared:	December 7, 2022
Trade or Proprietary Name:	CONDUIT™ LLIF Straight Inserters
Device Classification Name:	Implant insertion tool for intervertebral fusion device with bone graft, lumbar
Classification & Regulation #:	Class II per 21 CFR §888.3080
Product Code:	MAX
Classification Panel:	Spinal Devices (DHT6B)



DESCRIPTION OF THE DEVICE SUBJECT TO PREMARKET NOTIFICATION:

The CONDUIT™ LLIF Straight Inserters are designed for use during Lumbar Interbody Fusion surgery, specifically using a lateral approach. The instruments have been designed specifically to interface with the EIT Cellular Titanium® Cages (K201605). The reusable instruments are provided non-sterile and made from commonly used orthopedic materials with commonly used manufacturing processes. There are no changes to the implants or to other instruments provided with the implants.

INDICATIONS FOR USE

The CONDUIT™ LLIF Straight Inserters are intended to be used with the EIT Cellular Titanium® LLIF Cages.

EIT Cellular Titanium® LLIF Cage

The EIT Cellular Titanium® LLIF Cages with a microscopic roughened surface and micro and nano-scale features are indicated for intervertebral body fusion procedures in skeletally mature patients with degenerative disc disease (DDD) at one or two contiguous levels from L2-S 1. DDD is defined as discogenic back pain with degeneration of the disc confirmed by patient history and radiographic studies. These DDD patients may also have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level(s). These spinal implants are to be used with autogenous bone graft and/or allogeneic bone graft comprised of cancellous and/or corticocancellous bone graft. EIT Spine LLIF is to be used with supplemental fixation. Patients should have at least six (6) months of non-operative treatment prior to treatment with an intervertebral cage.

TECHNOLOGICAL CHARACTERISTICS

CONDUIT™ LLIF Straight Inserters are made from Stainless Steel per ASTM F899 and Radel in conformance with USP Class VI. The subject and predicate devices have nearly identical technological characteristics. Specifically, the following characteristics are identical between the subject and predicates:

- Indications for Use

- Materials of manufacture
- Sterility
- Compatibility

Primary Predicate Devices

510k Number	Trade or Proprietary or Model Name	Manufacturer	Predicate Type
K212823	DePuy CONDUIT™ Angled Inserters	DePuy Synthes	MAX

PERFORMANCE DATA

Non-clinical testing was conducted to confirm device performance per intended use including impact endurance test, and thread endurance test. The results of this non-clinical testing show that the performance of the CONDUIT™ LLIF Straight Inserters are sufficient for its intended use and is substantially equivalent to legally marketed predicate devices.

CONCLUSION

The overall technology characteristics and mechanical performance data lead to the conclusion that the CONDUIT™ LLIF Straight Inserters is substantially equivalent to the predicate device.
