



September 15, 2022

Smedtrum Medical Technology Co., Ltd.
Crimson Wu
Senior Regulatory Engineer
1F., No. 8, Ln. 97, Wugong Rd.,
New Taipei City, Xinzhuang Dist. 248016
Taiwan

Re: K221597

Trade/Device Name: CO2 Laser System

Regulation Number: 21 CFR 878.4810

Regulation Name: Laser Surgical Instrument For Use In General And Plastic Surgery And In
Dermatology

Regulatory Class: Class II

Product Code: GEX

Dated: August 3, 2022

Received: August 3, 2022

Dear Crimson Wu:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Jianting Wang
Acting Assistant Director
DHT4A: Division of General Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K221597

Device Name
CO2 Laser System

Indications for Use (Describe)

The CO2 Laser System is used for body soft tissue ablation, vaporization, excision and coagulation in dermatology, plastic surgery and general surgery.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Section 5 : 510(k) Summary CO₂ Laser System (K221597)

I. SUBMITTER

Smedtrum Medical Technology Co., Ltd.
1F., No. 8, Ln. 97, Wugong Rd.,
Xinzhuang Dist., New Taipei City 248016,
Taiwan (R.O.C.)

Contact Person

Crimson Wu

Position: Senior Regulatory Engineer

Tel: +886-2-2298-9578, Ext. 301

Fax: +886-2-2298-9426

E-mail: crimsonwu@smedtrum.com

Date of preparation: May 31th, 2022

II. DEVICE

Trade Name:	CO ₂ Laser System
Common or Usual Name:	Surgical Laser Device
Classification Name:	GEX–Powered Laser Surgical Instrument 21 C.F.R. § 878.4810, Device Class II

III. PREDICATE DEVICE

Trade Name:	CO ₂ Laser Therapy System
Common or Usual Name:	Surgical Laser Device
Classification Name:	GEX–Powered Laser Surgical Instrument 21 C.F.R. § 878.4810, Device Class II
Premarket Notification:	K201109 Sep 15 th , 2020

IV. DEVICE DESCRIPTION

CO₂ Laser System is a carbon dioxide surgical laser device intended for prescription use and is comprised mainly of three components: console, articulated arm and footswitch. CO₂ Laser System contains a radio-frequency excited CO₂ laser tube which generates laser sources at nominal wavelength of 10,600 nm. As CO₂ laser radiation is invisible; therefore, a low-power, visible aiming laser source (650 nm) is required to position the treatment laser beam. The output of laser beam is delivered through articulated arm to a fractional handpiece or normal handpiece. The control panel is in the form of an LCD touch screen in front of device and displays for operating and monitoring the laser. The physician activates laser emission by means of a footswitch. The fractional handpiece is suited mainly for general surgical application.

When the laser contacts human tissue, the laser energy is absorbed by water in the target tissue, resulting in a very rapid, highly localized temperature increase to the tissue. The instantaneous temperature-increase heats up the water and causes evaporation of the affected tissue. Some heat is absorbed by tissue adjacent to the treatment area, causing tissue coagulation which induces hemostasis as well as thermal stimulation of deep skin layers, that inducing fibroblast stimulation and neocollagenesis.

The CO₂ Laser System is capable of producing continuous-wave, pulsed laser or fractional laser radiation. The physician can optimize the parameter for different applications by adjusting the energy of laser pulse and output pattern.

V. INDICATIONS FOR USE

The CO₂ Laser System is used for body soft tissue ablation, vaporization, excision and coagulation in dermatology, plastic surgery and general surgery.

VI. COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH THE PREDICATE DEVICE

Feature	Proposed device	Predicate device (K201109)
<i>Device Name</i>	CO ₂ Laser System	CO ₂ Laser Therapy System
<i>Product Code</i>	GEX	GEX
<i>Regulation No.</i>	21 CFR 878.4810	21 CFR 878.4810
<i>Device Class</i>	Class II	Class II
<i>Indication for Use</i>	The CO ₂ Laser System is used for body soft tissue ablation, vaporization, excision and coagulation in dermatology, plastic surgery and general surgery.	The CO ₂ Laser Therapy System is used for body soft tissue vaporization and coagulation in dermatology and plastic surgery, general surgery, gynecology.
<i>Laser Type</i>	RF Sealed-off CO ₂	RF Sealed-off CO ₂
<i>Laser Classification</i>	Class 4	Class 4
<i>CO₂ Laser Wavelength</i>	10600 nm	10600 nm
<i>Aiming Beam Wavelength</i>	< 2 mw /650 nm /Semiconductor Laser LD	< 2 mw /650 nm /Semiconductor Laser LD
<i>Max. Aiming Beam Power Watts</i>	2 mW	2 mW
<i>Laser Delivery System</i>	Articulated arm with counterweight	Articulated arm with counterweight
<i>Beam Delivery Handpiece</i>	2 Normal Handpieces or 1 Fractional Handpiece	2 Normal Handpieces or 1 Fractional Handpiece
<i>Laser firing Controls</i>	LCD color Touchscreen Footswitch	LCD color Touchscreen Footswitch

Feature	Proposed device		Predicate device (K201109)	
<i>Max. Output Power Watts</i>	35 W		HS-411: 35 W HS-411A: 55 W	
<i>Scan Area Size</i>	2x2 mm~20x20 mm		2x2 mm~20x20 mm	
<i>Pulse Duration</i>	CW	NA	CW	NA
	Single	1~375 ms	Single	10~500ms
	Pulse	Pulse width: 5~375 ms Delay time: 1~500 ms	Pulse	On time: 5~500 ms Off time: 1-500 ms
	S-Pulse	Pulse width: 1~4 ms Delay time: 1~100 ms	S. Pulse	On time: 1~4 ms Off time: 1-100 ms
	U-pulse	Pulse width: 0.1-0.9 ms Delay time: 1-100 ms	U pulse	On time: 0.1-0.9 ms Off time: 1-100 ms
<i>Fractional Pulse Energy</i>	1-300 mJ/dot		1-300 mJ/dot	
<i>Spot Size</i>	120 μm (fractional)		150 μm (fractional)	
<i>Spot Density (DPA)/ cm²</i>	25, 36, 49, 64, 81, 100, 121, 144, 169, 196, 225, 256, 289, 324, 361, 400, 441, 484, 529, 784, 1024, 1521, 2025, 2500, 3025 dots.		25, 36, 49, 64, 81, 100, 121, 144, 169, 196, 225, 256, 289, 324, 361, 400, 441, 484, 529, 784, 1024, 1521, 2025, 2500, 3025 dots.	
<i>Operational Mode</i>	Fractional mode, and normal mode (CW, Single, Pulse, S. pulse, U. pulse)		Fractional mode, and normal mode (CW, Single, Pulse, S. pulse, U. pulse)	
<i>Cooling System</i>	Forced-air cooling		Forced-air cooling	
<i>Power Input</i>	100-240VAC, 50/60 Hz		100-240VAC, 50/60 Hz	
<i>Dimension</i>	188 cm x 61 cm x 53.4 cm (H x W x D)		113 cm x 45 cm x 50 cm (H x W x D)	
<i>Weight</i>	50 Kg		55 Kg	
<i>Software</i>	Yes. Verification and validation testing of the software are performed.		Yes	

VII. PERFORMANCE DATA

The CO₂ Laser System has been determined through engineering testing to verify laser energy output and electrical safety.

Electrical safety and electromagnetic compatibility

The test results demonstrated that the proposed device complies with the following standards:

IEC 60601-1: 2005+corr.1:2006+Corr.2.2007+A1:2012 Medical electrical equipment
- Part 1: General requirements for basic safety and essential performance

IEC 60601-1-2:2014 Medical electrical equipment -Part 1-2: General requirements for
basic safety and essential performance-Collateral Standard: Electromagnetic
disturbances - Requirements and tests

IEC 60825-1:2014 Safety of Laser products-Part 1: Equipment classification and
requirements

IEC 60601-2-22:2007(third edition)+A1:2012 for use in conjunction with IEC 60601-
1:2005 (third edition)+A1:2012 Medical electrical equipment - Part 2-22: Particular
requirements for the safety and essential performance of surgical, cosmetic,
therapeutic and diagnostic laser equipment

Software Verification and Validation Testing

Software verification and validation testing were conducted, and documentation was
provided as recommended by FDA's Guidance for Industry and FDA Staff, "Guidance
for the Content of Premarket Submissions for Software Contained in Medical Devices."
The software for this device was considered as a "moderate" level of concern since a
failure of the software could result in minor injury to a patient or to a user of the device.

Sterilization and Shelf-Life

The proposed device is not provided sterile and does not need to be sterilized. The
handpiece and the body are cleaned with a soft cloth moistened with isopropyl alcohol
or ethanol of 70% strength or higher. The proposed device is reusable and does not
have a restricted shelf-life.

VIII. CONCLUSION

The CO₂ Laser System has the same intended use, similar indications for use, the same
technological characteristics, the same energy used, and the same operating principles
as its predicates. The non-clinical data and performance testing reports in this
submission demonstrate that CO₂ Laser System meets the expected performance
requirements. Any difference between the subject and predicate device do not raise
new issues of safety or effectiveness. Based on above analysis, the CO₂ Laser System
is substantial equivalent to the cited predicate device.