

February 2, 2021

EcoGuard Inc. % Nickita Alexiades Consultant mdi Consultants, Inc. 55 Northern Blvd, Suite 200 Great Neck, New York 11021

Re: K202096

Trade/Device Name: 3-ply EcoGuard B with Earloop, 3-ply EcoGuard B with Tie-On Regulation Number: 21 CFR 878.4040 Regulation Name: Surgical Apparel Regulatory Class: Class II Product Code: FXX Dated: January 5, 2021 Received: January 13, 2021

Dear Nickita Alexiades:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <u>https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems</u>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<u>https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance</u>) and CDRH Learn (<u>https://www.fda.gov/training-and-continuing-education/cdrh-learn</u>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<u>https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice</u>) for more information or contact DICE by email (<u>DICE@fda.hhs.gov</u>) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For CAPT Elizabeth Claverie, M.S. Assistant Director DHT4B: Division of Infection Control and Plastic Surgery Devices OHT4: Office of Surgical and Infection Control Devices Office of Product Evaluation and Quality Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number *(if known)* K202096

Device Name

3-ply EcoGuard B with Earloop; 3-ply EcoGuard B with Tie-on

Indications for Use (Describe)

The Following EcoGuard Surgical Masks are intended to be worn to protect both the patient and healthcare personnel from the transfer of microorganisms, body fluids, and particulate material. These surgical masks are intended for use in infection control practices to reduce the potential exposure to blood and bodily fluid. These surgical masks are single use, disposable devices provided non-sterile.

3-ply EcoGuard B with Earloop, Model Number: ECO01 3-ply EcoGuard B with Tie-on, Model Number: ECO02

Type of Lise	(Select one	or hoth	as applicable)	
1900 01 030		01 00011,	us upplicusic)	

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer Paperwork Reduction Act (PRA) Staff *PRAStaff@fda.hhs.gov*

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY

The assigned 510(k) number is: K202096.

1. <u>Submitter's Identification:</u>

Manufacturer: EcoGuard Inc. Address: 700 S BATTLEGROUND AVE, SUITES 103 GROVER, NC 28073

Contact Person: Ms. Fang Wang Ecoguard Inc.

Date Summary Prepared: January 29, 2021

Official Correspondent: Mr. Nickita Alexiades mdi Consultants, Inc.

2. <u>Name of the Device:</u>

Device Name(s):	3-ply EcoGuard B with Earloop,	Model Number: ECO01
Model Number(s):	3-ply EcoGuard B with Tie-on,	Model Number: ECO02

Common Name: Surgical Mask Regulation Number: 878.4040 Regulation Name: Surgical Apparel Regulatory Class: II Product Code: FXX

3. Information for the 510(k) Cleared Device (Predicate Device):

Predicate Device: Disposable Surgical Face Mask K153496 Xiantao Rayxin Medical Products Co., Ltd.

4. <u>Device Description:</u>

The proposed surgical masks are available in two models with three-ply pleated earloop or tie-on (EcoGuard B) masks with ear loops and nose piece. Detail configurations of them are presented in Table I: Surgical Masks Description.

Product Model Number:	Product Model	ASTM Level
ECO01	3-Ply EcoGuard B with Earloop	Level 3
ECO02	3-Ply EcoGuard B with Tie-On	Level 3

Table 1: Surgical Masks Description

The inner and outer layers are made of spunbond polypropylene, and the middle layer is made of melt blown polypropylene filter. The ear loops or ties are held in place over the users' mouth and nose by two elastic ear loops or four ties welded to the surgical mask. Neither the elastic ear loops, nor ties are made with natural rubber latex. The nose piece in the layers of surgical mask is to allow the user to fit the surgical mask around their nose, which is made of malleable polyethylene wire. The surgical masks are sold nonsterile and are intended to be single use, disposable devices.

5. <u>Indications for Use:</u>

The following EcoGuard Surgical Masks are intended to be worn to protect both the patient and healthcare personnel from the transfer of microorganisms, body fluids, and particulate material. These surgical masks are intended for use in infection control practices to reduce the potential exposure to blood and bodily fluid. These surgical masks are single use, disposable devices provided non-sterile.

3-ply EcoGuard B with Earloop, Model Number: ECO01 3-ply EcoGuard B with Tie-on, Model Number: ECO02

6. <u>Comparison to the 510(k) Cleared Device (Predicate Device):</u>

Item	Subject Device	Predicate Device	Similar
nem	Subject Device	Fiedicale Device	
			Or Different
			Different
Manufacturer	EcoGuard Inc.	Xiantao Rayxin Medical	-
		Products Co., Ltd.	
510K number	K202096	K153496	-
Model Name	3-ply EcoGuard B	Disposable Surgical	Similar
	with Earloop	Face Mask	
	Model Number:		
	ECO01		
	3-ply EcoGuard B		
	with Tie-on		
	Model Number:		
	ECO02		
Classification	Class II Device	Class II Device	Similar
	Product Code: FXX	Product Code: FXX	
	(21 CFR878.4040)	(21 CFR878.4040)	
Indications for	The Following	The Disposable Surgical	Similar
Use	EcoGuard Surgical	Face Masks are	
	Masks are intended	intended to be worn to	
	to be worn to protect	protect both the patient	
	both the patient and	and healthcare	
	healthcare personnel	personnel from transfer	
	from transfer of	of microorganisms, body	
	microorganisms,	fluids and particulate	
	body fluids and	material. These face	
	particulate material.	masks are intended for	
	These surgical	use in infection control	
	masks are intended	practices to reduce the	
	for use in infection	potential exposure to	
	control practices to	blood and body fluids.	
	reduce the potential	This is a single use,	
	exposure to blood	disposable device(s),	
	and body fluids.	provided nonsterile.	
	-		
	These surgical		
	masks are single		
	use, disposable		

Table 2: Comparison to Predicate Device

	dovice provided		
	devices provided non-sterile.		
Model(s) &	EcoGuard B	Farloopa	Similar
. ,		Ear Loops Tie-On	Similar
Style	Ear Loops	Flat Pleated	
	Flat Pleated, 3 layers		
	ECO01	3 layers	
	Tio		
	Tie-On		
	Flat Pleated, 3 layers		
	ECO02		
Outer Facing	Spun-bond	Spun-bond	Similar
Layer	polypropylene	polypropylene	Cirinai
Layor			
Middle Layer	Melt blown	Melt blown	Similar
	polypropylene filter	polypropylene filter	
Inner Facing	Spun-bond	Spun-bond	Similar
Layer	polypropylene	polypropylene	
Nose Piece	Malleable	Malleable Aluminum	Different
	Polyethylene wire	wire	
Ear Loops or	Earloops:	Earloops:	Similar
Tie-On	Polyester	Polyester	
	Nylon/Spandex		
	Tie-On:		
	Polypropylene		
Color	Blue	Blue	Similar
Dimension	17.5cm±2cm	17.5cm±1cm	Similar
(Width)	6.89" ± 0.40"		
		0.5	
Dimension	9.5cm±2cm	9.5cm±1cm	Similar
(Length)	3.74" ± 0.40"	X	0. 1
OTC use	Yes	Yes	Similar
Sterility	Non-Sterile	Non-Sterile	Similar
Use	Single Use,	Single Use,	Similar
	Disposable	Disposable	Oinsiles
ASTM F2100	Level 3	Level 2	Similar
Level Non-Clinical	ASTM F1862	ASTM F1862	Similar
	ASTM F 1862 ASTM F2299	ASTM F 1862 ASTM F2299	Similar
Testing	ASTM F2299 ASTM F2101	ASTM F2299 ASTM F2101	
	ASTM F2101 ASTM F2100	ASTM F2101 ASTM F2100	
	MIL-M369454C 16 CFR 1610	MIL-M369454C 16 CFR 1610	
Biocompatibility	Cytotoxicity		Similar
	Irritation	Cytotoxicity Irritation	Similar
Testing			
1	Sensitization	Sensitization	

The difference in the materials does not raise additional questions for safety and effectiveness. Performance testing including biocompatibility evaluation has been performed on the final finished device

7. <u>Discussion of Non-Clinical Tests Performed for Determination of Substantial Equivalence are</u> <u>as follows:</u>

The EcoGuard Surgical Masks have been tested accordance with ASTM 2100-11 Standard Specification for Performance of Materials Used in Medical Face Masks. See Table 2 for a summary of other non-clinical testing.

Test	Purpose	Acceptance Criteria per ASTM	Subject Device Test Results- EcoGuard B	
		F2100-11 Level 3 (AQL = 4.0%)	ASTM F2100-11 Level 3	Average
ASTM F1862 Synthetic Blood	Determine synthetic blood penetration resistance	Level III	Pass at 160 mmHg (95/96) 32 Samples each from 3 non- consecutive lots	N/A
ASTM F2101 BFE	Determine the bacterial filtration efficiency	≥ 98%	Pass (96/96) 32 Samples each from 3 non- consecutive lots	99.9%
ASTM F2299 PFE at 0.1 micron	Determine submicron particulate filtration efficiency	≥ 99%	Pass (96/96) 32 Samples each from 3 non- consecutive lots	99.8%
Mil-M- 36954C Delta P	Determine breathing resistance or differential pressure	< 6.0 mm H2O/cm2	Pass (96/96) 32 Samples each from 3 non- consecutive lots	4.4 mm H20/cm2
16 CFR 1610 Flammability	Determine flammability or flame spread	Class 1	Class 1 Pass (96/96) 32 Samples each from 3 non- consecutive lots	N/A

All testing results have met ASTM F2100-11 2019 Level 3 acceptance criteria. The testing information demonstrates that the safety and effectiveness of the EcoGuard B Surgical Masks in their intended environment of use is supported by testing that was conducted in accordance with the FDA 2004 guidance "Surgical Masks - Premarket Notification [510(k)] Submissions", which outlines performance requirements for surgical masks.

(Standard) study, the device is noncytotoxic.	Table 5. Summary of Biocompatibility testing				
(Standard)study, the device is noncytotoxic.Irritation (Standard)Under the conditions of the study, the device isEcoGuard B: Pase	Test	Testing Findings	Result		
noncytotoxic.IrritationUnder the conditions of the study, the device isEcoGuard B: Pass	Cytotoxicity	Under the conditions of the	EcoGuard B: Pass		
IrritationUnder the conditions of the study, the device isEcoGuard B: Pass	(Standard)	study, the device is			
(Standard) study, the device is		noncytotoxic.			
	Irritation	Under the conditions of the	EcoGuard B: Pass		
nonirritating.	(Standard)	study, the device is			
		nonirritating.			
Sensitization Under the conditions of the EcoGuard B: Pass	Sensitization	Under the conditions of the	EcoGuard B: Pass		
(Standard) study, the device is	(Standard)	study, the device is			
nonsensitizing		nonsensitizing			

Table 3: Summary of Biocompatibility testing

None of the biocompatibility testing results demonstrated any safety hazards or any design characteristics that violated the requirements set forth in the Guidance for Industry regarding Premarket Notification Submissions of Surgical Masks. It was our conclusion that both 3-ply EcoGuard B with Earloop and 3-ply EcoGuard B with Tie-on surgical masks met all relevant testing requirements.

8. <u>Discussion of Clinical Tests Performed:</u>

No Clinical testing was performed.

9. <u>Conclusions:</u>

Based on the comparison and analysis above and the nonclinical tests performed, the subject device, EcoGuard B Surgical Masks, are as safe, as effective, and performs as well as the legally marketed predicate device, Xianto Rayxin Medical Products Disposable Surgical Mask cleared under K153496.