



July 12, 2021

Jiangxi Zhonghong Pulin Medical Products Co.,Ltd.
% Boyle Wang
Shanghai Truthful Information Technology Co., Ltd.
RM.608,No.738,Shangcheng Rd.,Pudong
Shanghai, Shanghai 200120
China

Re: K210723

Trade/Device Name: Nitrile Patient Examination Glove
Regulation Number: 21 CFR 880.6250
Regulation Name: Non-powdered patient examination glove
Regulatory Class: Class I, reserved
Product Code: LZA
Dated: June 15, 2021
Received: June 24, 2021

Dear Boyle Wang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Ryan Ortega
Acting Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K210723

Device Name

Nitrile Patient Examination Glove

Indications for Use (Describe)

The Nitrile Patient Examination Glove is a non-sterile disposable device intended for medical purposes that is worn on the examiner's hands or finger to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

(K210723)

This summary of 510(k) is being submitted in accordance with 21 CFR 807.92.

1.0 Submitter's Information

Name: Jiangxi Zhonghong Pulin Medical Products Co.,Ltd.
Address: Yinshawan Park,High-tech Industrial Park,Hukou County, Jiujiang
City,Jiangxi Province,China,332500
Phone Number: +86-19931401105
Contact: Zhou Ziyu
Date of Preparation: Jun.15,2021

Designated Submission Correspondent

Mr. Boyle Wang
Shanghai Truthful Information Technology Co., Ltd.
Room 608, No. 738 Shangcheng Rd., Pudong,Shanghai 200120 ,China
Tel: +86-21-50313932
Email: Info@truthful.com.cn

2.0 Device Information

Trade name: Nitrile Patient Examination Glove
Common name: Patient Examination Gloves
Classification name: Non-powdered patient examination glove
Model(s): XS,S, M, L, XL

3.0 Classification

Production code: LZA
Regulation number: 21CFR880.6250
Classification: Class I
Panel: General Hospital

4.0 Predicate Device Information

Manufacturer: Ever Global (Vietnam) Enterprise Corp
Device: Disposable Powder Free Nitrile Examination Glove, White/
Blue/ Black/ Pink Color
510(k) number: K171422

5.0 Indication for Use

The Nitrile Patient Examination Glove is a non-sterile disposable device intended for medical purposes that is worn on the examiner's hands or finger to prevent contamination between patient and examiner.

6.0 Device Description

The subject device is powder free nitrile examination gloves. The subject device is blue. The subject device is non-sterile.

7.0 Technological Characteristic Comparison Table

Table1-General Comparison

Item	Subject Device (K210723)	Predicated Device (K171422)	Remark
Product Code	LZA	LZA	Same
Regulation No.	21CFR880.6250	21CFR880.6250	Same
Class	I	I	Same
Intended Use	The Nitrile Patient Examination Glove is a non-sterile disposable device intended for medical purposes that is worn on the examiner's hands or finger to prevent contamination between patient and examiner.	The Nitrile Powder Free patient examination glove is a non-sterile disposable device intended for medical purposes that is worn on the examiner's hands or finger to prevent contamination between patient and examiner.	Same
Powdered or Powered free	Powdered free	Powdered free	Same
Design Feature	Ambidextrous	Ambidextrous	Same
Labeling Information	Single-use indication, powder free, device color, device name, glove size and quantity, Nitrile Glove Powder	Single-use indication, powder free, device color, device name, glove size and quantity, Disposable Powder Free Nitrile	Same

	Free Blue, Non-Sterile	Examination Glove, Non-Sterile	
--	------------------------	--------------------------------	--

Table2 Device Dimensions Comparison

Predicate Device(K171422)	Designation	Size					Tolerance	
		XS	S	M	L	XL		
	Length, mm	230	230	230	230	230	min	
	Width, mm	75	85	95	105	115	±5	
Thickness, mm:								
	Finger	0.05					min	
	Palm	0.05					min	
Subject Device (K210723)	Designation	Size					Tolerance	
		XS	S	M	L	XL		
		Length, mm	230	230	230	230	230	min
		Width, mm	70	80	95	110	120	±10
	Thickness, mm:							
		Finger	0.05					min
	Palm	0.05					min	
Remark	SIMILAR							

Analysis: The physical dimensions are little different with that of the predicate, but they all meet the requirements of ASTM D6319-19, so the differences do not raise any new safety or performance questions.

Table3 Performance Comparison

Item			Subject device (K210723)	Predicated device (K171422)	Remark
Colorant			Blue	White/ Blue/ Black/ Pink	Same
Physical Properties	Before Aging	Tensile Strength	14MPa, min	14MPa, min	Same
		Ultimate Elongation	500% min	500% min	Same
	After Aging	Tensile Strength	14MPa, min	14MPa, min	Same
		Ultimate Elongation	400%min	400%min	Same
	Comply with ASTM D6319			Comply with ASTM D6319	Same
Freedom from Holes			Be free from holes when tested in	Be free from holes when tested in	Same

	accordance with ASTMD5151 AQL=2.5	accordance with ASTMD5151 AQL=2.5	
Powder Content	Meet the requirements of ASTM D6124	Meet the requirements of ASTM D6124	Same

Table4 Safety Comparison

Item		Subject device (K210723)	Predicated device (K171422)	Remark
Material		Nitrile	Nitrile	Same
Biocompatibility	Irritation (ISO 10993-10:2010 Biological Evaluation of Medical Devices - Part 10: Tests For Irritation And Skin Sensitization)	Under the conditions of the study, not an irritant	Comply with ISO10993-10	Same
	Sensitization (ISO 10993-10:2010 Biological Evaluation of Medical Devices - Part 10: Tests For Irritation And Skin Sensitization)	Under conditions of the study, not a sensitizer.		
	Cytotoxicity (ISO 10993-5:2009 Biological Evaluation of Medical Devices - Part 5: Tests For In Vitro Cytotoxicity)	Under conditions of the study, device extract is not cytotoxic	/	Similar

8.0 Discussion of Non-clinical and Performance Testing

Non-clinical tests were conducted to verify that the proposed device met all design specifications. The test results demonstrated that the proposed device

complies with the following standards:

ISO 10993-10:2010 Biological Evaluation of Medical Devices - Part 10: Tests For Irritation And Skin Sensitization.

ISO 10993-5:2009 Biological Evaluation of Medical Devices - Part 5: Tests For In Vitro Cytotoxicity

ASTM D6124-06 (Reapproved 2017), Standard Test Method for Residual Powder on Medical Gloves

ASTM D5151-19, Standard Test Method for Detection of Holes in Medical Gloves.

ASTM D6319-19, Standard Specification for Nitrile Examination Gloves for Medical Application.

9.0 Discussion of Clinical and Performance Testing

Clinical testing is not needed for this device.

10.0 Conclusion

The conclusions drawn from the nonclinical tests demonstrate that the subject device is as safe, as effective, and performs as well as or better than the legally marketed predicated device.