



September 7, 2021

Bolder Surgical, LLC.
Nicholas Wong
Regulatory Affairs Manager
331 S. 104th Street, Suite 200
Louisville, Colorado 80027

Re: K211579

Trade/Device Name: CoolSeal Trinity (30 cm shaft), CoolSeal Trinity (37 cm shaft), CoolSeal Trinity (44 cm shaft)

Regulation Number: 21 CFR 878.4400

Regulation Name: Electrosurgical Cutting and Coagulation Device and Accessories

Regulatory Class: Class II

Product Code: GEI

Dated: August 6, 2021

Received: August 9, 2021

Dear Nicholas Wong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531 - 542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Long Chen, Ph.D.
Assistant Director
DHT4A: Division of General Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Indications for Use

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.

510(k) Number (if known)
K211579

Device Name
CoolSeal Trinity

Indications for Use (Describe)

The CoolSeal™ Trinity is a bipolar electrosurgical instrument intended for use in minimally invasive or open surgical procedures where ligation and division of vessels, tissue bundles, and lymphatics is desired. The CoolSeal™ Trinity can be used on vessels (arteries, veins, and vascular bundles) up to and including 7 mm in diameter. It is indicated for use in general surgery procedures including urologic, vascular, and gynecologic. It is indicated for use in adult and pediatric populations (infants, children, and adolescents). Procedures may include, but are not limited to, Nissen fundoplication, colectomy, cholecystectomy, adhesiolysis, hysterectomy, oophorectomy, etc. The CoolSeal™ Trinity has not been shown to be effective for tubal sterilization or tubal coagulation for sterilization procedures. Do not use the CoolSeal™ Trinity for these procedures. The device is contraindicated for use in ENT procedures.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary**SUBMITTER**

Bolder Surgical, LLC.
331 S. 104th Street, Suite 200
Louisville, CO 80027

Phone: 720-287-7130

Fax: 720-287-7135

Contact Person:

Nick Wong

Regulatory Affairs Manager

Date Prepared: May 19th, 2021

DEVICES

| | |
|-----------------------|--|
| Trade Name / Model #: | CoolSeal™ Trinity <ul style="list-style-type: none">• 30 cm / CSL-TR105-30• 37 cm / CSL-TR105-37• 44 cm / CSL-TR105-44 |
| Common or Usual Name: | Bipolar Vessel Sealing System |
| Classification Name: | Electrosurgical cutting and coagulation device and accessories |
| Regulatory Class: | II |
| Produce Code: | GEI |

PREDICATE DEVICES

Trade Name: CoolSeal™ Trinity
510(k): K203183

Trade Name: LigaSure™ Maryland
510(k) K170869

DEVICE DESCRIPTIONCoolSeal™ Trinity:

The CoolSeal™ Trinity, a Maryland Laparoscopic Sealer, Divider, and Dissector, with a 5 mm diameter shaft is designed for use with the CoolSeal™ Generator or any generator with the CoolSeal™ technology. The Trinity is provided sterile and is a single-use disposable instrument. The Trinity creates seals by application of radiofrequency (RF) electrosurgical energy to vascular structures (vessels and lymphatics) or tissue bundles interposed between its jaws. A blade within the instrument is surgeon-actuated to divide tissue. The double action jaws have been designed to dissect tissue, which includes separating tissue planes and widening openings as necessary for the surgical procedure. The CoolSeal™ Trinity includes 3 different shaft lengths: 30 cm, 37 cm, and 44 cm.

INDICATIONS FOR USECoolSeal™ Trinity:

The CoolSeal™ Trinity is a bipolar electrosurgical instrument intended for use in minimally invasive or open surgical procedures where ligation and division of vessels, tissue bundles, and lymphatics is desired. The CoolSeal™ Trinity can be used on vessels (arteries, veins, and vascular bundles) up to and including 7 mm in diameter. It is indicated for use in general surgery procedures including urologic, vascular, and gynecologic. It is indicated for use in adult and pediatric populations (infants, children, and adolescents). Procedures may include, but are not limited to, Nissen fundoplication, colectomy, cholecystectomy, adhesiolysis, hysterectomy, oophorectomy, etc. The CoolSeal™ Trinity has not been shown to be effective for tubal sterilization or tubal coagulation for sterilization procedures. Do not use the CoolSeal™ Trinity for these procedures. The device is contraindicated for use in ENT procedures.

COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH THE PREDICATE DEVICES

Table 1 below presents the comparison of technological characteristics between the subject device and the two identified predicate devices.

| Table 1. presents the subject Device compared to its two predicates, CoolSeal Trinity (K203183) and LigaSure Maryland (K170869) | | | |
|---|--|---|---|
| Description | CoolSeal™ Trinity (Subject Device) | CoolSeal™ Trinity (Primary Predicate/K203183) | LigaSure Maryland (Secondary Predicate/K170869) |
| Indications for Use | <p>The CoolSeal™ Trinity is a bipolar electro-surgical instrument intended for use in minimally invasive or open surgical procedures where ligation and division of vessels, tissue bundles, and lymphatics is desired. The CoolSeal™ Trinity can be used on vessels (arteries, veins, and vascular bundles) up to and including 7 mm in diameter. It is indicated for use in general surgery procedures including urologic, vascular, and gynecologic. It is indicated for use in adult and pediatric populations (infants, children, and adolescents). Procedures may include, but are not limited to, Nissen fundoplication, colectomy, cholecystectomy, adhesiolysis, hysterectomy, oophorectomy, etc.</p> <p>The CoolSeal™ Trinity has not been shown to be effective for tubal sterilization or tubal coagulation for CoolSeal™ Trinity for these procedures. The device is contraindicated for use in ENT procedures.</p> | <p>The CoolSeal™ Trinity is a bipolar electro-surgical instrument intended for use in minimally invasive or open surgical procedures where ligation and division of vessels, tissue bundles, and lymphatics is desired. The CoolSeal™ Trinity can be used on arteries up to and including 6 mm, veins, and vascular bundles up to and including 7 mm in diameter. It is indicated for use in general surgery procedures including urologic, vascular, and gynecologic. It is indicated for use in adult and pediatric populations (infants, children, and adolescents).</p> <p>The CoolSeal™ Trinity has not been shown to be effective for tubal sterilization or tubal coagulation for CoolSeal™ Trinity for these procedures. The device is contraindicated for use in ENT procedures.</p> | <p>The LigaSure Sealer/Divider is a bipolar electro-surgical instrument intended for use in minimally invasive or open surgical procedures where ligation and division of vessels, tissue bundles, and lymphatics is desired. The LigaSure Sealer/ Divider can be used on vessels (arteries, veins, and vascular bundles) up to and including 7 mm in diameter. It is indicated for use in general surgery procedures and such surgical specialties as urologic, vascular, thoracic, and gynecologic. Procedures may include, but are not limited to, Nissen fundoplication, colectomy, cholecystectomy, adhesiolysis, hysterectomy, oophorectomy, etc.</p> <p>The LigaSure system has not been shown to be effective for tubal sterilization or tubal coagulation for sterilization procedures. Do not use the LigaSure system for these procedures.</p> |
| Where used (environment) | Operating Room | | Operating Room |

| Description | CoolSeal™ Trinity (Subject Device) | CoolSeal™ Trinity (Primary Predicate/K203183) | LigaSure Maryland (Secondary Predicate/K170869) |
|--------------------------------------|--|---|--|
| Intended User | Surgeons | | |
| Anatomical Sites | Vessels, tissue bundles, and lymphatics | | |
| Anatomical Size | Arteries, veins, and vascular bundles up to and including 7 mm in diameter | Arteries up to and including 6 mm, veins, and vascular bundles up to and including 7 mm in diameter | Arteries, veins, and vascular bundles up to and including 7 mm in diameter |
| Patient Population | Adult and pediatric populations (infants, children, and adolescents). | | |
| Power Source | Bipolar energy platform | | |
| Primary Functions | Grasp, Dissect, Seal, Divide | | |
| Mechanism of Grasping | Hand actuated lever allows user to open or close | | |
| Mechanism of Dissection (Separation) | Bilateral jaw allows the user to separate planes of tissue | | |
| Mechanism of Action (Sealing) | Seal is created by application of RF energy to structures interposed between the jaws of the instrument. | | |
| Seal Activation | Button on the sealer instrument handle activated by thumb | | |
| Knife Activation | Cutting trigger – non-energized | | |
| Automatic sealing cycle | Yes | | |
| Rated Voltage (V _{peak}) | 190 V _{peak} | | |
| Shaft Diameter | 5 mm | | |
| Shaft Length | 30 cm, 37 cm, 44 cm | | |
| | | | Vessels, tissue bundles, and lymphatics |
| | | | Not specified |
| | | | Bipolar energy platform |
| | | | Grasp, Dissect, Seal, Divide |
| | | | Hand actuated lever allows user to open or close |
| | | | Unilateral jaw allows the user to separate planes of tissue |
| | | | Seal is created by the application of RF energy to structures interposed between the jaws of the instrument. |
| | | | Button on the sealer instrument handle activated by full compression of the lever or optional footswitch pedal |
| | | | Cutting trigger – non-energized |
| | | | Yes |
| | | | 288 V _{peak} |
| | | | 5 mm |
| | | | 23 cm, 37 cm, 44 cm |

| Description | CoolSeal™ Trinity (Subject Device) | CoolSeal™ Trinity (Primary Predicate/K203183) | LigaSure Maryland (Secondary Predicate/K170869) |
|------------------------------|---|---|---|
| Shaft Rotation | >360° | | 159°-359° |
| Seal Length | 19 mm | | 20 mm |
| How Supplied | Single-use disposable | Single-use disposable | Single-use disposable |
| Tissue Contact Materials | Stainless steel, titanium, polymers, silicone, polymer adhesives, and insulating coatings | Stainless steel, titanium, polymers, silicone, polymer adhesives, and insulating coatings | Stainless steel, titanium, polymers, silicone, polymer adhesives, and insulating coatings |
| Surgical Approach | Open or laparoscopic | Open or laparoscopic | Open or laparoscopic |
| Sterilization | Ethylene Oxide | Ethylene Oxide | Ethylene Oxide |
| Sterility Assurance Level | 10 ⁻⁶ | 10 ⁻⁶ | 10 ⁻⁶ |

PERFORMANCE DATA

The following performance data were provided in support of the substantial equivalence determination.

Biocompatibility TestingCoolSeal™ Trinity:

The biocompatibility evaluation for the CoolSeal™ Trinity was conducted in accordance with ISO 10993-1, Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process,” as recognized by FDA.

Electrical Safety and Electromagnetic Compatibility (EMC)

The CoolSeal™ system complies with relevant clauses of the ANSI/AAMI/IEC 60601-1 and ANSI/AAMI/IEC 60601-2-2 standards for electrical safety and ANSI/AAMI/IEC 60601-1-2 standard for EMC.

Mechanical and Functional Testing

Mechanical, electrical, and functional testing was carried out to verify that the proposed device performed as expected.

***Ex-vivo* and *In-vivo* Vessel Burst Pressure**CoolSeal™ Trinity:

Ex-vivo burst pressure testing of excised fresh porcine blood vessels was conducted on the subject device to demonstrate effective bipolar electrosurgical vessel sealing performance.

***In-vivo* Thermal Spread Comparison**CoolSeal™ Trinity:

Open laparotomy was performed in a porcine model using the subject device and predicate device. Each device was used to seal vessels and tissue bundles. Samples were excised for three-dimensional histological assessments (e.g., length, width, and depth) to quantify thermal spread in seals created by both devices. This study demonstrated that the subject device is as safe and effective as the predicate device.

Chronic Animal StudyCoolSeal™ Trinity:

Chronic animal studies were conducted to assess safety and performance of bipolar vessel sealing with the subject device over the minimum 21-day survival period. All animals survived without any complications. All vessel sealing effects on tissue maintained chronic hemostasis and healed as anticipated.

CONCLUSIONS

Based on a review of performance data, comparison of the device classification, intended use, operating principles, technological characteristics, sterility, and biocompatibility, the subject device is safe, as effective, and performs as well as the legally marketed predicate devices.