



March 25, 2020

Arterys Inc.
% Sharon Cholowsky
Director of Regulatory and Compliance
51 Federal Street, Suite 305
SAN FRANCISCO CA 94107

Re: K192437

Trade/Device Name: Arterys MICA
Regulation Number: 21 CFR 892.2050
Regulation Name: Picture archiving and communications system
Regulatory Class: Class II
Product Code: QIH, LLZ
Dated: February 15, 2020
Received: February 19, 2020

Dear Ms. Cholowsky:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see

<https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For

Thalia T. Mills, Ph.D.
Director
Division of Radiological Health
OHT7: Office of In Vitro Diagnostics
and Radiological Health
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K192437

Device Name
Arterys MICA

Indications for Use (Describe)

Arterys MICA software is a medical diagnostic application that displays, processes, stores, and transfers DICOM and non-DICOM medical data. It provides the capability to store images and patient information, and perform filtering, digital manipulation, and quantitative measurements. The client software is designed to run on standard personal and business computers and on monitors/screens that meet appropriate technical specifications for image diagnosis.

Arterys MICA includes an optional Cardio AI module which is used to analyze the heart and its major vessels using multi-slice, multi-phase, and velocity-encoded cardiovascular magnetic resonance (MR) images. It provides clinically relevant and reproducible, quantitative data, and has been tested and validated on MR images acquired from both 1.5T and 3.0 T MR Scanners.

Arterys MICA includes an optional Oncology AI module which provides analytical tools to help the user assess and document changes in morphological activity at diagnostic and therapy follow-up examinations. It is a tool used to support the oncological workflow by helping the user confirm the absence or presence of lesions, including evaluation, quantification, follow-up, and documentation of any such lesions.

Arterys MICA software is intended to be used as a support tool by trained healthcare professionals to aid in diagnosis. It is intended to provide image and related information that is interpreted by a trained professional to render findings and/or diagnosis, but it does not directly generate any diagnosis or potential findings.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

1. General Information

Date Prepared: September 4, 2019

Date Updated: March 19, 2020

Submitter Information

Company Name	Arterys Inc.
Company Address	51 Federal St., Suite 305 San Francisco, CA 94107
Contact Person	Sharon Cholowsky Director of Regulatory & Compliance
Contact Information	Email: regulatory@arterys.com Phone: 1 (650) 319-7230

Proposed Device

Proprietary Name	<i>Arterys[®] MICA</i>
Common Name	Medical image processing software
Model Number	AMM6
Classification Name	System, Image Processing, Radiological
Regulation Number	21 CFR 892.2050
Product Code	QIH/LLZ
Regulatory Class	II

Predicate & Reference Devices

Predicate Device	<i>Arterys[®] MICA, K182034</i> Product Code LLZ
Reference Device	<i>Medis MR-CT VVA, K140587</i> Product Code LLZ

2. Device Description

Arterys MICA, already cleared as per the predicate, is a dedicated software application used as a Digital Imaging and Communications in Medicine (DICOM) and non-DICOM information and data management system. Pre-existing DICOM images, such as CT or MR, are uploaded into

Arterys MICA from a PACS or a scanner. The software has two components: i) client, and ii) server. The client software (i) can be used in a Chrome desktop web browser. The server software (ii) runs on the Linux operating system.

The *Viewer* application of *Arterys MICA* is designed around a modular architecture of separate components that make up a basic image viewer. These components include the Worklist, from which studies are selected and opened, the Uploads list that displays all uploaded studies for the current organization, and the basic image display itself, which allows for viewing and working with 2D and 3D images.

Functionality provided by the *Viewer* is extended by the additional *Cardio AI* and *Oncology AI* (*Oncology AI: Lung AI* and *Oncology AI: Liver AI*) application modules which add support for specific clinical workflows:

- **Cardiac Workflow Module:** evaluates multi-slice and multi-phase velocity-encoded cardiovascular MR images to quantify blood flow and ventricular function.
- **Oncology Workflow Module:** supports the oncological workflow by helping the user confirm the absence or presence of lesions including evaluation, quantification, follow-up and documentation of any such lesions within MR or CT images.

NOTE: The clinician retains the ultimate responsibility for making the pertinent diagnosis based on their standard practices and visual comparison of images. *Arterys MICA* software is a complement to these standard procedures.

3. Indications for Use

Indications for Use Statement for *Arterys MICA* is as follows:

Arterys[®] *MICA* software is a medical diagnostic application that displays, processes, stores, and transfers DICOM and non-DICOM medical data. It provides the capability to store images and patient information, and perform filtering, digital manipulation, and quantitative measurements. The client software is designed to run on standard personal and business computers and on monitors/screens that meet appropriate technical specifications for image diagnosis.

Arterys MICA includes an optional *Cardio AI* module which is used to analyze the heart and its major vessels using multi-slice, multi-phase, and velocity-encoded cardiovascular magnetic resonance (MR) images. It provides clinically relevant and reproducible, quantitative data, and has been tested and validated on MR images acquired from both 1.5T and 3.0 T MR Scanners.

Arterys MICA includes an optional *Oncology AI* module which provides analytical tools to help the user assess and document changes in morphological activity at diagnostic and therapy follow-up examinations. It is a tool used to support the oncological workflow by helping the user confirm the absence or presence of lesions, including evaluation, quantification, follow-up, and documentation of any such lesions.

Arterys MICA software is intended to be used as a support tool by trained healthcare professionals to aid in diagnosis. It is intended to provide image and related information that is interpreted by a trained professional to render findings and/or diagnosis, but it does not directly generate any diagnosis or potential findings.

4. Predicate Device Comparison

The indications for use of the proposed device *Arterys MICA* is similar to the predicate device. They are intended to be used as a support tool by trained healthcare professionals to aid in diagnosis. The devices are intended to provide image and related information that is interpreted by a trained professional to render findings and/or diagnosis, but it does not directly generate any diagnosis or potential findings.

The indications for use of the proposed device consists of the indications for use statement of the predicate, *Arterys MICA* (K182034), with a phrase removed to allow for mammography diagnostic viewing and a phrase added regarding minimum technical specifications for monitors/screens. The differences in the indications for use don't raise different questions of safety and effectiveness.

The *Arterys MICA* software has the similar technological characteristics as the predicate device and has the same uses and applications as the predicate device. Differences include minor incremental updates to some of the existing software features for user experience improvements. These software changes do not impact safety or efficacy of the device.

5. Performance Data

Safety and performance of *Arterys MICA* has been evaluated and verified in accordance with software specifications and applicable performance standards through software verification and validation testing. The software verification and validation activities were performed as per *IEC 62304:2006/AC:2015- Medical device software – Software life cycle processes* and *ISO 14971:2007 Medical devices -- Application of risk management to medical devices*, in addition to the FDA Guidance documents *Guidance for the Content of Premarket Submissions for Software*

Contained in Medical Devices and Content of Premarket Submission for Management of Cybersecurity in Medical Devices.

Hundreds of software verification and validation tests, including the display quality of mammography images and a performance test comparison to *Medis MR-CT VVA*, were repeatedly conducted throughout the software development effort, with any defects logged and traced to the failed tests. Defects were then fixed or were assessed for approval in the released product.

6. Conclusion

Based on the information submitted in this premarket notification, and based on the indications for use, technological characteristics and performance testing, *Arterys MICA* raises no new or different questions of safety and effectiveness and is substantially equivalent to the predicate device in terms of safety, efficacy and performance.