



July 31, 2020

Safey Medical Devices Pvt Ltd  
Taher Moiyed  
CEO and Founder  
PAP-S-47&48, Chakan MIDC – II,  
Pune 410501, Maharashtra India

Re: K200832  
Trade/Device Name: Safey Peak Flow Meter  
Regulation Number: 21 CFR 868.1860  
Regulation Name: Peak-flow meter for spirometry  
Regulatory Class: Class II  
Product Code: BZH

Dear Taher Moiyed:

The Food and Drug Administration (FDA) is sending this letter to notify you of an administrative change related to your previous substantial equivalence (SE) determination letter dated July 30, 2020. Specifically, FDA is updating this SE Letter as an administrative correction (incorrect contact information).

Please note that the 510(k) submission was not re-reviewed. For questions regarding this letter please contact Rachana Visaria, OHT1: Office of Ophthalmic, Anesthesia, Respiratory, ENT and Dental Devices, (240) 402-5628, [Rachana.Visaria@fda.hhs.gov](mailto:Rachana.Visaria@fda.hhs.gov).

Sincerely,

**Rachana Visaria -S**

for Michael Ryan  
Director  
DHT1C: Division of ENT, Sleep Disordered  
Breathing, Respiratory and  
Anesthesia Devices  
OHT1: Office of Ophthalmic, Anesthesia,  
Respiratory, ENT and Dental Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health



July 30, 2020

Safey Medical Devices Pvt Ltd  
Taher Moiyed  
CEO and Founder  
3rd Floor, Office 303, Nyati Emporium, S no 105 H no 4A,  
Baner, Pune, Maharashtra 411045  
India

Re: K200832

Trade/Device Name: Safey Peak Flow Meter  
Regulation Number: 21 CFR 868.1860  
Regulation Name: Peak-flow meter for spirometry  
Regulatory Class: Class II  
Product Code: BZH  
Dated: June 30, 2020  
Received: June 30, 2020

Dear Taher Moiyed:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

  
Rachana Visaria -S

for Michael Ryan

Director

DHT1C: Division of ENT, Sleep Disordered  
Breathing, Respiratory and  
Anesthesia Devices

OHT1: Office of Ophthalmic, Anesthesia,  
Respiratory, ENT and Dental Devices

Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K200832

Device Name  
Safey Peak Flow Meter

Indications for Use (Describe)

Safey Peak Flow Meter is intended to measure Peak Expiratory Flow (PEF) and Forced Expiratory Volume in one second (FEV1) in home healthcare environment.

The device is designed for children greater than five years of age, adolescent and adult subjects.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

### 1) Applicant Information

<b>Manufacturer</b>	Safey Medical Devices Pvt Ltd PAP-S-47&48, Chakan MIDC – II, Pune 410501, Maharashtra India
<b>Contact Information</b>	Taher Ali Moiyed, CEO and Founder taher@safeymedicaldevices.com +44 7855 535353
<b>Date of preparation</b>	07/29/2020

### 2) Device Information

	<b>Subject Device</b>
<b>Proprietary name</b>	Safey Peak Flow Meter
<b>Common Name</b>	Peak Flow meter
<b>Classification</b>	Peak-flow meter for spirometry Regulatory Class: Class II; per 21 CFR 868.1860 Product Code: BZH
<b>Predicate Device</b>	Smart One (K181666)

### 3) Device Description

Safey Peak Flow Meter is an over-the-counter medical device to help respiratory patients keep track of their lung health. This device measures Peak Expiratory Flow (PEF) and Forced Expiratory Volume in one second (FEV1). Safey Peak Flow Meter is a pocket device intended for home use and operates on two AAA type standard alkaline batteries.

Safey Peak Flow Meter works on infrared interrupt concept. The turbine consists of a vane which rotates clockwise or anti-clockwise depending on the direction of flow into the turbine. The device consists of Infrared pairs which detects the direction and speed of rotation of the vane, which is further calculated to PEF and FEV1. The device connects with a Medical Mobile Application (Safey App) using BLE (Bluetooth Low Energy) to display the test results to the User.

#### **3.1 Mobile Medical Device (Safey App)**

Safey Peak Flow Meter connects to the smartphone with Safey App installed using Bluetooth (low energy). Upon conducting the tests as per instruction, the test result information is synched to the Safey App in real time. The App displays and stores this information for future use.

The Safey App is also a medication reminder tool. It helps users by reminding to take their medications on time. The Safey App also has the functionality which helps users to connect with friends and family members to share medication reminder related information.

#### 4) Indications for Use

Safey Peak Flow Meter is intended to measure Peak Expiratory Flow (PEF) and Forced Expiratory Volume in one second (FEV1) in home healthcare environment.

The device is designed for children greater than five years of age, adolescent and adult subjects.

#### 5) Comparison of Technological Characteristics

The following is a comparison between the predicate and the subject device.

	<b>Smart One</b>	<b>Safey Peak Flow Meter</b>	<b>Comparison</b>
510(k) number	K181666 (Predicate)	K200832	
Intended Use	Smart One is intended for home use by patients to monitor PEF (Peak Expiratory Flow) and FEV1 (Forced Expiratory Volume in one second). The device is designed for children greater than five years of age, adolescent and adult subjects.	Safey Peak Flow Meter is intended for home use by patients to monitor PEF (Peak Expiratory Flow) and FEV1 (Forced Expiratory Volume in one second). The device is designed for children greater than five years of age, adolescent and adult subjects.	Same
Type of Use	OTC	OTC	Same
Classification	BZH	BZH	Same
Size	109x49x21 mm	95x58x18.12 mm	Different dimensions
Operating Principal	Infrared Interrupts	Infrared Interrupts	Same
Method of communication	Bluetooth Low Energy	Bluetooth Low Energy	Same
Measured Values	PEF and FEV1	PEF and FEV1	Same
Volume Accuracy	3% or 0.1 L whichever is greater	3% or 0.1 L whichever is greater	Same
Peak Flow Accuracy	10% or 24 L/m (0.40 L/s) whichever is greater	10% or 24 L/m (0.40 L/s) whichever is greater	Same
Mouthpiece type	Reusable	Reusable	Same
Power source	2x Alkaline AAA Batteries	2x Alkaline AAA Batteries	Same
Flow and Volume Accuracy Standards	As per ATS/ERS Standards	As per ATS/ERS Standards	Same
IP Rating	IP22	IP22	Same
Maximum peak flow	16 L/s	16 L/s	Same
Maximum FEV1	10L	10L	Same
Test Feedback	Real time feedback and graphical representation of flow	Real time feedback and graphical representation of flow	Equivalent
Display screen	Mobile App	Mobile App	Same
Applicable standards	Electrical Safety IEC 60601-1	Electrical Safety IEC 60601-1	Equivalent

	Electromagnetic Compatibility IEC 60601–1–2 ATS Standardization of spirometry 1994 Update	Electromagnetic Compatibility IEC 60601–1–2 IEC 60601-1-6 IEC 60601-1-11 ATS Standardization of spirometry 2005	
--	--	--	--

In comparison to the predicate device, Safey Peak Flow Meter has some differences mainly in the shape, size and color of the devices. There are no differences in features, accuracy or any other factor which effects the safety and efficacy of the device.

## 6) Performance Data

A series of non-clinical tests are conducted on Safey Peak Flow Meter. They include: -

Standard	Description
IEC 60601-1:2005+AMD1:2012	General requirements for basic safety and essential performance
IEC 60601-1-2:2014(4th Edition)	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests
IEC 60601-1-6:2010, AMD1:2013	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability
IEC 60601-1-11:2015	General requirements for basic safety and essential performance -Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment
ATS Standardization of Spirometry	The Pocket Spirometer device was tested on a Flow/Volume Simulator according to American Thoracic Society (ATS) Document “Standardization of Spirometry -2005”. The results obtained show that Safey Peak Flow Meter and Safey Pocket Spirometer display results within ATS limits.
ISO 10993-1	Biocompatibility of the materials has been tested for cytotoxicity, irritation, and sensitization according to ISO 10993-1: 2009, following FDA’s guidance document “Use of International Standard ISO 10993-1, “Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process”.
IEC 62304	The evaluation of Software Development Life Cycle of the Software programmed into Safey Peak Flow Meter, Safey Pocket Spirometer and the Safey App was conducted as per IEC 62304. The software verification and validation was conducted and documented as per “Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices”
FCC Part 15 Subpart B and C	The device was tested as per 47 CFR Part 15 for intentional and unintentional radiators.



All performance tests conducted on Safey Peak Flow Meter passed.

Software verification and validation testing were conducted, and documentation was provided as recommended by FDA's Guidance, "Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices", and the software for this device was considered as a "moderate" level of concern.

In addition, the following FDA guidance documents were also followed in this submission:

- Guidance For Labeling Peak Flow Meters For Over The Counter Sale, Version 1.0, 1993
- Radio Frequency Wireless Technology in Medical Devices, Guidance for Industry and Food and Drug Administration Staff, August 4, 2013
- Design Considerations for Devices intended for home Use, Nov 24, 2014
- Postmarket management of Cybersecurity in Medical Devices, Dec 28, 2016

## **7) Conclusion**

Based on these results, it is our determination that the device is substantially equivalent to predicate device.