



May 7, 2021

Meditech Gloves SDN BHD
Wan Hassan
Assistant Manager - QA/RA
PT 3345, Jalan Permata 1/3, Arab Malaysian Industrial Park
Nilai, Negeri Sembilan 71800
Malaysia

Re: K210755

Trade/Device Name: Powder Free Nitrile Patient Examination Gloves, Non-Sterile (Blue)
Regulation Number: 21 CFR 880.6250
Regulation Name: Non-Powdered Patient Examination Glove
Regulatory Class: Class I, reserved
Product Code: LZA
Dated: March 5, 2021
Received: March 15, 2021

Dear Wan Hassan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Clarence W. Murray, III, Ph D.
Acting Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K210755

Device Name
Powder Free Nitrile Patient Examination Gloves, Non-Sterile (Blue)

Indications for Use (Describe)

A patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASTaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY

1.0 Submitter:

Name: Dr Effendi Tenang
Address: Meditech Gloves Sdn Bhd
PT3345, Jalan Permata 1/3, Arab Malaysian Industrial Park,
71800 Nilai, Negeri Sembilan, MALAYSIA
Phone No.: +60 6 799 7742
Fax No.: +60 6 799 7749

Date of Summary Prepared: May 4, 2021

2.0 Name of the device:

Powder Free Nitrile Patient Examination Gloves, Non-Sterile (Blue)
Common Name: Examination Gloves
Classification Name: Patient Examination Gloves (21 CFR 880.6250 product code LZA)

3.0 Identification of The Legally Marketed Devices that equivalency is claimed:

Dermagrip Powder Free Blue Nitrile Examination Gloves
510(k): K133168
Regulatory Class I
Product Code: LZA

4.0 Description of The Device:

Powder Free Nitrile Patient Examination Glove, Non-Sterile (Blue) meet all the requirements of ASTM standard D6319-19 and FDA 21 CFR 880.6250.

The powder free nitrile examination glove is manufactured from synthetic rubber latex. Inner surface of gloves undergoes surface treatment process to produce a smooth surface that assists the user in donning the gloves with ease without using any lubricant such as powder on the glove surface. The glove is ambidextrous, i.e., can be worn on right hand or left hand.

5.0 Intended Use of the Device:

A patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

510(k) SUMMARY

6.0 Comparison of the Technological Characteristics of the Device:

The Powder Free Nitrile Examination Gloves, Non-Sterile (Blue) are summarized with the following technological characteristics compared to ASTM D6319 or equivalent standards.

CHARACTERISTICS	STANDARDS	DEVICE PERFORMANCE		COMPARISON ANALYSIS
		PREDICATE	CURRENT	
		BLUE	BLUE	
510(k) Number		K133168	K210755	N/A
Manufacturer(s)	-	WRP Asia Pacific Sdn Bhd	Meditech Gloves Sdn Bhd	Same
Material	ASTM D6319-10	Nitrile	Nitrile	Same
Color	-	Blue	Blue	Same
Texture	-	Finger textured	Finger textured	Same
Physical Properties	ASTM D6319-10			
<u>Before Aging</u> Tensile Strength : Ultimate Elongation :		14MPa min 500% min	14MPa min 500% min	Same
<u>After Aging</u> Tensile Strength : Ultimate Elongation :		14MPa min 400% min	14MPa min 400% min	Same
Thickness - Finger - Palm - Cuff	ASTM D6319-19	0.07-0.10mm 0.07-0.09mm 0.06-0.08mm	0.07- 0.10mm 0.06- 0.09mm 0.06- 0.08mm	Similar
Powder Free	ASTM D6124-06	≤ 2 mg/glove	≤ 2 mg/glove	Same

510(k) SUMMARY

CHARACTERISTICS	STANDARDS	DEVICE PERFORMANCE		COMPARISON ANALYSIS
		PREDICATE	CURRENT	
		BLUE	BLUE	
Biocompatibility	Primary Skin Irritation – ISO 10993-10:2010	<p>Passes (Not a primary skin irritant)</p> <p>There was no erythema or oedema noted on test site after (24±2), (48±2) and (72±2) hours. The primary Irritation Index (PII) was “0”</p>	<p>Passes (Not a primary skin irritant)</p> <p>There was no erythema or oedema noted on test site after (1±0.1), (24±2), (48±2) and (72±2) hours. The primary Irritation Index (PII) was “0”. Also, no mortality after 72 hours. The gloves considered negligible.</p>	Similar
Biocompatibility	Dermal Sensitization- ISO 10993-10:2010	<p>Passes (Not a contact sensitizer)</p> <p>There was no positive allergic reaction observed during the challenge phase (at 24 hours and 48 hours) in animals treated with the test material and negative control.</p>	<p>Passes (Not a contact sensitizer)</p> <p>There was no positive allergic reaction observed during the challenge phase (at 0, 24 hours and 48 hours) in animals treated with the test material and negative control.</p>	Similar
Biocompatibility	Cytotoxicity – MEM Elution, ISO 10993-5:2009	<p>Exhibit severe cytotoxicity reactivity at 100%, 66%, 44% and 30% extract concentration.</p> <p>Slight cytotoxicity reactivity at 20% and no cytotoxicity reactivity at 15% extract concentrations.</p>	<p>Exhibit severe cytotoxicity reactivity at 100%, 50%, and 25% extract concentration.</p> <p>No cytotoxicity reactivity at 12.5%, 6.25% and 3.125% extract concentrations.</p>	Similar

510(k) SUMMARY

CHARACTERISTICS	STANDARDS	DEVICE PERFORMANCE		COMPARISON ANALYSIS
		PREDICATE	CURRENT	
		BLUE	BLUE	
Biocompatibility	Acute Systemic Toxicity, ISO 10993-11:2017 (E)	Not Available	Passes (no adverse biological reaction) No mortality was observed (72±2) hours	Not Available
Watertight (1000ml)	ASTM D5151-19	Inspection Level 1, AQL 1.5	Inspection Level 1, AQL 1.5	Same
Intended use		A patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.	A patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.	Same
Size	Medical Glove Guidance Manual - Labeling	Extra Small Small Medium Large Extra Large	Extra Small Small Medium Large Extra Large	Same
Single Use	Medical Glove Guidance Manual - Labeling	Single use	Single use	Same

510(k) SUMMARY

There are no significant differences between the two products and they are the same or similar in terms of intended use, materials, design, manufacturing methods.

7.0 Summary of Non-Clinical Performance Data

The performance test data of the non-clinical tests for this powder free nitrile examination gloves is summarized as per below.

510(k) SUMMARY

Test Method	Standard	Purpose of Testing	Acceptance Criteria			Results		Status
				Before aging	After aging	Before aging	After aging	
Physical Properties	ASTM D412-16 (Standard Test Method for Vulcanized Rubber and Thermoplastic Elastomers-Tension)	To evaluate the tensile (tension) properties of glove.	Tensile Strength	Min 14 MPa	Min 14 MPa	15.3	14.8	Pass
			Ultimate Elongation	Min 500%	Min 400%	500	460	Pass

Test Method	Standard	Purpose of Testing	Acceptance Criteria		Results		Status
Dimension	ASTM D3767-03(2020) Standard Practice for Rubber – Measurement of Dimensions	To measure the length, width, and thickness of glove	Length	Min 240mm	Min 240		Pass
			Width:				
			XS	70 ± 10 mm	Ave = 72 mm		Pass
			S	80 ± 10 mm	Ave = 84 mm		Pass
			M	95 ± 10 mm	Ave = 95 mm		Pass
			L	110 ± 10 mm	Ave = 102 mm		Pass
			XL	>110	Ave = 113 mm		Pass
Thickness	Finger – 0.05mm Palm – 0.05mm	Thickness	Finger – 0.10mm Palm – 0.06mm			Pass	

Test Method	Standard	Purpose of Testing	Acceptance Criteria	Results	Status
-------------	----------	--------------------	---------------------	---------	--------

510(k) SUMMARY

Water Tight	ASTM D 5151-19 Standard Test Method for Detection of Holes in Medical Gloves	To detect holes that leak water and thereby compromise the usefulness of the glove	Sample Size : 315 Inspection Level : GI AQL : 1.5 Acceptance No : 10	This batch sampling is 150,001 to 500,000. Hence according to single sampling plan GI, the sample to be drawn is under Code M equivalent to 315 pieces with accept 10 and reject 11. During the test , 6 pieces were found with leaks. Hence it falls within the acceptance criteria.	Pass
-------------	---	---	---	---	------

Test Method	Standard	Purpose of Testing	Acceptance Criteria	Results	Status
Residual Powder	ASTM D6124-06 (2017) Standard Test Method for Residual Powder on Medical Gloves	To determine the amount of residual powder and non- powder solids found on gloves	Less than 2 mg per glove	Sample size : 5 pcs Requirement : <2mg/glove Result : 1.5 mg/glove	Pass

510(k) SUMMARY

8.0 Summary of Clinical Performance Data

Not applicable - Clinical data was not used to assess performance of the subject device.

9.0 Conclusion

The conclusions drawn from the nonclinical tests demonstrate that the subject device, Powder Free Nitrile Patient Examination Gloves, Non-Sterile (Blue) (K210775), is as safe, as effective, and performs as well as or better than the legally marketed predicate device, Dermagrip Powder Free Blue Nitrile Examination Gloves (K133168).