



Maxx Orthopedics, Inc.
Priscilla Herpai
Regulatory Manager
2460 General Armistead Ave, Suite 100
Norristown, Pennsylvania 19403

October 29, 2020

Re: K192989

Trade/Device Name: Libertas E-XLPE Modular Liner

Regulation Number: 21 CFR 888.3353

Regulation Name: Hip Joint Metal/Ceramic/Polymer Semi-Constrained Cemented Or Nonporous
Uncemented Prosthesis

Regulatory Class: Class II

Product Code: OQI, LZO, OQG

Dated: September 18, 2020

Received: October 1, 2020

Dear Priscilla Herpai:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Vesa Vuniqi, M.S.
Assistant Director
DHT6A: Division of Joint Arthroplasty Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.

Indications for Use

510(k) Number (if known)

K192989

Device Name

Libertas™ E-XLPE Modular Liner

Indications for Use (Describe)

The Libertas™ Hip Replacement System is intended for use in total hip arthroplasty. Total hip arthroplasty is intended to provide patient mobility and reduce pain by replacing the damaged hip joint articulation in patients where there is evidence of sufficient sound bone to fix and support the components.

Total hip replacement is indicated for the following conditions:

- Non-inflammatory degenerative joint diseases including osteoarthritis, post traumatic arthritis and avascular necrosis.
- Rheumatoid arthritis.
- Congenital hip dysplasia.
- Acute traumatic fracture of the femoral head or neck.
- Certain cases of Ankylosis.
- Dislocation of the hip.
- Correction of functional deformity.
- Revision of failed joint reconstruction or treatment.
- Treatment of nonunion, femoral neck and trochanteric fractures of the proximal femur.

Note

- The Modular Shell, Uncemented Stem and Taper Uncemented Femoral Stem are intended for press-fit, uncemented use only.
- The Cemented stem is intended for cemented use only.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



5. 510(k) SUMMARY

510(k) Summary

This 510(k) Summary is submitted in accordance with the requirements of 21 CFR Part 807, Section 92.

5.1 Applicant:

Maxx Orthopedics Inc.
2460 General Armistead Ave
Suite 100, Norristown, PA 19403
USA

5.2 Maxx Contact Person:

Priscilla Herpai
Regulatory Manager
Maxx Orthopedics
E-mail: priscilla.herpai@maxxortho.com
Phone: +484-342-0092 x 507

5.3 Alternate contact Person:

Gayathri Nair
Senior Manager- Regulatory Affairs/ Quality Assurance
Meril Healthcare Private Limited
E mail: gayathri.nair@merillife.com
Cell: +91 9909033393

5.4 Date prepared: September 20, 2019

5.5 Device information:

Proprietary Name:	Libertas™ E-XLPE Modular Liner
Common / Usual Name:	Hip Joint Prosthesis
Classification name:	Hip joint metal/Ceramic/Polymer Semi-constrained Cemented or Nonporous Uncemented Prosthesis (21 CFR 888.3353)
Product Code:	OQI, LZ0, OQG
Device Class:	Class II



5.6 Predicate Devices:

Component	Equivalent device category	Manufacturer	Trade name	510(k)
Vitamin E containing HXLPE	Primary Predicate device	Maxx Orthopedics Inc., USA	Libertas™ – Hip Replacement System	K180973
Modular Liner (E-XLPE Modular liner)	Predicate Device	Maxx Orthopedics Inc., USA	Libertas™ Acetabular Hooded Liner	K183365
	Reference device	Corin, USA	Trinity Acetabular cup System ECIMA Liners	K111481

5.7 Device Description:

This 510k is intended to add a Modular Liner made from Vitamin E containing HXLPE (E-XLPE) to the Libertas™ – Hip Replacement System and Libertas™ Acetabular Hooded Liner cleared under K180973 and K183365, respectively. These Modular Liners are designed to be used with the Libertas components already cleared under K180973. These are available in different sizes i.e. MA, MB, MD, MF, MH, MJ, MK and different variants for each size Viz, neutral offset, “+4 mm offset”, “Elevated wall”, “10° Oblique”, “15° Oblique” and “+4 mm offset 10° Oblique”.

5.8 Indications for use:

The Libertas™ Hip Replacement System is intended for use in total hip arthroplasty. Total hip arthroplasty is intended to provide patient mobility and reduce pain by replacing the damaged hip joint articulation in patients where there is evidence of sufficient sound bone to fix and support the components.

Total hip replacement is indicated for the following conditions:

- Non-inflammatory degenerative joint diseases including osteoarthritis, post traumatic arthritis and avascular necrosis.
- Rheumatoid arthritis.
- Congenital hip dysplasia.
- Acute traumatic fracture of the femoral head or neck.
- Certain cases of Ankylosis.
- Dislocation of the hip.
- Correction of functional deformity.



- Revision of failed joint reconstruction or treatment.
- Treatment of nonunion, femoral neck and trochanteric fractures of the proximal femur.

Note:

- The Modular Shell, Uncemented Stem and Taper Uncemented femoral stem are intended for press-fit, uncemented use only.
- The Cemented stem is intended for cemented use only.

5.9 Comparison of technological characteristics:

The Libertas™ E-XLPE Modular Liner is substantially equivalent to the previously cleared predicate devices based on similarities in intended use, device design/technological characteristics, materials, and sterilization method.

5.10 Non clinical Performance data:

Non-clinical testing conducted to evaluate device function/mechanical performance and to demonstrate substantial equivalence.

- Axial disassembly (push out) test (ASTM F1820:13)
- Offset pull out (Lever out) test (ASTM F1820:13)
- Torque out disassembly test (ASTM F1820:13)
- Impingement Test (ASTM F2582:14)
- Range of Motion (ISO 21535-2007/Amd 1:2016)
- Material Characterization of E-XLPE (ASTM F2565:13, ASTM F2695:12, ASTM F2759:11, ASTM F648:14, ASTM F2003:15, ISO 5834-3:19, ISO 5834-1:19 and ISO 5834-2:19)
- Extraction Testing
- Biocompatibility testing

5.11 Conclusion

Based on performance testing results and similarities in intended use, device design/technological characteristics, materials, and sterilization method, the Libertas™ E-XLPE Modular Liner is considered substantially equivalent to the previously cleared predicate and reference devices.