



June 3, 2020

Argon Medical Devices, Inc.
Ana Jimenez-Hughes
Sr. Regulatory Affairs Specialist
1445 Flat Creek Road
Athens, Texas 75751

Re: K200268
Trade/Device Name: Halo™ Single-Loop Snare Kit
Regulation Number: 21 CFR 870.5150
Regulation Name: Embolectomy Catheter
Regulatory Class: Class II
Product Code: MMX
Dated: April 22, 2020
Received: April 24, 2020

Dear Ms. Jimenez-Hughes:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or post-marketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for

Gregory O'Connell
Assistant Director
DHT2C: Division of Coronary
and Peripheral Intervention Devices
OHT2: Office of Cardiovascular Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K200268

Device Name

Halo™ Single-Loop Snare Kit

Indications for Use (Describe)

The Halo™ Single-Loop Snare Kit is intended for use in the cardiovascular system or hollow viscous to retrieve and manipulate foreign objects.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Date Prepared: January 31, 2020

Company: Argon Medical Devices, Inc.
1445 Flat Creek Road
Athens, Texas 75751 USA
Facility Registration number: 1625425

Contact: Ana Jimenez-Hughes
Sr. Regulatory Specialist
Phone: 903-676-4276
Fax: 903-677-9396
Email: ana.hughes@argonmedical.com

Device Trade Name: Halo™ Single-Loop Snare Kit

Device Common Name: Percutaneous Retrieval Device

Device Classification: Device, Percutaneous Retrieval
Product code, MMX
21 CFR 870.5150
Class II
Review Panel: Cardiovascular Devices

Predicate Device(s): *Primary:* K972511 Amplatz Goose Neck Snare Kit/Catheter
Reference: K122088 Merit ONE Snare™ Endovascular Snare System

Description of the Device: Halo™ Single-Loop Snare Kit contains: (1) Snare, (1) Snare Catheter, (1) Introducer and (1) Torque Handle.

The snare is constructed of a flexible and radiopaque loop. The pre-formed snare loop can be introduced through the snare catheter without risk of snare deformation because of the snare's super-elastic construction. The snare catheter is constructed of flexible tubing and contains a radiopaque marker band.

Indication for Use: The Halo™ Single-Loop Snare Kit is intended for use in the cardiovascular system or hollow viscous to retrieve and manipulate foreign objects.

Technological Characteristics:

A comparison of the technological characteristics of the subject device and the predicate devices shows the Halo™ Single-Loop Snare Kit to be substantially equivalent to the current marketed predicate devices.

Equivalence is based upon the product performance, design and intended use. The Halo™ Single-Loop Snare Kit and the predicate devices have similar materials of construction, dimensional specifications, designs and sterilization process.

Performance Tests (Non-Clinical):

No performance standards have been established under section 514, performance standards, of the Food, Drug and Cosmetic Act for these devices. A series of testing was conducted in accordance with protocols based on requirements outlined in guidances and industry standards and the below were shown to meet the acceptance criteria that were determined to demonstrate substantial equivalence.

The following tests were performed under the specified testing parameters to support the Halo™ Single-Loop Snare Kit substantial equivalence.

Performance Testing, including:

- Tensile strength
- Liquid leakage
- Air leakage
- Corrosion Resistance
- System Tip Flexibility
- Tip Flexibility – Snare & Catheter
- Snare Flexing & Fracture Test
- Catheter Flexural Modulus
- Catheter Kink Test
- Marker Band Pull Test
- Torque Strength Test
- Simulative Use
- Radiopacity
- Particulate
- Luer Testing
- Shipping Test

Biocompatibility Testing, including:

- Cytotoxicity (ISO 10993-5)
- Sensitization (ISO 10993-10)
- Intracutaneous Irritation (ISO 10993-10)
- Acute Systemic Toxicity (ISO 10993-11)
- Material Mediated Pyrogen (ISO 10993-11)
- Hemocompatibility (ISO10993-4)
 - ASTM Hemolysis – Direct and Indirect Contact
 - Complement Activation, SC5b-9
 - Platelet and Leucocyte Counts
 - Partial Thromboplastin Time (PTT)

**Substantial
Equivalence:**

Based on the Indication for Use, design, and safety and performance testing, the Halo™ Single-Loop Snare Kit meets the requirements for its intended use and is substantially equivalent to the predicate devices.

Conclusion:

The results of all testing demonstrate that the Halo™ Single-Loop Snare Kits are substantially equivalent to the predicate devices.
