



November 17, 2020

Sorin Group Italia S.r.l.
Luigi Vecchi
Director, Regulatory Affairs
Via statale 12 Nord, 86
Mirandola, Modena 41037
Italy

Re: K200612

Trade/Device Name: R501 aortic root cannula without vent line, R502 aortic root cannula with vent line
Regulation Number: 21 CFR 870.4210
Regulation Name: Cardiopulmonary Bypass Vascular Catheter, Cannula, Or Tubing
Regulatory Class: Class II
Product Code: DWF
Dated: October 14, 2020
Received: October 15, 2020

Dear Luigi Vecchi:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for Fernando Aguel
Assistant Director
DHT2B: Division of Circulatory Support,
Structural and Vascular Devices
OHT2: Office of Cardiovascular Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K200612

Device Name
R501 Aortic Root Cannula without vent line
R502 Aortic Root Cannula with vent line

Indications for Use (Describe)

The Aortic root cannulae are intended for use in adult and pediatric patients to cannulate the aortic root for cardioplegic solution delivery into the coronary arteries and for venting of the heart during cardiopulmonary bypass surgery for periods of up to six hours

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Section 8 – 510(k) Summary

(in accordance with 21 CFR 807.92)

510(k) Number: K200612

I. Applicant Information

Applicant:

SORIN GROUP ITALIA S.R.L.
Via Statale 12 Nord, 86
Mirandola, MO
41037 Italy

Contact Person:

Luigi Vecchi
Director, Regulatory Affairs
Tel: +39 0535 29957
e-mail: luigi.vecchi@livanova.com

Application Correspondent:

SORIN GROUP ITALIA S.R.L.
Via Statale 12 Nord, 86
Mirandola, MO
41047 Italy

Contact Person:

Luigi Vecchi
Director, Regulatory Affairs
Tel: +39 0535 29957
e-mail: luigi.vecchi@livanova.com

Date Prepared:

February 13th 2020

II. Subject Device Identification

Proprietary Name: R501 Aortic Root Cannula without vent line
R502 Aortic Root Cannula with vent line

Common/Usual Name: Cardiopulmonary bypass vascular cannula

Classification Name: Cardiopulmonary bypass vascular catheter,
cannula or tubing

Regulation Number: 21 CFR 870.4210

Product Code: DWF

Classification: Class II

Classification Panel: Cardiovascular

III. Predicate Device

The R501-R502 Aortic Root Cannula is substantially equivalent to the following cleared predicate device. Both models have the same fundamental scientific technology and intended use:

510(k) Number:	K861310
Proprietary Name:	R501 Aortic Root Cannula without vent line R502 Aortic Root Cannula with vent line
Common/Usual Name:	Cardiopulmonary bypass vascular cannula
Classification Name:	Cardiopulmonary bypass vascular catheter, cannula or tubing
Regulation Number:	21 CFR 870.4210
Product Code:	DWF
Classification:	Class II
Classification Panel:	Cardiovascular

IV. Device Description

The R501-R502 Aortic Root Cannula (hereinafter referred to as R501-R502 cannulae) are aortic root cannulae intended to be used to cannulate the aortic root, to vent and to deliver cardioplegic solution during heart bypass operations

The device is available in two styles: with a vent line (R502) and without a vent line (R501). Both models consist of a flexible Polyvinylchloride (PVC) tubing body, a PVC flanged tip and luer connector. The introducer is made of stainless steel with a PVC obturator. The vent line model have an additional feature constituted by a vent line provided as integral part of the device . This attribute makes it possible to vent and also introduce cardioplegic solutions using the same device. These cannulae are available in a range of sizes as per table below:

Model	Tip size (internal diameter)	Product designation
R501 (aortic root without vent line)	1.5 mm	R501-15
	2.0 mm	R501-20
	2.6 mm	R501-26
R502 (aortic root with vent line)	1.5 mm	R502-15
	2.0 mm	R502-20
	2.6 mm	R502-26

The R501-R502 cannulae are a modified version of the currently marketed R501-R502 Aortic Root cannulae (K861310).

V. Indications for Use

The Aortic root cannulae are intended for venting the aortic root for delivery of cardioplegic solution and for rapid and secure perfusion into the coronary arteries during cardiopulmonary surgery for periods of up to six hours.

VI. Summary of Technical Characteristics

The **R501-R502 Cannulae** have the same fundamental technological characteristics, principles of operation and control mechanisms as the unmodified devices.

The **R501-R502 Cannulae** are provided with a different material of one specific component of the device (the flanged tip): the PVC material used for this component in the **R501-R502 Cannulae** is without Phthalates while the original PVC material used in the unmodified devices does contain Phthalates

No other design changes have been made to the device. Except for the change of the tip material, the R501-R502 cannulae utilize the same materials as the unmodified cannulae.

The devices are ethylene oxide sterilized and have a non-pyrogenic fluid path. They are for single use only.

VII. Non-Clinical Performance Data

SORIN GROUP ITALIA S.R.L. has conducted extensive verification and validation testing of the **R501-R502 Cannulae**, as a cardiopulmonary bypass arterial cannula capable of providing adequate perfusion to the patient during surgery.

The **R501-R502 Cannulae** complies with all the applicable voluntary standards related to cardiopulmonary bypass arterial cannulae. The device passed all the testing in accordance with national and international standards.

VIII. Clinical Performance Data

No clinical testing was conducted in support of the **R501-R502 Cannulae**, as the indications for use and technical characteristics are equivalent to those of the predicate device, which has been on the market for several years with proven safety and efficacy of use. The non-clinical testing summarized in this submission supports the substantial equivalence of the subject device with the predicate device when used according to its intended use.

IX. Statement of Substantial Equivalence

Based on equivalent intended use and technological characteristics, as well as on equivalent performance testing, the **R501-R502 Cannulae** can be deemed to be substantially equivalent to its predicate device, the Unmodified **R501-R502 Cannulae**, cleared under **K861310**.

The **R501-R502 Cannulae**, as designed and manufactured, does not raise new questions regarding safety and effectiveness as compared to its predicate device and is determined to be substantially equivalent to its predicate device, the Unmodified **R501-R502 Cannulae**.