



November 23, 2020

Alphatec Spine, Inc.
Cynthia Dorne
Manager, Regulatory Affairs
5818 E1 Camino Real
Carlsbad, California 92008

Re: K203056

Trade/Device Name: Invictus™ Spinal Fixation System
Regulation Number: 21 CFR 888.3070
Regulation Name: Thoracolumbosacral Pedicle Screw System
Regulatory Class: Class II
Product Code: NKB, KWP
Dated: October 7, 2020
Received: October 8, 2020

Dear Cynthia Dorne:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin O'Neill, M.B.E.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K203056

Device Name

Invictus™ Spinal Fixation System

Indications for Use (Describe)

The Invictus Spinal Fixation System is intended for posterior, non-cervical fixation in skeletally mature patients as an adjunct to fusion for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e. Fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion.

When used for posterior non-cervical pedicle screw fixation in pediatric patients, the Invictus Spinal Fixation System implants are indicated as an adjunct to fusion to treat progressive spinal deformities (i.e., scoliosis, kyphosis, or lordosis) including idiopathic scoliosis, neuromuscular scoliosis, and congenital scoliosis. Additionally, the Invictus Spinal Fixation System is intended to treat pediatric patients diagnosed with the following conditions: spondylolisthesis / spondylolysis, fracture caused by tumor and/or trauma, pseudarthrosis, and/or failed previous fusion. Pediatric pedicle screw fixation is limited to a posterior approach.

The Invictus Spinal Fixation System is intended to be used with autograft and/or allograft.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

This 510(k) summary of safety and effectiveness is being submitted in accordance with the requirements of 21 CFR 807.92.

I. SUBMITTER: Alphatec Spine, Inc.
5818 El Camino Real
Carlsbad, CA 92008
Phone: (760) 431-9286
Fax: (760) 431-0289

Contact Person: Cynthia Dorne
Manager, Regulatory Affairs
Contact Phone: (760) 494-6740

Date Summary Prepared: October 6, 2020

II. DEVICE

Name of Device: Invictus™ Spinal Fixation System
Common or Usual Name: Thoracolumbosacral pedicle screw system
Classification Name: Thoracolumbosacral pedicle screw system (21 CFR 888.3070)
Spinal interlaminar fixation orthosis (21 CFR 888.3050)
Regulatory Class: Class II
Product Code: NKB, KWP

III. LEGALLY MARKETED PREDICATE DEVICES

510(k)	Product Code	Trade Name	Manufacturer
Primary Predicate Device			
K192938	NKB, KWP	Invictus™ Spinal Fixation System	Alphatec Spine
Additional Predicate Devices			
K181677	NKB, KWP	Kodiak™ Spinal Fixation System	Alphatec Spine

IV. DEVICE DESCRIPTION

The *Invictus Spinal Fixation System* is a thoracolumbosacral pedicle screw system designed to be implanted through a posterior surgical approach. The implants are manufactured from titanium alloy (Ti-6Al-4V ELI) per ASTM F136, commercially pure titanium (CP Ti Grade 2) per ASTM F67 and cobalt chromium (Co-28Cr-6Mo) alloy per ASTM F1537. The Invictus System consists of a variety of shapes and sizes of screws, hooks, rods, connectors, and cross-connectors to create a rigid construct as an adjunct to fusion for temporary internal fixation and stabilization of the thoracic, lumbar and sacral spine.

The purpose of this submission is to add new components to the *Invictus Spinal Fixation System*.

V. INDICATIONS FOR USE

The Invictus Spinal Fixation System is intended for posterior, non-cervical fixation in skeletally mature patients as an adjunct to fusion for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e. Fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion.

When used for posterior non-cervical pedicle screw fixation in pediatric patients, the Invictus Spinal Fixation System implants are indicated as an adjunct to fusion to treat progressive spinal deformities (i.e., scoliosis, kyphosis, or lordosis) including idiopathic scoliosis, neuromuscular scoliosis, and congenital scoliosis. Additionally, the Invictus Spinal Fixation System is intended to treat pediatric patients diagnosed with the following conditions: spondylolisthesis / spondylolysis, fracture caused by tumor and/or trauma, pseudarthrosis, and/or failed previous fusion. Pediatric pedicle screw fixation is limited to a posterior approach.

The Invictus Spinal Fixation System is intended to be used with autograft and/or allograft.

VI. TECHNOLOGICAL COMPARISON TO PREDICATES

The technological design features of the subject implants were compared to the predicates in intended use, indications for use, design, function and technology and it was demonstrated that they are substantially equivalent.

VII. PERFORMANCE DATA

Nonclinical testing performed on the Invictus Spinal Fixation System supports substantial equivalence to other predicate devices. The following testing was performed:

- ASTM F1717 Dynamic Compression Bending
- ASTM F1798 Static Flexion-Extension Moment (M_y)

The results demonstrate that the subject Invictus Spinal Fixation System is substantially equivalent to other predicate devices for nonclinical testing.

Clinical Information

Not applicable; determination of substantial equivalence is not based on an assessment of clinical performance data.

VIII. CONCLUSION

Based upon the information provided in this 510(k) submission, it has been determined that the subject devices are substantially equivalent to legally marketed devices in regards to indications for use, intended use, design, technology, and performance.