



August 25, 2021

Pulse Biosciences, Inc.
Punam Gollamudi
Director US Regulatory Affairs
3957 Point Eden Way
Hayward, California 94545

Re: K211444

Trade/Device Name: CellFX System
Regulation Number: 21 CFR 878.4400
Regulation Name: Electrosurgical Cutting and Coagulation Device and Accessories
Regulatory Class: Class II
Product Code: GEI
Dated: August 9, 2021
Received: August 10, 2021

Dear Punam Gollamudi:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531 - 542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Long Chen, Ph.D.
Assistant Director
DHT4A: Division of General Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K211444

Device Name

CellFX® System

Indications for Use (Describe)

The CellFX® System is intended for dermatological procedures requiring ablation and resurfacing of the skin.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY

510(k) Summary**I. Submitter**

Applicant: Pulse Biosciences, Inc.
3957 Point Eden Way
Hayward, CA 94545
Phone: (510) 906-4649

Contact Person: Punam Gollamudi
Director of US Regulatory Affairs
Pulse Biosciences, Inc.
Phone: (415) 305-4404

Date Prepared: August 24, 2021

II. DEVICE INFORMATION

Trade Name: CellFX[®] System
Regulation Number 21 CFR § 878.4400
Regulation Name: Electrosurgical Cutting and Coagulation Device and
Accessories
Regulation Class: Class II
Product Code: GEI
Classification Panel: General and Plastic Surgery

III. PREDICATE DEVICE

K203299 – CellFX[®] System, Pulse Biosciences, Inc.
No reference devices were used in this submission.

IV. DEVICE DESCRIPTION

The Pulse Biosciences[®] CellFX[®] System is a proprietary platform technology. The CellFX System consists of the CellFX Console, CellFX Handpiece, CellFX Treatment Tips, and CellFX Software. The CellFX System delivers nanosecond duration electrical energy pulses that disrupt the function of cells leading to regulated cell death, while sparing non-cellular tissue. The CellFX System delivers a series of timed, nanosecond electrical pulses (referred to as a “Cycle”) to ablate and resurface tissue areas in dermatologic conditions.

The CellFX Console is capable of delivering short electric pulses at amplitudes up to 15 kV and pulse widths up to 700 ns. The electrical energy pulses are applied directly to targeted tissue using sterile Treatment Tips with stainless steel microneedles. The treatment parameters are selected by the user through a user interface on the Touchscreen Display of the CellFX Console.

V. INDICATIONS FOR USE STATEMENT

The CellFX[®] System is intended for dermatological procedures requiring ablation and resurfacing of the skin.

VI. COMPARISON OF TECHNOLOGICAL CHARACTERISTICS WITH THE PREDICATE DEVICE

To support a determination of substantial equivalence, Pulse Biosciences performed verification and validation testing demonstrating the subject device performs as intended. The design, materials, and energy source remain the same. Based on the data provided, the additional Treatment Tips do not raise any new or different questions of safety and effectiveness.

	Subject Device	Predicate Device (K203299)
Trade Name	CellFX [®] System	CellFX [®] System
Prescription Only	Yes	Yes
Regulation Number	878.4400	878.4400
Product Code	GEI	GEI
Indications for Use		
Indications for Use Statement	The CellFX System is intended for dermatological procedures requiring ablation and resurfacing of the skin.	The CellFX System is intended for dermatological procedures requiring ablation and resurfacing of the skin.
Technological Characteristics		
Mechanism of Action	Nano-Pulse Stimulation: ultrafast nanosecond electrical pulses to the targeted tissue via treatment tips with an array of microneedles	Nano-Pulse Stimulation: ultrafast nanosecond electrical pulses to the targeted tissue via treatment tips with an array of microneedles
Pulse Frequency	1 Hz -10 Hz	1 Hz -10 Hz
Pulse Amplitude/Voltage	300V to 15kV	300V to 15kV
Pulse Width	100 ns – 700 ns	100 ns – 700 ns
Power Input	100-240 VAC, 50/60 Hz, 2A	100-240 VAC, 50/60 Hz, 2A
Power Output	30 Watts	30 Watts
Energy Density Range	1.5 x 1.5 mm – 110-190 mJ/mm ³ 2.5 x 2.5 mm – 60-110 mJ/mm ³ 5.0 x 5.0 mm – 45-85 mJ/mm ³	1.5 x 1.5 mm – 110-190 mJ/mm ³ 2.5 x 2.5 mm – 60-110 mJ/mm ³ 5.0 x 5.0 mm – 45-85 mJ/mm ³
Use of Local Anesthesia	Yes	Yes
Sterile Tip	EO Sterilization	EO Sterilization
Handpiece	Single Handpiece for 2-Row Tips	Single Handpiece for 2-Row Tips

System Physical Characteristics		
Height	132 cm	132 cm
Length	46 cm	46 cm
Width	53 cm	53 cm
Weight	54 kg	54 kg
Power Input	100-240 VAC, 50/60 Hz, 2A	100-240 VAC, 50/60 Hz, 2A
Tip Characteristics		
Treatment Tip	2-Row	2-Row
Microneedle Configuration		
1.5 x 1.5 mm	2 rows of 3 needles per row	2 rows of 3 needles per row
2.5 x 2.5 mm	2 rows of 3 or 4 needles per row	2 rows of 3 needles per row
5.0 x 5.0 mm	2 rows of 4 or 6 needles per row	2 rows of 4 needles per row
Tip Insertion Depth	1 mm and 2 mm	2 mm

VII. PERFORMANCE DATA

The following performance data were provided in support of the substantial equivalence determination.

Biocompatibility Testing

Biocompatibility testing was not required as there were no changes to the device compared to the predicated that would impact biocompatibility.

Electrical safety and electromagnetic compatibility (EMC)

The CellFX System complies with IEC 60601-1, IEC 60601-1-2, 60601-2-2, IEC 60601-1-6, IEC 62366, and IEC 62304. All test conditions were performed as outlined within the IEC standards. All electrical safety and EMC tests passed.

Bench Testing

The performance testing for the CellFX System (Console, Handpiece and Tip) verified that the CellFX System performed as intended and met product specifications.

Software Verification and Validation Testing

Software verification and validation testing were conducted and documentation was provided as recommended by FDA's Guidance for Industry and FDA Staff, "Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices." The software for this device was considered as a "moderate" level of concern since a failure of the device could directly result in minor injury to the patient or operator. Completed software testing supports the safety and effectiveness of the device.

Cybersecurity controls have been implemented to mitigate the risk of malware being introduced into the CellFX System as recommended by FDA's Draft Guidance for

Industry and FDA Staff, “Content of Premarket Submissions for Management of Cybersecurity in Medical Devices.”

Animal Study

An animal safety study was conducted to validate the safety and performance of the CellFX System.

Validation Report for CellFX System Using 2-Row Tips in Swine Skin

An animal study was conducted to characterize treatment ablation areas to validate the safety and performance of the CellFX System. The study was conducted on two Yucatan pigs that underwent skin treatments with the subject device including 2-Row CellFX Treatment Tips (1.5mm, 2.5mm, and 5.0mm).

Safety

All CellFX treatments were successfully completed with no adverse events. Pathology results support the safety of the CellFX System with acceptable histological findings and no evidence of thermal damage at CellFX treatment sites.

Performance

All test articles met the acceptance criteria for device performance. The treatment ablation areas met the acceptance criteria and successfully achieved degeneration and necrosis of active treatment sites while the epidermis remained intact in the majority of the treatment sites. The device performed as intended.

Conclusion

CellFX treatments in a swine skin animal model were shown to be safe and performed as intended. Therefore, this animal study data confirmed that the device performs as intended and does not raise safety concerns.

CONCLUSION

The CellFX System has the same indication for use, technological characteristics and principles of operation as its predicate device. The non-clinical performance testing and *in vivo* animal data provided in this submission demonstrate and support that the CellFX System is as safe and as effective as the predicate device. Therefore, the CellFX System is substantially equivalent to the predicate device.