



January 11, 2022

Stryker GmbH
Danielle Jannuzzi Madureira
Staff Regulatory Affairs Specialist
325 Corporate Drive
Mahwah, New Jersey 07430

Re: K213199

Trade/Device Name: Asnis® III Cannulated Screw System and Asnis® PRO Cannulated Screw System
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: HWC, HTN
Dated: December 17, 2021
Received: December 20, 2021

Dear Danielle Jannuzzi Madureira:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, MPH
Assistant Director
DHT6C: Division of Restorative, Repair
and Trauma Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K213199

Device Name

Asnis® III Cannulated Screw System

Indications for Use (Describe)

The Asnis® III Cannulated Screw System intended for fracture fixation of small and long bones and of the pelvis. The system is not intended for spinal use.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Indications for Use

510(k) Number (if known)

K213199

Device Name

Asnis® PRO Cannulated Screw System

Indications for Use (Describe)

The Asnis® PRO Cannulated Screw System intended for fracture fixation of small and long bones and of the pelvis. The system is not intended for spinal use.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

K213199
510(k) Summary

Proprietary Name: Asnis® III Cannulated Screw System and Asnis® PRO Cannulated Screw System

Common Name: Bone Screw (primary)
Washer, Bolt Nut

Regulation Description: Smooth or threaded metallic bone fixation fastener (primary)
Single/multiple component metallic bone fixation appliances and accessories

Regulation Number: 21 CFR 888.3040 (primary), 21 CFR 888.3030

Classification Product Code: HWC (primary), HTN

Device Class: II

Sponsor: Stryker GMBH
Bohnackerweg 1
2545 Selzach, Switzerland

Contact Person: Danielle Jannuzzi Madureira, PhD
Staff Regulatory Affairs Specialist
Dr. Homer Stryker Strasse, 1
CH-2545 Selzach, Switzerland
Phone: +41 79 890 02 89
Fax: +41 32 641 66 60

Date: September 24, 2021

Primary Predicate Device: Asnis® III Cannulated Screw System (K000080)

Additional Predicate Device: Biomet Cannulated Screw System (K140891)

Device Description: The Asnis® III Cannulated Screw System, previously cleared in K000080, K024060, consists of self-tapping cannulated screws and the corresponding washers. All devices in the system are provided sterile and non-sterile. The thread diameters are 4.0, 5.0, 6.5, and 8.0 mm. They are either fully or partially threaded. All screws are self-drilling and self-tapping. There are corresponding washers to the 4.0 and 5.0 mm screws respectively and one washer fitting for both diameters, 6.5 and 8.0 mm. Screws and washers are made of stainless steel and titanium alloy.

The Asnis® PRO Cannulated Screw System consists of self-tapping cannulated screws. All devices in the system are provided sterile and non-sterile. The thread diameters are 6.5, and 8.0 mm. They are either fully or partially threaded. All screws are self-drilling and self-tapping. Screws are made of stainless steel and titanium alloy.

The subject of this bundled traditional submission is to introduce minor design specification changes; MRI Labeling, shelf-life update of existing package to 10 years, and re-branding Ø6.5/8.0 mm long screws (130-180 mm to Asnis® PRO Cannulated Screw System) to the Asnis® III Cannulated Screw System, cleared in K000080 and K024060 and to introduce a new system Asnis® PRO Cannulated Screw System.

- Indications for Use:** The Asnis® III Cannulated Screw System intended for fracture fixation of small and long bones and of the pelvis. The system is not intended for spinal use.
The Asnis® PRO Cannulated Screw System intended for fracture fixation of small and long bones and of the pelvis. The system is not intended for spinal use.
- Comparison to Predicate Device:** The intended use of the modified devices, as described in its labeling, has not changed because of the modifications proposed in the present submission. The introduction of minor design specification changes, MRI Labeling, shelf-life update of existing package to 10 years, nor re-branding Ø6.5/8.0 mm long screws (130-180 mm) to Asnis® PRO Cannulated Screw System nor the new system (Asnis® PRO Cannulated Screw System) introduction does not alter the fundamental scientific technology shared by both the subject devices, Asnis® III Cannulated Screw System and Asnis® PRO Cannulated Screw System, and predicate device, Asnis® III Cannulated Screw System.
- Performance Data (Nonclinical):** *Non-Clinical Performance and Conclusions:*
No mechanical testing was deemed necessary as the design modifications and the new system introduction do not create a new worst case. All bench tests performed in accordance with ASTM F543 and previously presented in Asnis® III Cannulated Screw System (K024060), remain true and accurate. Tests performed to establish compatibility with a magnetic resonance environment:
- Magnetically Induced Displacement per ASTM F2052
 - Magnetically Induced Torque per ASTM F2213
 - RF Heating per ASTM F2182
 - Image Artifacts per ASTM F 2119
- Clinical Performance and Conclusions:*
Clinical data and conclusions were not needed for this device.
- Conclusion:** The subject devices have the same intended use and indications for use as the predicate device. The subject devices use the same operating principle, incorporate the same basic design and labeling and are manufactured and sterilized using the same materials and processes as the predicate device. Except for the modifications described in this submission the subject devices are identical to the predicate device, and the performance data and analyses demonstrate that:
- any differences do not raise new questions of safety and effectiveness as established with performance testing; and
 - the subject devices are at least as safe and effective as the legally marketed predicate devices