



December 14, 2021

Lisa Vershave
Regulatory Affairs Manager
PerkinElmer Inc
940 Winter Street
Waltham, MA 02451

Re: EUA202791/S002
Trade/Device Name: PKamp Respiratory SARS-CoV-2 RT-PCR Panel 1
Dated: October 27, 2021
Received: October 28, 2021

Dear Ms. Vershave:

This is to notify you that your request to update the Instructions for Use (IFU) of the PKamp Respiratory SARS-CoV-2 RT-PCR Panel 1 to update the reagent stability claim from 12 to 24 months, is granted. Upon review, we concur that the data and information submitted in EUA202791/S002 supports the requested updates for use with the PKamp Respiratory SARS-CoV-2 RT-PCR Panel 1. By submitting this EUA revision for review by the Food and Drug Administration (FDA), you have complied with the Conditions of Authorization stated in the letter authorizing the emergency use of the PKamp Respiratory SARS-CoV-2 RT-PCR Panel 1 issued on October 6, 2021.

Sincerely yours,

Uwe Scherf, M.Sc., Ph.D.
Director, Division of Microbiology Devices
OHT7: Office of In Vitro Diagnostics and Radiological Health
Office of Product Evaluation and Quality
Center for Devices and Radiological Health