

January 29, 2020

GE Healthcare Finland OY Joel Kent Senior Regulatory Affairs Manager Kuortaneenkatu 2 00510 Helsinki Finland

Re: K191323

Trade/Device Name: CARESCAPE B850 Regulation Number: 21 CFR 870.1025

Regulation Name: Arrhythmia Detector and Alarm (Including ST-Segment Measurement and Alarm)

Regulatory Class: Class II

Product Code: MHX, BZK, BZL, BZQ, CAP, CBQ, CBR, CBS, CCK, CCL, DPS, DPZ, DQA, DQK,

DRT, DSI, DSJ, DSK, DXG, DXN, FLL, GWJ, GWQ, KOI, KRB, MLD, NHO,

NHP, NHQ, OLT, OLW, OMC, ORT

Dated: December 23, 2019 Received: December 26, 2019

#### Dear Joel Kent:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <a href="https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm">https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm</a> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

K191323 - Joel Kent Page 2

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <a href="https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products">https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products</a>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <a href="https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems">https://www.fda.gov/medical-device-problems</a>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<a href="https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance">https://www.fda.gov/training-and-continuing-education/cdrh-learn</a>) and CDRH Learn (<a href="https://www.fda.gov/training-and-continuing-education/cdrh-learn">https://www.fda.gov/training-and-continuing-education/cdrh-learn</a>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<a href="https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice">https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice">https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice</a>) for more information or contact DICE by email (<a href="DICE@fda.hhs.gov">DICE@fda.hhs.gov</a>) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Jennifer Shih
Assistant Director (Acting)
Division of Cardiac Electrophysiology,
Diagnostics and Monitoring Devices
Office of Cardiovascular Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

**Enclosure** 

#### DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration

## Indications for Use

Form Approved: OMB No. 0910-0120 Expiration Date: 06/30/2020

See PRA Statement below.

510(k) Number (if known)		
K191323		
Device Name		
CARESCAPE B850		
Indications for Use (Describe)		

The CARESCAPE B850 is a multi-parameter patient monitor intended for use in multiple areas within a professional healthcare facility.

The CARESCAPE B850 is intended for use on adult, pediatric, and neonatal patients and on one patient at a time. The CARESCAPE B850 is indicated for monitoring of:

- hemodynamic (including ECG, ST segment, arrhythmia detection, ECG diagnostic analysis and measurement, invasive pressure, non-invasive blood pressure, pulse oximetry, cardiac output (thermodilution and pulse contour), temperature, mixed venous oxygen saturation, and central venous oxygen saturation),
- respiratory (impedance respiration, airway gases (CO2, O2, N2O, and anesthetic agents), spirometry, gas exchange),
- neurophysiological status (including electroencephalography, Entropy, Bispectral Index (BIS), and neuromuscular transmission).

The CARESCAPE B850 can be a stand-alone monitor or interfaced to other devices. It can also be connected to other monitors for remote viewing and to data management software devices via a network.

The CARESCAPE B850 is able to detect and generate alarms for ECG arrhythmias: atrial fibrillation, accelerated ventricular rhythm, asystole, bigeminy, bradycardia, ventricular couplet, irregular, missing beat, multifocal premature ventricular contractions (PVCs), pause, R on T, supra ventricular tachycardia, tachycardia, trigeminy, ventricular bradycardia, ventricular fibrillation/ventricular tachycardia, ventricular tachycardia, and VT>2. The CARESCAPE B850 also shows alarms from other ECG sources.

The CARESCAPE B850 also provides other alarms, trends, snapshots and events, and calculations, and can be connected to displays, printers and recording devices.

The CARESCAPE B850 is intended for use under the direct supervision of a licensed healthcare practitioner, or by personnel trained in proper use of the equipment in a professional healthcare facility.

Contraindications for using the monitor

The CARESCAPE B850 is not intended for use in a controlled MR environment.

Type of Use (Select one or both, as applicable)			
Prescription Use (Part 21 CFR 801 Subpart D)	Over-The-Counter Use (21 CFR 801 Subpart C)		
CONTINUE ON A SEPARATE PAGE IF NEEDED.			

This section applies only to requirements of the Paperwork Reduction Act of 1995.

#### \*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\*

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer Paperwork Reduction Act (PRA) Staff PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



GE Healthcare Finland Oy Kuortaneenkatu 2 00510 Helsinki Finland T: +358 10 39411

### 510(k) Summary

In accordance with 21 CFR 807.92 the following summary of information is provided:

# Owner/Contact/Date (807.92(a)(1):

Date: January 21, 2020

Owner/Submitter: GE Healthcare Finland Oy.

Kuortaneenkatu 2 00510 Helsinki FINLAND

Phone: +358 10 39411

Primary Contact Person: Joel Kent

Senior Regulatory Affairs Manager

GE Healthcare

Phone: 617-851-0943

E-mail: joel.kent@ge.com

Secondary Contact Person: Anssi Ruokonen

Regulatory Affairs Leader GE Healthcare Finland Oy

Kuortaneenkatu 2 00510 Helsinki

Finland

Phone: + 358 10 394 3686

E-mail: Anssi.ruokonen@ge.com

Device names (807.92(a)(2)):

Trade Name: CARESCAPE B850

Common/Usual Name: Multiparameter patient monitor (monitor, physiological, patient

(with arrhythmia detection or alarms)

- Classification Names: 21 CFR 870.1025 Arrhythmia detector and alarm (including STsegment measurement and alarm)
  - 21 CFR 868.1850 spirometer, monitoring (w/wo alarm)
  - 21 CFR 868.1730 computer, oxygen-uptake
  - 21 CFR 868.2375 monitor, breathing frequency
  - 21 CFR 868.2600 monitor, airway pressure (includes gauge and/or alarm)
  - 21 CFR 868.1500 analyzer, gas, enflurane, gaseous-phase (anesthetic conc.)
  - 21 CFR 868.1500 analyzer, gas, desflurane, gaseous-phase (anesthetic conc.)
  - 21 CFR 868.1500 analyzer, gas, sevoflurane, gaseous-phase (anesthetic conc.)
  - 21 CFR 868.1500 analyzer, gas, isoflurane, gaseous-phase (anesthetic conc.)
  - 21 CFR 868.1700 analyzer, gas, nitrous-oxide, gaseous phase (anesthetic conc.)
  - 21 CFR 868.1620 analyzer, gas, halothane, gaseous-phase (anesthetic conc.)
  - 21 CFR 868.1400 analyzer, gas, carbon-dioxide, gaseous-phase
  - 21 CFR 868.1720 analyzer, gas, oxygen, gaseous-phase
  - 21 CFR 870.2340 electrocardiograph
  - 21 CFR 870.2710 oximeter, ear
  - 21 CFR 870.2700 oximeter
  - 21 CFR 870.1425 computer, diagnostic, programmable
  - 21 CFR 870.2300 monitor, cardiac (incl. cardiotachometer & rate alarm)
  - 21 CFR 870.1025 detector and alarm, arrhythmia
  - 21 CFR 870.1100 alarm, blood-pressure
  - 21 CFR 870.1110 computer, blood-pressure
  - 21 CFR 870.1435 computer, diagnostic, pre-programmed, single-function
  - 21 CFR 870.1130 system, measurement, blood-pressure, noninvasive
  - 21 CFR 870.2910 thermometer, electronic, clinical
  - 21 CFR 882.1900 stimulator, auditory, evoked response
  - 21 CFR 882.1400 full-montage standard electroencephalograph
  - 21 CFR 868.2775 stimulator, nerve, peripheral, electric
  - 21 CFR 870.1915 probe, thermodilution
  - 21 CFR 870.1025 monitor, st segment with alarm
  - 21 CFR 882.1400 non-normalizing quantitative
  - electroencephalograph software
  - 21 CFR 882.1400 index-generating electroencephalograph software
  - 21 CFR 882.1400 reduced- montage standard electroencephalograph

21 CFR 882.1400 burst suppression detection software for electroencephalograph

Product Code: MHX

Subsequent Product Codes: BZK, BZL, BZQ, CAP, CBQ, CBR, CBS, CCK, CCL, DPS,

DPZ, DQA, DQK, DRT, DSI, DSJ, DSK, DXG, DXN, FLL, GWJ, GWQ, KOI, KRB, MLD, NHO, NHP, NHQ, OLT, OLW,

OMC, ORT

Predicate Device(s)

(807.92(a)(3): K131414 CARESCAPE Monitor B850

<u>Device Description</u> (807.92(a)(4)):

CARESCAPE B850 V3 is a new version of a modular multiparameter patient monitoring system. The monitor includes a separate 19-inch touchscreen display, the central processing unit (also called CPU) and the module frame F5 or F7. CARESCAPE B850 V3 is equipped with an "ePort" interface that supports use of PDM or CARESCAPE ONE patient data acquisition modules for patient monitoring. In addition, the F5 module frame has five module slots, and the F7 module frame has seven module slots where patient data acquisition modules (E-Module type), can be connected to perform patient monitoring. The CARESCAPE B850 V3 includes features and subsystems that are optional or configurable.

Intended Use: (807.92(a)(5):

Indications (from labeling)

The CARESCAPE B850 is a multi-parameter patient monitor intended for use in multiple areas within a professional healthcare facility.

The CARESCAPE B850 is intended for use on adult, pediatric, and neonatal patients and on one patient at a time.

The CARESCAPE B850 is indicated for monitoring of:

- hemodynamic (including ECG, ST segment, arrhythmia detection, ECG diagnostic analysis and measurement, invasive pressure, non-invasive blood pressure, pulse oximetry, cardiac output (thermodilution and pulse contour), temperature, mixed venous oxygen saturation, and central venous oxygen saturation),
- respiratory (impedance respiration, airway gases (CO2, O2, N2O, and anesthetic agents), spirometry, gas exchange), and
- neurophysiological status (including electroencephalography, Entropy, Bispectral Index (BIS), and neuromuscular transmission).

The CARESCAPE B850 can be a stand-alone monitor or interfaced to other devices. It can also be connected to other monitors for remote viewing and to data management software devices via a network.

The CARESCAPE B850 is able to detect and generate alarms for ECG arrhythmias: atrial fibrillation, accelerated ventricular rhythm, asystole, bigeminy, bradycardia, ventricular couplet, irregular, missing beat, multifocal premature ventricular contractions (PVCs), pause, R on T, supra ventricular tachycardia, tachycardia, trigeminy, ventricular bradycardia, ventricular fibrillation/ventricular tachycardia, ventricular tachycardia, and VT>2. The CARESCAPE B850 also shows alarms from other ECG sources.

The CARESCAPE B850 also provides other alarms, trends, snapshots and events, and calculations, and can be connected to displays, printers and recording devices.

The CARESCAPE B850 is intended for use under the direct supervision of a licensed healthcare practitioner, or by personnel trained in proper use of the equipment in a professional healthcare facility.

# Contraindications for using the monitor

The CARESCAPE B850 is not intended for use in a controlled MR environment.

Technology (807.92(a)(6)):

CARESCAPE B850 with CARESCAPE Software version 3 incorporates updated hardware and a new software platform.

The fundamental function and operation of the proposed CARESCAPE B850 V3 monitor are unchanged compared to CARESCAPE Monitor B850 with ESP V2 software (K131414). There are no new types of monitored parameters introduced compared to the predicate B850 monitor.

A summary of the main changes compared to the predicate are listed below in the comparison table.

# **Subject Device and Predicate Device Comparison**

Specification	CARESCAPE Monitor B850 with ESP V2 software (K131414)	CARESCAPE B850 with CSP software version V3	Differences
Patient type	Adult, pediatric & neonatal	Adult, pediatric & neonatal	Identical
Use environments		Within a professional healthcare facility (Not intended for MRI)	Identical
Size (H x W x D) & Weight	91 x 401 x 340 mm (3.6 x 15.8 x 13.5 in) and weight 7.5 kg (16.5 lbs)	91 x 401 x 340 mm (3.6 x 15.8 x 13.5 in) and weight 7.5 kg (16.5 lbs)	Identical
Processor	1.06GHz	Freescale ARM Cortex-A9	Equivalent The CARESCAPE Bx50 V3 monitors have an updated common CPU platform.
Module Housing		Up to 12 optional E-module slots with F5 and F7 Frames. One slide mount for acquisition module	Identical
Display/screen	19" Active matrix color TFT LCD	19" Active matrix color TFT LCD	Identical
Waveforms and parameter windows	individual waveforms and up to 20, if horizontal parameter	Standard view: Up to 8 individual waveforms and up to 20, if horizontal parameter area turned on.	Identical
Modules	CAiO, E-CAiOV, E-CAiOVX, E-COV, E-P, E-PP, E-PT, E-COP, E-COPSv, E-EEG, E-Entropy,	E-BIS, E-COP, E-COPSv, E-PiCCO, E-EEGX, E-Entropy, E-Masimo, E-miniC, E-NMT, E-NSATX, E-PP, E-PT, E-sCAiO, E-sCAiOVX, E-sCAiOVE, E-sCO, E-sCOV, E-sCOVX, PDM, CARESCAPE ONE	Equivalent Removed support for several legacy E/M multiparameter hemodynamic acquisition modules as many newer acquisition modules are included, covering the same parameters. Added support for CARESCAPE ONE and E-EEGX acquisition modules.  Added support for E-sCAiOVX and E-sCOVX modules (K150245).

Specification	CARESCAPE Monitor B850 with ESP V2 software (K131414)	CARESCAPE B850 with CSP software version V3	Differences
Available parameters	detection, ECG diagnostic analysis and measurement, invasive pressure, non-invasive blood pressure, pulse oximetry, cardiac output (thermodilution and pulse contour), temperature, mixed venous oxygen saturation, and central venous oxygen saturation, impedance respiration, airway gases (CO2, O2, N2O, and anesthetic agents), spirometry, gas exchange,	ECG, ST segment, arrhythmia detection, ECG diagnostic analysis and measurement, invasive pressure, non-invasive blood pressure, pulse oximetry, cardiac output (thermodilution and pulse contour), temperature, mixed venous oxygen saturation, and central venous oxygen saturation, impedance respiration, airway gases (CO2, O2, N2O, and anesthetic agents), spirometry, gas exchange, electroencephalography, Entropy, Bispectral Index (BIS), neuromuscular transmission.	Identical
EK-Pro arrhythmia detection algorithm	EK-Pro V13		Equivalent CARESCAPE B850 V3 uses an EK-ProV14 arrhythmia analysis algorithm compared to the EK-Pro V13 used in the predicate monitors
Graphical user interface	GE Healthcare Common User Interface (CUI) Requirements and Style Version 6		A new color scheme for the display and slight visual adjustments have been incorporated, in line with a GE Healthcare wide user interface design guideline. The overall user interface layout, structure, operations, and workflow remains the same as the predicate.
	networked laser printer. Printings for waveforms, alarms waveforms, numeric	Local recorder/printer and networked laser printer. Printings for waveforms, alarms waveforms, numeric trends	Identical

Alarms A fu	ystems  Alarm management core unctionalities:	systems  Alarm management core	Identical
Alarms A fu	ystems  Alarm management core unctionalities:	systems  Alarm management core	
fu	unctionalities:		Ei1 '
of A A	Adjustment of alarm settings Alarm On/Off functionality nd audio silencing	of alarms Adjustment of alarm settings Possibility to set critical alarm limits Alarm On/Off functionality and audio silencing	Equivalent The predicate already included the options to configure some alarm settings. In V3, more flexibility was added for the user to adjust alarm priorities and alarm criteria for additional parameter alarms, including additional alarm delay options, critical alarm options, alarm acknowledgement and latched alarm indicators, and a pause monitoring & central function. In general, options for tailored/specific alarm management schemes have been added supporting clinicians in their goals of reducing alarm fatigue while maintaining safety. Default setting are according to IEC 60601-1-8.
Specification C	CARESCAPE Monitor B850 with ESP V2 software	CARESCAPE B850 with CSP software version V3	Differences

	(K131414)		
Remote Alarm	Alarm Interface to the Remote	Alarm Interface to the Remote	Equivalent
Device	Alarm Box with Remote Light		CARESCAPE RAD
		RAD	replaces the Remote
	,		Alarm Box with Remote
			Light (RAB RL) device.
			The CARESCAPE RAD
			is the new accessory for
			use with the
			CARESCAPE Bx50 V3
			The proposed accessory
			communicates with the
			CARESCAPE Bx50 V3
			via a standard USB
			connection. The
			CARESCAPE RAD is
			intended for relaying
			1
			primary alarm signals from the host device to
			an external distributed
			alarm system, i.e. a
			nurse call system or a
			remote alarm light. The CARESCAPE Bx50 V3
			alarm functionality is
			unaffected when using
			the CARESCAPE RAD
			as the new accessory
			simply receives data
			from the monitor (one-
			way communication) to
			indicate when an alarm
			is active or not.
Networking		CARESCAPE Network	Equivalent
capability		LAN/VLAN	
		Single wire network	The single wire network
		configuration supported for	configuration simplifies
		CARESCAPE Networks	the installation and
			maintenance of
			CARESCAPE Bx50 V3
			patient monitors. The
			single wire network
			configuration has no
			impact on clinical
			monitoring.
NI A I C	101 7 1001 7	101 T 1001 T	
Network interface	Tubase I, 100base I	10baseT, 100baseT	Identical

<u>Determination of</u> <u>Substantial Equivalence</u> (807.92(b)(1):

Summary of Non-Clinical Tests:

Bench testing related to software, hardware and performance including applicable consensus standards was conducted on the

CARESCAPE B850, demonstrating the design meets the specifications.

The hardware bench testing included electromagnetic compatibility, electrical safety, environmental and usability.

The CARESCAPE B850 has been found to substantially equivalent to the predicate device(s) for the intended users, uses and use environments. Extensive usability work has been completed for CARESCAPE B850 and the predicate devices including critical task identification through use-based hazard analysis, multiple rounds of formative usability testing and summative testing, among other activities.

Software testing included software design, development, verification, validation and traceability.

Patient safety, security, and privacy risks have been addressed in the design and development of CARESCAPE B850 including a Security Risk Assessment and Threat model. This includes system integrity controls, access controls, audit controls, network controls, and remote service controls which map to the General Principles and Security Capabilities outlined in the FDA Guidance for Content of Premarket Submissions for Management of Cybersecurity in Medical Devices Guidance for Industry and Food and Drug Administration Staff Document issued on October 2, 2014.

Testing was completed on CARESCAPE B850 to show the device can withstand network storm, i.e. continue to monitor patients without rebooting when connected by wire to either MC, IX, or both networks which suffer broadcast storm traffic.

Clinical (807.92(b)(2)): Summary of Clinical Tests:

The subject of this premarket submission, CARESCAPE B850 did not require clinical studies to support substantial equivalence.

Conclusion (807.92(b)(3)): GE Healthcare considers the CARESCAPE B850 to be substantially equivalent to the predicate device(s).