



October 28, 2020

L&K BIOMED Co., Ltd.
KiHyang Kim
Official Correspondent
#201, 202 16-25, Dongbaekjungang-ro 16 beon-gil,
Giheung-gu
Yongin-si, Gyeonggi-do 17015
Republic Of Korea

Re: K200790

Trade/Device Name: LnK Spinal Fixation System, OpenLoc-L Spinal Fixation System
Regulation Number: 21 CFR 888.3070
Regulation Name: Thoracolumbosacral Pedicle Screw System
Regulatory Class: Class II
Product Code: NKB, KWQ, KWP
Dated: September 28, 2020
Received: September 28, 2020

Dear KiHyang Kim:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin O'Neill, M.B.E.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K200790

Device Name

LnK Spinal Fixation System
OpenLoc-L Spinal Fixation System

Indications for Use (Describe)

Spinal Fixation System is non-cervical spinal fixation devices intended for use as posterior pedicle screw fixation systems (T1-S2/ilium), or as an anterolateral fixation system (T8-L5). All components in the system are limited to skeletally mature patients. These devices are intended as an adjunct to fusion for all of the following indications: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis); tumor; stenosis, and failed previous fusion (pseudoarthrosis)

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) SUMMARY

The following 510(k) summary is being submitted as required by 21 CFR 807.92(a):

1. Manufacturer

Submitter:	L&K BIOMED Co., Ltd.
	#201, 202 16-25, Dongbaekjungang-ro 16 beon-gil, Giheung-gu, Yongin-si, Gyeonggi-do, 17015, Korea
	Phone. +82-10-5477-0325
Contact Person:	KiHyang Kim e-mail: khkim@lnkbiomed.com

2. Device Identification

Trade Name	LnK Spinal Fixation System OpenLoc-L Spinal Fixation System
Classification Name	Thoracolumbosacral pedicle screw system
Class	Class II
Product Code	NKB, KWP, KWQ
Common Name	Thoracolumbosacral pedicle screw system Spinal interlaminar fixation orthosis Spinal intervertebral body fixation orthosis
Regulation No.	21 CFR 888.3070
Panel	Orthopedic

3. Predicate or legally marketed devices which are substantially equivalent

The additional components of the LnK Spinal Fixation System and OpenLoc-L Spinal Fixation System are considered substantially equivalent to the predicate devices. The systems have same design, materials, scientific technology, and indications for use.

Primary Predicate:

LnK Spinal Fixation System (K143363)

Additional Predicates:

LnK Spinal Fixation System and OpenLoc-L Spinal Fixation System (K171813, K183168)

LnK Spinal Fixation System (K120270)

4. Description of the Device

The Spinal Fixation System is available in various lengths. This system is comprised of screws, set screws, rods, crosslinks, connectors and hooks. The screws are available from 4.0 mm to 10.5 mm diameters with lengths ranging from 20 mm to 150 mm. The rods are available from 5.0 mm, 5.5 mm, 6.0 mm and 6.35 mm diameter with lengths ranging from 40 mm to 600 mm. Both straight rods and curved rods have four types of design that consist of standard type, hex type, stopper type and double stopper.

5. Materials

The LnK Spinal Fixation System and OpenLoc-L Spinal Fixation System are manufactured from Ti-6Al-4V ELI titanium alloy (ASTM F136) or Cobalt-28Chromium-6Molybdenum-4Vanadium ELI (ASTM F1537). This this is the same material used in the predicate devices.

6. Indication for Use

Spinal Fixation System is non-cervical spinal fixation devices intended for use as posterior pedicle screw fixation systems (T1-S2/ilium), or as an anterolateral fixation system (T8-L5). All components in the system are limited to skeletally mature patients. These devices are intended as an adjunct to fusion for all of the following indications: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis); tumor; stenosis, and failed previous fusion (pseudoarthrosis)

7. Performance Testing

The additional components to be added through this submission do not require additional mechanical testing. None of the additional components is the worst case of the LnK Spinal Fixation System and OpenLoc-L Spinal Fixation System. Therefore, we substitute mechanical test data of additional components of LnK Spinal Fixation System and OpenLoc-L Spinal Fixation System with the predicate device (K120270, K143363, K171813, K183168).

8. Summary of Technology Characteristics

LnK Spinal Fixation System and OpenLoc-L Spinal Fixation System are substantially equivalent to the predicate devices in terms of design, materials and indications for use.

9. Substantial Equivalence

LnK Spinal Fixation System and OpenLoc-L Spinal Fixation System were shown to be substantially equivalent to the predicate devices in indications for use, design, function and materials used.

10. Conclusion

The additional components of LnK Spinal Fixation System and OpenLoc-L Spinal Fixation System perform as well as the predicate devices. Therefore, the additional components are substantially equivalent to the predicate devices (K120270, K143363, K171813 and K183168).