

July 14, 2022

Church & Dwight Co., Inc. % Dawn Reilly-O'Dell Principal Full Circle Regulatory Consulting, LLC 107 Casablanca Court Cary, NC 27519

Re: K221431

Trade/Device Name: TROJANTM Her Pleasure Warming male natural rubber latex condom with

warming lubricant

Regulation Number: 21 CFR§ 884.5300

Regulation Name: Condom

Regulatory Class: II Product Code: HIS Dated: May 12, 2022 Received: May 17, 2022

Dear Dawn Reilly-O'Dell:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to https://www.fda.gov/medical-device-problems.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance) and CDRH Learn (https://www.fda.gov/training-and-continuing-education/cdrh-learn). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice">https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Jason R. Roberts, Ph.D.
Assistant Director
DHT3B: Division of Reproductive,
Gynecology and Urology Devices
OHT3: Office of GastroRenal, ObGyn,
General Hospital and Urology Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration

Indications for Use

510(k) Number (if known)

Form Approved: OMB No. 0910-0120

Expiration Date: 06/30/2023 See PRA Statement below.

K221431	
Device Name TROJAN™ Her Pleasure Warming male natural rubber latex condom with warming lubricant	
Indications for Use (Describe) The TROJAN TM Her Pleasure Warming natural rubber latex condom v for prophylactic purposes (to prevent pregnancy and the transmission of	
Type of Use (Select one or both, as applicable) Prescription Use (Part 21 CFR 801 Subpart D)	Over-The-Counter Use (21 CFR 801 Subpart C)
CONTINUE ON A SEPARATE PAGE IF NEEDED.	

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer Paperwork Reduction Act (PRA) Staff PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Submitter Name: Church & Dwight Co., Inc.

Submitter Address: 500 Charles Ewing Boulevard, Ewing, NJ 08628

Submitter Contact Person: Supreet Sahota-Bhatti Sr. Manager, Global Regulatory Affairs (US)

Church & Dwight Co., Inc., 469 North Harrison Street, Princeton, NJ 08543

Tel: (609)806-7893

Preparer Name: Dawn Reilly-O'Dell, RAC, MPH Principal, Full Circle Regulatory Consulting, LLC

Email: dreilly@fullcirclereg.com

Date Prepared: May 12, 2022

Trade Name: TROJAN™ Her Pleasure Warming male natural rubber latex condom with

warming lubricant

Common Name: Natural Rubber Latex Condom

Product Code: HIS (condom)

Regulatory Class II

Classification Name: Condom (21 CFR § 884.5300)

Predicate Device: K131887: TROJAN™ Jaguar Male Natural Rubber Latex Condom with Warming

and Tingling Lubricant

The predicate device has not been subject to a design related recall.

Reference Devices: K120249: TROJAN™ MAGNUM Male Latex Condom with Warming Lubricant

K092586: Male Natural Latex Condom

Description of Device:

The TROJAN™ Her Pleasure Warming natural rubber latex condom with warming lubricant is made of a natural rubber latex sheath, which completely covers the penis with a fitted membrane and has a warming lubricant applied directly to the condom to create a perception of warming. The condom is a bulbous shaped condom with 11 rows of continuous ribs on the shaft and 9 rows of continuous ribs on the bulb. The TROJAN™ Her Pleasure Warming natural rubber latex condom with warming lubricant meets the specifications of ASTM D3492-16. The condom has a nominal length of 185 mm and a nominal flat width of 54 mm, measured 30 mm from the open end. The bulbous part at the closed end of the condom has a flat width of 65 mm.

Intended Use of the Device:

The TROJAN™ Her Pleasure Warming natural rubber latex condom with warming lubricant is used for contraception and for prophylactic purposes (to prevent pregnancy and the transmission of sexually transmitted infections).

Predicate Device Comparison:

The subject and predicate device have the same intended use and indications for use. The subject and predicate device share the same product code, indications for use, design, shape, dimensions, shelf-life, and primary packaging. Both are lubricated natural rubber latex condoms, but the formulations of their natural rubber latex and lubricant differ. These differences in technological characteristics do not raise different questions of safety and effectiveness, and both the natural rubber latex and lubricant formulations of the subject device are the same as the natural rubber latex and lubricant formulations of legally marketed, 510(k) cleared reference devices.

Summary of Performance Testing

Physical Testing Data: Three (3) lots of TROJAN™ Her Pleasure Warming natural rubber latex condom with warming lubricant were tested at baseline and met airburst specifications of ASTM D3492-16 Standard Specifications for Rubber Contraceptives (Male Condoms).

Shelf Life: Stability of the TROJAN™ Her Pleasure Warming natural rubber latex condom with warming lubricant was established from results of physical testing data using a protocol that followed 21 CFR §801.435 and met the requirements of both ASTM D3492-16 and ISO 4074:2015. Based on the evaluation of the results of the physical testing data, the expiration date has been initially set at 36 months (about 3 years) and will be then verified through real-time stability through five (5) years in compliance with FDA expiration labeling requirements in 21 CFR §801.435.

Summary of Safety Testing

Biocompatibility: Biocompatibility evaluations were performed on the TROJAN™ Her Pleasure Warming natural rubber latex condom with warming lubricant in accordance with ISO 10993-1:2018, Biological Evaluation of Medical Devices.

Table 1: Biocompatibility Evaluations Conducted

Evaluation (biological endpoint)	ISO Standard
Cytotoxicity ISO Elution Method	ISO 10993-5
Rabbit Vaginal Irritation	ISO 10993-10
Rabbit Penile Irritation	ISO 10993-10
Primary Rabbit Skin Irritation	ISO 10993-10
Guinea Pig Maximization Sensitization	ISO 10993-10
Acute Systemic Toxicity	ISO 10993-11

The results of these evaluations demonstrate that the subject device is biocompatible.

Conclusion:

Based on the results of the testing and data described above, the TROJAN™ Her Pleasure Warming natural rubber latex condom with warming lubricant is as safe and effective as the predicate device. The performance data support a determination of substantial equivalence.