



August 26, 2020

SeaSpine Orthopedics Corporation  
Aly Alvarez  
Sr. Regulatory Affairs Specialist  
5770 Armada Drive  
Carlsbad, California 92008

Re: K202064

Trade/Device Name: Admiral ACP System  
Regulation Number: 21 CFR 888.3060  
Regulation Name: Spinal Intervertebral Body Fixation Orthosis  
Regulatory Class: Class II  
Product Code: KWQ  
Dated: July 24, 2020  
Received: July 27, 2020

Dear Ms. Alvarez:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin O'Neill, M.B.E.  
Acting Assistant Director  
DHT6B: Division of Spinal Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K202064

Device Name

Admiral ACP System

Indications for Use (Describe)

The Admiral ACP System is intended for anterior cervical fixation (C2-T1) for the following indications:

- Degenerative Disc Disease (DDD) (defined as neck pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies),
- Spondylolisthesis,
- Trauma (i.e., fracture or dislocation),
- Spinal Stenosis,
- Deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis),
- Tumor,
- Pseudoarthrosis,
- and failed previous fusion.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

### Contact Details

Applicant Name: SeaSpine Orthopedics Corporation

Address: 5770 Armada Drive, Carlsbad CA

Phone number: (619) 884-4342 (Aly) or (949) 855-7175 (Cindy)

Fax number: (760) 683-6874

Contact Person: Aly Alvarez, Sr. Regulatory Affairs Specialist

Additional Contact: Cindy Fong, Regulatory Affairs Coordinator

Date Prepared: July 24, 2020

### Device Name

Trade Name: Admiral ACP System

Common Name: Cervical Plate

Classification Name: Spinal Intervertebral Body Fixation Orthosis (21 CFR 888.3060)

Class: II

Product Code: KWQ

### Legally Marketed Predicate Devices

510(k) Number	Product Code	Trade Name	Manufacturer
<b>Primary Predicate Device</b>			
K173521	KWQ	SeaSpine Cabo ACP (Anterior Cervical Plate) System	SeaSpine Orthopedics Corporation
<b>Additional Predicate Device</b>			
K112206	KWQ	Cardiff Anterior Cervical Plate System	SeaSpine, Inc. (formerly known as Integra LifeSciences Corporation)

## Device Description

The Admiral ACP System consists of a variety of bone plates and screws. Fixation is achieved by inserting bone screws through the openings in the plate into the vertebral bodies of the cervical spine. The Admiral plates include locking pins that cover the heads of the bone screws to reduce the potential for screw back-out. The locking pins come preassembled to the plate. The Admiral ACP System implant components are made from titanium alloy per ASTM F136. Associated instruments are available to facilitate the implantation of the device.

## Indications for Use

The Admiral ACP System is intended for anterior cervical fixation (C2-T1) for the following indications:

- Degenerative Disc Disease (DDD) (defined as neck pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies),
- Spondylolisthesis,
- Trauma (i.e., fracture or dislocation),
- Spinal Stenosis,
- Deformities or curvatures (i.e., scoliosis, kyphosis, and/or lordosis),
- Tumor,
- Pseudoarthrosis,
- and failed previous fusion.

## Summary of Technological Characteristics

The Admiral ACP System is equivalent to the cited predicate devices in regard to components, device description, intended use/indications for use, technological characteristics (operating principle, design, materials, sterility, manufacturing, etc.) and performance (mechanical).

The implants are used to treat the same conditions, have essentially the same precautions and contraindications for use, and represent a basic design concept in terms of safety and effectiveness, and differ only in design details and not functionality.

## Non-Clinical Testing

The Admiral ACP System demonstrated equivalent performance to the predicate system through mechanical testing in static axial compression bending and torsion, and dynamic axial compression bending per ASTM F1717.

## Conclusions

The submitted data demonstrate that the Admiral ACP System is substantially equivalent to the cited legally marketed predicate device.