



April 11, 2022

OrthoFundamentals, LLC  
Howard Schrayer  
Regulatory Consultant  
8 Lookout  
Hilton Head Island, Massachusetts 29928

Re: K212962

Trade/Device Name: TRELLIS™ SI Joint Fusion System  
Regulation Number: 21 CFR 888.3040  
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener  
Regulatory Class: Class II  
Product Code: OUR  
Dated: March 8, 2022  
Received: March 9, 2022

Dear Howard Schrayer:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin O'Neill, M.B.E.  
Assistant Director  
DHT6B: Division of Spinal Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## INDICATIONS FOR USE

DEPARTMENT OF HEALTH AND HUMAN SERVICES Food and Drug Administration <b>Indications for Use</b>	Form Approved: OMB No. 0910-0120 Expiration Date: 06/30/2023 See PRA Statement below.
---	---

510(k) Number (if known)  
K212962

Device Name  
TRELIS™ SI Joint Fusion System

Indications for Use (Describe)

The OrthoFundamentals TRELIS™ SI Joint Fusion and Stabilization Implant System is intended for fixation of sacroiliac joint disruptions, and intended for sacroiliac joint fusion for conditions including:

Sacroiliac joint disruptions

Degenerative sacroiliitis

To augment immobilization and stabilization of the sacroiliac joint in skeletally mature patients undergoing sacropelvic fixation as part of a lumbar or thoracolumbar fusion and

Acute, non-acute, and non-traumatic fractures involving the sacroiliac joint.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)  Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASstaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**510(k) SUMMARY  
(PER 21 CFR 807.92)**

**General Company Information**

Name: OrthoFundamentals, LLC  
Contact: Howard Schrayer  
Address: 173 Governors Avenue  
Medford, MA 02155  
Telephone: 609.273.7350

**Date Prepared:** April 11, 2022

**General Device Information**

Product Name: TRELIS™ SI Joint Fusion System  
Classification: Smooth Or Threaded Metallic Bone Fixation Fastener  
Product code: OUR – Class II  
21 CFR 888.3040

**Predicate Device**

Primary Predicate

OrthoFix Firebird SI  
510(k) K203138

Additional Predicates

Synthes 6.5mm Cannulated Screw  
510(k) K029132

Keystone Dental Genesis Implant System  
510(k) K101545

SI Bone iFuse Implant System  
510(k) K193524

Globus SI-LOK Sacroiliac Joint Fixation System  
510(k) K183119

Zyga Technology, Inc Slmmetry Sacroiliac Joint Fusion System  
510(k) K141549

## Device Description

The TRELLIS™ SI Joint Fusion System is a multiple component system consisting of sterile single use accessory instruments and sterile individually packaged screws of various lengths and diameters. The screws are designed to compress and stabilize the sacroiliac joint during the fusion process. They are self-tapping and are designed with a porous mid-shaft region to allow bone to grow onto and into its surface. The cannulated screws have axial fenestrations that allow bone through growth and packing with autograft bone or allograft material. The screws are designed to promote bony fusion onto and into the implant through the macro-surface topography which is generated through EBM additive manufacturing. The fenestration geometry harvests autograft during insertion which helps create an environment for bony in-growth and on-growth.

## Indications for Use

The OrthoFundamentals TRELLIS™ SI Joint Fusion System is indicated for fixation of sacroiliac joint disruptions, and intended for sacroiliac joint fusion for conditions including:

- sacroiliac joint disruptions
- degenerative sacroiliitis
- to augment immobilization and stabilization of the sacroiliac joint in skeletally mature patients undergoing sacropelvic fixation as part of a lumbar or thoracolumbar fusion and
- acute, non-acute, and non-traumatic fractures involving the sacroiliac joint.

## Substantial Equivalence Bench Testing

A series of laboratory studies (bench tests) have been conducted to verify the suitability of the TRELLIS™ SI Joint Fusion System for its intended use, establish Substantial Equivalence with the predicate devices and confirm the reproducibility of packaging.

These tests include:

Test	Finding
Screw Pull-Out Force Testing	Substantially Equivalent
Insertion and Removal Torque Testing	Pass
Torsion Properties Testing	Substantially Equivalent
Static and Fatigue Bend Testing	Substantially Equivalent
Package Seal Strength Verification	Pass
Surface Characterization	Substantially Equivalent
Biocompatibility Justification	Pass
Component Testing Plan and Justification for Test Coupon Location	Pass
Porous Surface Lattice Abrasion Resistance	Pass
Porous Surface Lattice Stereological Evaluation	Pass
Porous Surface Lattice Mechanical Testing	Pass

## Similarities and Differences

**OrthoFundamentals**  
**TRELLIS™ SI Joint Fusion**  
**System**  
**510(k) TBD**

**Synthes**  
**6.5mm Cannulated Screw**  
**510(k) K021932**

**OrthoFix**  
**Firebird SI**  
**510(k) K203138**

### Similarities

Indicated for fixation of sacroiliac joint disruptions, and intended for sacroiliac joint fusion

Indicated for SI joint disruptions

Indicated for fixation of sacroiliac joint disruptions, and intended for sacroiliac joint fusion

Product Code:  
Class II  
21 CFR 888.3040

Product Code:  
Class II  
21 CFR 888.3040

Product Code:  
Class II  
21 CFR 888.3040

Prescription use in OR setting

Prescription use in OR setting

Prescription use in OR setting

Cannulated screw-based implant for implantation over a k-wire

Cannulated screw-based implant for implantation over a k-wire

Cannulated screw-based implant for implantation over a k-wire

Implant uses threads for primary means of fixation

Implant uses threads for primary means of fixation

Implant uses threads for primary means of fixation

Implant is available in a variety of lengths and diameters

Implant is available in a variety of lengths

Implant is available in a variety of lengths and diameters

Device is fabricated from Titanium-6Aluminum-4Vanadium ELI per ASTM F3001-14

Device is fabricated from stainless steel or titanium alloy

Device is fabricated from Titanium-6Aluminum-4Vanadium ELI per ASTM F3001-14

Device is additively manufactured

Device is traditionally machined

Device is additively manufactured

Device is available in 9mm and 11mm diameters

Device is available in 6.5mm diameters

Device is available in 9mm, 11mm and 12mm diameters

Device is available in lengths 30mm to 70mm

Device is available in lengths 30mm to 200mm

Device is available in lengths 25mm to 70mm

Device is provided with single-use, sterile accessory delivery tools

Device provided with non-sterile accessories

Device provided with non-sterile accessories that may be resterilized (9mm screw is also non-sterile)

Device is provided sterile by gamma radiation and stored at room temperature for single patient use

Device is provided non-sterile

Device is provided in sterile and non-sterile configurations

Initial shelf-life of 1 year based on packaging validations

No shelf-life provided

No shelf-life provided

### **Performance Data**

No animal or clinical testing has been conducted.

This submission supports the position that the OrthoFundamentals TRELIS™ SI Joint Fusion System is substantially equivalent to the Primary and Additional predicates listed above. A number of other predicate devices list the same range of clinical uses. See above for specific non-clinical testing conducted to demonstrate substantial equivalence of the TRELIS™ SI Joint Fusion System.

### **Conclusions**

OrthoFundamentals, LLC believes that the information provided establishes that similar legally marketed devices have been used for the same clinical application as the TRELIS™ SI Joint Fusion System. The device is substantially equivalent to the referenced predicates.