



January 12, 2022

GA Health Company Limited  
Cindy Ye  
Chief Executive Officer  
Unit 18, 21/F, Metropole Square, 2 On Yiu Street, Shatin  
Hong Kong,  
CHINA

Re: K213095  
Trade/Device Name: andorate Suction Valve, andorate Air/Water Valve  
Regulation Number: 21 CFR 876.1500  
Regulation Name: Endoscope and accessories  
Regulatory Class: Class II  
Product Code: ODC  
Dated: December 13, 2021  
Received: December 16, 2021

Dear Cindy Ye:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

*for*

Shanil P. Haugen, Ph.D.  
Assistant Director  
DHT3A: Division of Renal, Gastrointestinal,  
Obesity and Transplant Devices  
OHT3: Office of GastroRenal, ObGyn,  
General Hospital and Urology Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K213095

Device Name

andorate® Suction Valve  
andorate® Air/Water Valve

Indications for Use (Describe)

The single use andorate® Suction Valve is used to control the suction function of an endoscope during GI endoscopic procedures.

The single use andorate® Air/Water Valve is used to control the air / water function of an endoscope during GI endoscopic procedures.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

The 510(k) summary of safety and effectiveness information is being submitted in accordance with the requirements of the Safe Medical Devices Act of 1990 and 21 CFR 807.92.

### 1. Submission Sponsor

Submitter's Name: GA Health Company Limited  
Submitter's Address: Unit 18, 21/F, Metropole Square  
2 On Yiu Street, Shatin, N.T,  
Hong Kong, CHINA  
Establishment Registration No.: 3014749926

### 2. Sponsor Contact

Cindy Ye  
Chief Executive Officer  
Telephone: +852 2833 9010  
Email: cindy@gahealth.net

Lam Wing Yu  
Senior Regulatory Affairs Specialist  
Telephone: +852 2833 9010  
Email: rainy@gahealth.net

### 3. Date Prepared

16 November 2021

### 4. Device Identification

Device Name: andorate® Suction Valve  
Common Name: Suction valve for endoscope  
Classification Number: 21 CFR 876.1500  
Classification Name: Endoscope and accessories  
Product Code: ODC  
Product Code Name: Endoscope channel accessory  
Regulation Class: 2  
Device Panel: Gastroenterology/Urology

Device Name: andorate® Air/Water Valve  
Common Name: Air/Water valve for endoscope  
Classification Number: 21 CFR 876.1500  
Classification Name: Endoscope and accessories  
Product Code: ODC  
Product Code Name: Endoscope channel accessory  
Regulation Class: 2  
Device Panel: Gastroenterology/Urology

**5. Predicate Device Identification**

Predicate Device 510(k) No.: K180341  
 Predicate Device Trade Name: FUJIFILM 600 Series Endoscope EG-600WR v2  
 Predicate Device Product Code: FDS – Endoscope and Accessories

**6. Device Description:**

The subject devices are intended for single-use and are supplied sterile. Table 1 shows the components included in the submission.

Table 1 – Components included in the Submission

Components	Qty	Product Code	Regulation Number	Regulatory Classification
andorate® Suction Valve	1	ODC – Endoscope channel accessory	21 CFR 876.1500	2
andorate® Air/Water Valve	1	ODC – Endoscope channel accessory	21 CFR 876.1500	2

The suction valve is designed to be attached to the suction port of the endoscope and the air/water valve is designed to be attached to the air/water port of the endoscope. The activation of the suction valve allows the user to aspirate excess fluids or other debris obscuring the endoscope image, while the activation of the air/water valve allows the user to control air and water flow to assist in cleansing the lens during procedures.

There were no prior submissions for the subject devices.

**7. Intended Use:**

The single use andorate® Suction Valve is used to control the suction function of an endoscope during GI endoscopic procedures.

The single use andorate® Air/Water Valve is used to control the air / water function of an endoscope during GI endoscopic procedures.

**8. Technological Characteristics**

Table 2 summarizes the suction valve and air/water valve technological characteristics as compared to the predicate devices.

Table 2 – Summary of design, features and principles of operation and technological characteristics between the subject device and predicate devices

Specification	Predicate Device	Subject Device	Substantial Equivalence
Product code	FDS	ODC	Substantial Equivalent
Regulatory Classification	2	2	Identical
Regulation No	21 CFR 876.1500	21 CFR 876.1500	Identical
Regulation Description	Endoscope and accessories	Endoscope and accessories	Identical
Intended Use	<p>This product is the suction button used in combination with our endoscopes in medical facilities. Do not use this product for any other purpose.</p> <p>This product is the air/water button used in combination with our endoscopes in medical facilities. Do not use this product for any other purpose.</p>	<p>The single use andorate® Suction Valve is used to control the suction function of an endoscope during GI endoscopic procedures.</p> <p>The single use andorate® Air/Water Valve is used to control the air / water function of an endoscope during GI endoscopic procedures.</p>	Substantial Equivalent
Environment of Use	Hospital and or clinics	Hospital and or clinics	Identical
Single Use or Reusable	Reusable	Single Use	Substantial Equivalent
Material	Fluorine Resin, Stainless Steel, Silicone, Thermoplastics	Polycarbonate, Acrylonitrile Butadiene Styrene, Thermoplastic Elastomer, Silicone Rubber, Stainless Steel	Substantial Equivalent
Manufacturing method	Injection molding, overmolding	Injection molding, compression molding, overmolding	Substantial Equivalent
Packaging	Packaged in a plastic bag	Packaged in a sealed pouch	Substantial Equivalent
Sterilization	Supplied non-sterile Sterilized after each use	Yes	Substantial Equivalent
Shelf Life	1 year	1 year	Substantial Equivalent

## 9. Performance Test

The bench testing was performed to support substantial equivalence on both the subject device and the predicate device. The performance data demonstrated that the subject devices met established specifications in the following non-clinical tests.

### Suction Valve

1. Endoscope Compatibility Test
2. Suction Flow Test
3. Vacuum Leak Test
4. Pressing Force Test

#### Air/water Valve

1. Endoscope Compatibility Test
2. Air Flow Test
3. Water Flow Test
4. Water Leakage Test
5. Pressing Force Test
6. Backflow Prevention Test

#### **10. Biocompatibility**

The biocompatibility of the subject devices was conducted in accordance with the FDA guideline "Use of International Standard ISO 10993-1, "Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process". Biocompatibility testing is conducted on subject device in accordance with the ISO 10993 standard. The test result shows that both the subject devices are biocompatible.

#### **11. Conclusion**

The subject devices have the same intended use as the predicate devices. Based on the technological characteristics and overall performance of the devices in bench testing, there are no significant differences exist between the subject devices and the predicate devices. The subject devices do not raise any new issues of safety and effectiveness. From a clinical perspective and comparing design specifications, the subject devices and the predicate device are substantially equivalent.