



September 14, 2022

SprintRay Inc
Sara Moghtadernejad
Regulatory Affairs Manager
2705 Media Center Drive, Suite 100A
Los Angeles, California 90065

Re: K220979

Trade/Device Name: SprintRay Denture Base
Regulation Number: 21 CFR 872.3760
Regulation Name: Denture Relining, Repairing, Or Rebasing Resin
Regulatory Class: Class II
Product Code: EBI
Dated: August 16, 2022
Received: August 16, 2022

Dear Sara Moghtadernejad:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For Michael E. Adjodha, M.ChE.
Assistant Director
DHT1B: Division of Dental and
ENT Devices
OHT1: Office of Ophthalmic, Anesthesia,
Respiratory, ENT and Dental Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K220979

Device Name

SprintRay Denture Base

Indications for Use (Describe)

SprintRay Denture Base resin is a light-curable polymerizable resin intended to be used for the fabrication and repair, of full and partial removable dentures and baseplates. The material is an alternative to traditional denture base material.

Fabrication of dental prosthetics with SprintRay Denture Base resin requires computer-aided design and manufacturing system that includes the following components not part of the device: oral casting impression, digital denture base file created in an optical impression system, 3D printer, and curing light equipment.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

***DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW. ***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human
Services Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY-K220979

SprintRay Denture Base

Submitter: SprintRay Inc.
2705 Media Center Drive, Suite 100A
Los Angeles, CA 90065

Phone:

Contact Person: Sara Moghtadernejad

Date Prepared: March 31, 2022

Name of Device: SprintRay Denture Base

Common or Usual Name: Denture Base, Prescription

Classification Name: Denture Base, Prescription

Regulatory Class: Class II

Regulation Number: 21 CFR 872.3760

Product Code: EBI

Predicate Devices

Dentca Denture Base (K143033)

Device Description

SprintRay Denture Base consists of a curable dental acrylate resin that is designed to be used in conjunction with a scanned 3D image of a patient's teeth, and 3D printer assembly, to locally manufacture out a dental appliance in dental offices based on the clinician's judgment of patient need.

SprintRay Denture Base resin is intended exclusively for professional dental work. SprintRay Denture Base resin is offered in following shades/colors:

- Light Pink
- Original Pink
- Light Meharry
- Original Meharry
- Extra Light Pink
- Medium Pink
- Dark Pink

- Dark Meharry
- Deep Dark Meharry

SprintRay Denture Base is designed to meet appropriate ISO standards for flexibility, sorption, and solubility to withstand prolonged use in the oral cavity. It is delivered non-sterile, and instructions are provided on cleaning the material prior to providing it to a patient. Curing is performed with a UV lamp. The appliance is then cleaned, trimmed, and verified to fit in the dental office before the patient leaves.

Intended Use / Indications for Use

SprintRay Denture Base resin is a light-curable polymerizable resin intended to be used for the fabrication and repair, of full and partial removable dentures and baseplates. The material is an alternative to traditional denture base material.

Fabrication of dental prosthetics with SprintRay Denture Base resin requires computer-aided design and manufacturing system that includes the following components not part of the device: oral casting impression, digital denture base file created in an optical impression system, 3D printer, and curing light equipment.

Summary of Technological Characteristics

Light-Denture Based curing of a 3D printed acrylate resin is the technological principle for both the subject and predicate devices. SprintRay Denture Base resin is poured into a 3D printer, which relies on scanned images of the patient's oral cavity to produce a dental appliance. At a high level, the subject and predicate devices are based on the following same technological elements:

- Are a pourable acrylate resin
- Are used in conjunction with 3D printers, which rely on common 3D images to define the fabricated dental appliance
- Are cured prior to final trimming and cleaning
- Are used for the fabrication of orthodontic and dental appliances

The following technological differences exist between the subject and predicate devices:

- Differences in acrylate resin material

Performance Data

The following performance data were provided in support of the substantial equivalence determination.

Biocompatibility Testing

The biocompatibility evaluation for SprintRay Denture Base was conducted in accordance with the FDA Blue Book Memorandum #G95-1 and International Standard ISO 10993-1 and ISO 7405, as recognized by FDA. The battery of testing included the following tests:

- Genotoxicity
- Cytotoxicity
- Acute Systematic Toxicity
- Sensitization
- Irritation

SprintRay Denture Base is considered tissue contacting for a period longer than 30 days (a removable prosthesis).

Bench Testing

SprintRay Denture Base was tested for conformity with the industry consensus standard ISO 20795-1. The battery of testing included the following tests:

- Flexural Strength and Modulus
- Water Sorption and Solubility
- Stability
- Residual Methyl Methacrylate Monomers
- Homogeneity
- Surface Characteristics
- Shape Capability, Translucency, and Polishability
- Freedom from Porosity
- Color Stability

In all instances, SprintRay Denture Base functioned as intended and the outcomes were as expected.

Conclusions

SprintRay Denture Base resin is as safe and effective as its predicate device. SprintRay Denture Base has the same intended use and indication, and similar technological characteristics, and principles of operation as its predicate device. The minor technological differences between SprintRay Denture Base and its predicate devices raise no new issues of safety or effectiveness. Performance data demonstrate that SprintRay Denture Base is as safe and effective as the predicate device. Thus, SprintRay Denture Base is substantially equivalent.