



July 15, 2022

Canon Inc.  
% Saori Sawaki  
Business Manager, Regulatory Consultant  
Ken Block Consulting  
800 E. Campbell Road, Suite 202  
RICHARDSON TX 75081

Re: K221876

Trade/Device Name: Digital Radiography CXDI-Pro, Digital Radiography D1  
Regulation Number: 21 CFR 892.1680  
Regulation Name: Stationary x-ray system  
Regulatory Class: Class II  
Product Code: MQB  
Dated: June 27, 2022  
Received: June 28, 2022

Dear Saori Sawaki:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Laurel Burk, Ph.D.  
Assistant Director  
Diagnostic X-Ray Systems Team  
OHT 8B: Division of Radiological Imaging Devices and  
Electronic Products  
OHT8: Office of Radiological Health  
Office of Radiological Health  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K221876

Device Name

DIGITAL RADIOGRAPHY CXDI-Pro

DIGITAL RADIOGRAPHY D1

Indications for Use (Describe)

The DIGITAL RADIOGRAPHY CXDI-Pro / D1 provides digital image capture for conventional film/screen radiographic examinations. This device is intended to capture, for display, radiographic images of human anatomy, and to replace radiographic film/screen systems in all general purpose diagnostic procedures. This device is not intended for mammography applications.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**510(k) SUMMARY**

Applicant/ Sponsor: Canon Inc.  
9-1 Imaikami-cho  
Nakahara-ku, Kanagawa 211-8501 JAPAN

Contact Person: Mr. Akira Hirai  
General Manager  
TEL: 81-3-3758-2111  
FAX: 044-739-6695  
[hirai.akira@mail.canon](mailto:hirai.akira@mail.canon)

Date Prepared: June 27, 2022

Submission Type: Special 510(k) Submission

Proposed Device      Manufacturer:            Canon Inc.  
                                 Trade Name:            DIGITAL RADIOGRAPHY CXDI-Pro  
                                                                    DIGITAL RADIOGRAPHY D1  
                                 Common Name:        Solid State X-Ray Imager (Flat Panel/Digital Imager)  
                                 Classification Name: Stationary X-ray System  
                                 Product Code:        MQB  
                                 Regulation:           892.1680, Stationary X-ray System

Predicate Device:    Clearance:              K220098  
                                 Manufacturer:        Canon Inc.  
                                 Trade Name:         DIGITAL RADIOGRAPHY CXDI-Pro  
                                                                    DIGITAL RADIOGRAPHY D1  
                                 Common Name:        Solid State X-ray Imager (Flat Panel/Digital Imager)  
                                 Classification Name: Stationary X-ray System  
                                 Product Code:        MQB  
                                 Regulation:           892.1680, Stationary X-ray System

Device Description:    The DIGITAL RADIOGRAPHY CXDI-Pro, also called the DIGITAL RADIOGRAPHY D1, (hereinafter referred to as CXDI-Pro) is a solid-state x-ray imager. The CXDI-Pro is a series of detectors, and in the predicate device (K220098) consists of the CXDI-703C Wireless and CXDI-403C Wireless detectors, also called the AR-D3543W and AR-D4343W detectors respectively. The detectors intercept x-ray photons, and the scintillator emits visible spectrum photons that illuminate an array of photodetectors that create electrical signals. After the electrical signals are generated, the signals are converted to digital values. The digital values are sent to the PC via a wired or wireless connection, converted to images with the CXDI Control Software, and then displayed on the PC/monitors. The PC/monitors used with the CXDI-Pro are not a part of this submission.

                                 The subject of this Special 510(k) submission is a change to the predicate device, including compatibility with the Multi Box and Status Indicator, DIGITAL RADIOGRAPHY CXDI-Elite components that were cleared under K213780. The Status Indicator can be used when the proposed device is used in combination with the Multi Box. In addition, the CXDI Control Software has been updated from V3.10.0.8 to V3.10.2.2 to fix some bugs. Together, these changes make up the CXDI-Pro.

**510(k) SUMMARY**

Indications for Use:

The DIGITAL RADIOGRAPHY CXDI-Pro / DIGITAL RADIOGRAPHY D1 provides digital image capture for conventional film/screen radiographic examinations. This device is intended to capture, for display, radiographic images of human anatomy, and to replace radiographic film/screen systems in all general purpose diagnostic procedures. This device is not intended for mammography applications.

Summary of Technological Characteristics:

Comparison with the predicate device shows the technological characteristics of the CXDI-Pro is substantially equivalent to the predicate device.

The major differences between the CXDI-Pro and the predicate are differences in the components and a newer version of the CXDI control software. The proposed indications for use statement is identical to the indications for use statement of the predicate device.

	Proposed Device		Predicate Device		
Trade Name	DIGITAL RADIOGRAPHY CXDI-Pro / D1		DIGITAL RADIOGRAPHY CXDI-Pro / D1		
510(k) Submitter [Number]	Canon Inc. [TBD]		Canon Inc. [K220098]		IDENTICAL
Indication for Use	The DIGITAL RADIOGRAPHY CXDI-Pro / DIGITAL RADIOGRAPHY D1 provides digital image capture for conventional film/screen radiographic examinations. This device is intended to capture, for display, radiographic images of human anatomy, and to replace radiographic film/screen systems in all general purpose diagnostic procedures. This device is not intended for mammography applications.		The DIGITAL RADIOGRAPHY CXDI-Pro / DIGITAL RADIOGRAPHY D1 provides digital image capture for conventional film/screen radiographic examinations. This device is intended to capture, for display, radiographic images of human anatomy, and to replace radiographic film/screen systems in all general purpose diagnostic procedures. This device is not intended for mammography applications.		IDENTICAL
Application	General Radiography		General Radiography		IDENTICAL
Components	Detector (Sensor) Documentation (Manuals) Battery Charger (BC-1A/BC-01) Battery Pack (LB-4A) X-ray Interface Box (XB-1A) Power Box (PB-01) Wiring Cable (WC-01) PC Connection Cable (CP-01) Ready Indicator (RI-3A) Detector Stand (DS-01)		Detector (Sensor) Documentation (Manuals) Battery Charger (BC-1A/BC-01) Battery Pack (LB-4A) X-ray Interface Box (XB-1A) Power Box (PB-01) Wiring Cable (WC-01) PC Connection Cable (CP-01) Ready Indicator (RI-3A) Detector Stand (DS-01)		IDENTICAL
Other Compatible Components	Multi Box (MB-02) Status Indicator (SI-01/SI-4A)		N/A		MODIFIED
Detector Sensor	CXDI-703C Wireless AR-D3543W	CXDI-403C Wireless AR-D4343W	CXDI-703C Wireless AR-D3543W	CXDI-403C Wireless AR-D4343W	IDENTICAL
External Dimensions	384 x 460 x 15.5 mm	460 x 460 x 15.7 mm	384 x 460 x 15.5 mm	460 x 460 x 15.7 mm	IDENTICAL
Case Material	Magnesium alloy		Magnesium alloy		IDENTICAL
Detector Technology	TFT		TFT		IDENTICAL
Pixel Pitch	140µm		140µm		IDENTICAL
Scintillator	CsI(Tl) [Cesium Iodide doped with Thallium]		CsI(Tl) [Cesium Iodide doped with Thallium]		IDENTICAL
Spatial Resolution	35% [MTF@2lp/mm]		35% [MTF@2lp/mm]		IDENTICAL
DQE	58% [@0.5 lp/mm, 3.5 uGy]		58% [@0.5 lp/mm, 3.5 uGy]		IDENTICAL
IP Level	IP55		IP55		IDENTICAL
Software	CXDI Control Software V3.10.2.2		CXDI Control Software V3.10.0.8		MODIFIED
Wireless Communication	IEEE 802.11n/ac/a/g/b 2.4GHz/5GHz Bluetooth Low Energy		IEEE 802.11n/ac/a/g/b 2.4GHz/5GHz Bluetooth Low Energy		IDENTICAL
Photographing Mode	Standard Synchronization Mode, Non Generator Connection Mode		Standard Synchronization Mode, Non Generator Connection Mode		IDENTICAL

**510(k) SUMMARY**

Summary of  
Non-Clinical /  
Test Data:

The fundamental scientific technology of the CXDI-Pro has not been modified. The changes are the addition of optional accessories, a Multi Box and Status Indicator; the update to the detector firmware from 01.00.02.00 to 01.01.03.00; and the update to the CXDI Control Software from 3.10.0.8 to 3.10.2.2.

Evaluation of the addition of the new detector and optional accessories confirmed that the changes did not impact CXDI-Pro conformance with the U.S. Performance Standard for radiographic equipment and with relevant voluntary safety standards for Electrical safety and Electromagnetic Compatibility testing, specifically IEC standards 60601-1, 60601-1-2, 60601-1-6, and 60601-2-54.

These verification/validation activities successfully demonstrated that the device continues to meet the standards for the areas impacted by the modifications to the predicate device and raises no new questions regarding either safety or effectiveness when compared to the predicate device. Clinical testing is not necessary for the current submission, based on the minor differences from the predicate device. Adequate verification and validation activities should be sufficient to demonstrate that the CXDI-Pro works as intended. Therefore, the verification/validation activities conducted support a determination of substantial equivalence for the CXDI-Pro.

The proposed device follows the applicable elements of the following FDA guidance documents: *Guidance for the Submission of 510(k)s for Solid State X-ray Imaging Devices*, *Guidance for the Content of Premarket Submissions for Software contained in Medical Devices*, *Content of Premarket Submissions for Management of Cybersecurity in Medical Devices*, and *Use of International Standard ISO 10993-1, "Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process."*

Conclusion:

Canon Inc. considers the DIGITAL RADIOGRAPHY CXDI-Pro / D1 to be substantially equivalent to the predicate device listed above. This conclusion is based on the similarities in primary intended use, principles of operation, functional design, and established medical use.