



Exactech, Inc.  
Zach Sharrah  
Manager, Regulatory Affairs  
2320 NW 66th Court  
GAINSVILLE, FL 32653

April 29, 2020

Re: K192097

Trade/Device Name: Exactech® Equinnox® Stemless Humeral Components  
Regulation Number: 21 CFR 888.3660  
Regulation Name: Shoulder joint metal/polymer semi-constrained cemented prosthesis  
Regulatory Class: Class II  
Product Code: PKC  
Dated: March 26, 2020  
Received: March 30, 2020

Dear Zach Sharrah:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Michael Owens  
Acting Assistant Director  
DHT6A: Division of Joint  
Arthroplasty Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K192097

Device Name

Exactech® Equinoxe® Stemless Humeral Components

Indications for Use (Describe)

The Equinoxe Stemless Shoulder System is indicated for use in skeletally mature individuals with degenerative diseases of the glenohumeral joint where anatomic shoulder arthroplasty is determined by the surgeon to be the preferred method of treatment.

Clinical indications for anatomic total shoulder arthroplasty are as follows:

- Osteoarthritis, osteonecrosis or post-traumatic degenerative problems
- Congenital abnormalities in the skeletally mature
- Primary and secondary necrosis of the humeral head
- Pathologies where arthrodesis or resectional arthroplasty of the humeral head are not acceptable
- Revisions of humeral prostheses when other treatments or devices have failed (where adequate fixation can be achieved)
- To restore mobility from previous procedures (e.g. previous fusion)

In the USA, the Equinoxe Stemless Shoulder is only indicated for total shoulder arthroplasty.

The Equinoxe Stemless Shoulder humeral components are indicated for press-fit, uncemented use.

When used in total shoulder arthroplasty, the Equinoxe Stemless Shoulder System is intended to be used with the cemented Equinoxe glenoid components.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**Exactech® Equinox® Stemless Humeral Components  
Traditional 510(k) – 510(k) Summary**

**Company:** Exactech®, Inc  
2320 NW 66<sup>th</sup> Court  
Gainesville, FL 32653

**Date:** July 30, 2019

**Contact Person:** Zach Sharrah  
Manager, Regulatory Affairs

Phone: (352) 377-1140  
Fax: (352) 378-2617

**Proprietary Name:** Exactech® Equinox® Stemless Humeral Components

**Common Name:** Shoulder Prosthesis

**Classification Name:** Prosthesis, Total Anatomic Shoulder, Uncemented  
Metaphyseal Humeral Stem with No Diaphyseal Incursion,  
Semi-Constrained, 21 CFR 888.3660, Class II, Product  
Code PKC

**Legally Marketed Device to Which Substantial Equivalence Is Claimed:**

- Exactech Equinox Stemless Shoulder (K173388)

**Reference Devices**

- Tornier Simplici Shoulder System (K143552)
- Equinox Humeral Heads (K042021, K140063)

**Device Description**

The Equinox Stemless Humeral Components are intended to be used with Exactech Equinox Stemless Humeral Heads and the Equinox glenoid components for use in Total Shoulder Arthroplasty. The Exactech glenoid components are indicated for cemented use. The proposed Equinox Stemless Humeral Components are additively manufactured from Ti-6Al-4V and have porous regions. The Equinox Stemless Humeral Components are available in in three sizes, with lengths between 17mm and 24mm.

## Exactech® Equinox® Stemless Humeral Components Traditional 510(k) – 510(k) Summary

### Indications for Use

The Equinox Stemless Shoulder System is indicated for use in skeletally mature individuals with degenerative diseases of the glenohumeral joint where anatomic shoulder arthroplasty is determined by the surgeon to be the preferred method of treatment.

Clinical indications for anatomic total shoulder arthroplasty are as follows:

- Osteoarthritis, osteonecrosis or post-traumatic degenerative problems
- Congenital abnormalities in the skeletally mature
- Primary and secondary necrosis of the humeral head.
- Pathologies where arthrodesis or resectional arthroplasty of the humeral head are not acceptable
- Revisions of humeral prostheses when other treatments or devices have failed (where adequate fixation can be achieved)
- To restore mobility from previous procedures (e.g. previous fusion)

In the USA, the Equinox Stemless Shoulder is only indicated for total shoulder arthroplasty.

The Equinox Stemless Shoulder humeral components are indicated for press-fit, uncemented use.

When used in total shoulder arthroplasty, the Equinox Stemless Shoulder System is intended to be used with the cemented Equinox glenoid components.

### Summary of Technological Characteristics

The rationale for substantial equivalence is based on consideration of the following device use and characteristics:

- **Indications for Use.** The proposed Exactech Equinox Stemless Shoulder components and the predicate devices have similar indications for use.
- **Materials/Surface Finish/Coatings.** The proposed Exactech Equinox Stemless Humeral Components and the predicate devices are composed of the similar biocompatible substrate materials, and the same or similar surface finish for permanent implants.
- **Design Features.** The proposed Exactech Equinox Stemless Shoulder components and the predicate devices have similar design features.
- **Dimensions.** The proposed Exactech Equinox Stemless Shoulder components and the predicate devices are dimensionally comparable.
- **Sterilization.** The proposed Exactech Equinox Stemless Shoulder components and the predicate devices are provided sterile for single use only.
- **Performance Requirements.** The proposed Exactech Equinox Stemless Shoulder components and the predicate devices conform to recognized performance standards for total shoulder replacement devices.

**Exactech® Equinox® Stemless Humeral Components  
Traditional 510(k) – 510(k) Summary**

**Non-Clinical Testing**

The following engineering analyses were performed to demonstrate that the Exactech Equinox Stemless Shoulder perform as intended and are substantially equivalent to the identified predicate devices:

- Fatigue Testing
- Axial Pull Out Testing
- Torque out Testing
- Taper Disengagement Testing
- Porous Structure Characterization

Pyrogen testing was conducted in accordance with USP <161>, USP <85>, and ANSI/AAMI ST72 to ensure the proposed Equinox Stemless Shoulder components meet recommended limits per FDA's *Guidance Document Submission and Review of Sterility Information in Premarket (510(k)) Submission for Devices Labeled as Sterile*.

**Substantial Equivalence Conclusion**

Based on consideration of indications for use, technological characteristics, and results of non-clinical testing, it was concluded that the Exactech Equinox Stemless Shoulder demonstrates substantial equivalence to the referenced predicate devices.