



February 26, 2021

AzurMeds Inc.
Jean-Marie Berger
Chief Commercial Officer
4809 N Ravenswood Avenue Suite 119
Chicago, Illinois 60604

Re: K210268

Trade/Device Name: LTJ Screws and Washer
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener
Regulatory Class: Class II
Product Code: HWC, HTN
Dated: January 29, 2021
Received: February 1, 2021

Dear Jean-Marie Berger:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, M.P.H.
Assistant Director
DHT6C: Division of Restorative, Repair
and Trauma Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K210268

Device Name

LTJ Screws and Washer

Indications for Use (Describe)

The LTJ Screws (3.75mm solid and 4.5mm cannulated) are intended to be used in skeletally mature patient as stand-alone bone screws for bone reconstruction, osteotomy, arthrodesis, joint fusion, ligament fixation, fracture repair and fracture fixation, appropriate for the size of the device. Optional washers are available for the 3.75 solid screws.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



510(k) Summary

AzurMeds., Inc's LTJ Screws and Washer

Date prepared: January 29, 2021

Assigned 510(k) number: -

Company: AzurMeds Inc.
4809 N Ravenswood Ave 119
CHICAGO, IL, 60640,
USA
Phone: + 1 (773)-564-9324
Cell: + 1 (612) 810-7221

Contact: Jean-Marie Berger

Trade/Proprietary Name: LTJ Screws and Washer

Regulation & Classification Name: (Primary) 21 CFR 888.3040 - Smooth or threaded metallic bone fixation fastener
21 CFR 888.3030 - Single/multiple component metallic bone fixation appliances and accessories

Device Classification: Class II

Product Code & Common Name: HWC (Screw, Fixation, Bone)
HTN (Washer, Bolt Nut)

Predicate Devices: (Primary predicate): LTJ Screws and Washer (K200987)
(Reference predicate): Monster Screw System™ (K124027)

Device Description: The purpose of this Special 510(k) is to add additional references to the LTJ Screws and Washer system. These new references are provided in titanium alloy compliant with ASTM F136 and have the same design and dimensions as the cleared LTJ Screws and Washer provided in stainless steel alloy compliant with ASTM F138.

The LTJ Screws are partially threaded solid headless 3.75 mm screws and partially threaded cannulated headed 4.5 mm screws provided in various length ranging from 26 mm to 50 mm. The headless implants can be

associated to an LTJ diameter 8 mm Washer. It is essential to handle and insert implants with LTJ instruments specifically designed for this purpose.

Intended Use and indications for use:

The LTJ Screws (3.75mm solid and 4.5mm cannulated) are intended to be used in skeletally mature patient as stand-alone bone screws for bone reconstruction, osteotomy, arthrodesis, joint fusion, ligament fixation, fracture repair and fracture fixation, appropriate for the size of the device. Optional washers are available for the 3.75 solid screws.

Technology Comparison and Non-Clinical Performance Testing:

Mechanical testing was performed under ASTM F543 requirements. These tests showed the subject device to be substantially equivalent in terms of performance to the predicate LTJ Screws and Washer (K200987).

Conclusion:

There are no substantial differences between the new references of LTJ Screws and Washer and the predicate devices with respect to intended use and technological characteristics, including design, materials of manufacture, mechanical properties, and intended effect.

Therefore, the LTJ Screws and Washer can be found substantially equivalent to the cited predicate, as it does not raise new questions of safety and effectiveness.