



March 26, 2021

Gramercy Extremity Orthopedics, LLC
% Tamala Wampler
Regulatory Consultant
Novus Management Group, LLC
4480 Lake Forest Drive, Ste. 412
Blue Ash, Ohio 45242

Re: K210534

Trade/Device Name: GEO FirstFUSE 1st MTP Joint Arthrodesis Plating System
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And Accessories
Regulatory Class: Class II
Product Code: HRS, HWC
Dated: February 15, 2021
Received: February 24, 2021

Dear Tamala Wampler:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For: Shumaya Ali, M.P.H.
Assistant Director
DHT6C: Division of Restorative, Repair
and Trauma Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K210534

Device Name

GEO FirstFUSE 1st MTP Joint Arthrodesis Plating System

Indications for Use (Describe)

The GEO FirstFUSE 1st MTP Joint Arthrodesis Plating System is indicated for use in the stabilization and fixation of the first metatarsal-phalangeal joint in the foot for surgical fusion (arthrodesis), osteotomy, nonunion, malunion or revision surgery.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



GEO FirstFUSE 1st MTP Joint Arthrodesis Plating System 510(k) Summary

Submitted by/ Sponsor:	Gramercy Extremity Orthopedics, LLC. 840 F. Avenue #104 Richardson, TX 75081 USA 972-908-9808	Contact Person:	Michael Simpson CEO 855-436-2278
Date Prepared:	February 12, 2021		
Trade Name:	GEO <i>FirstFUSE</i> 1 st MTP Joint Arthrodesis Plating System		
Common Name:	Plate, Fixation, Bone		
Classification Code Name & Reference:	HRS	Plate, Fixation, Bone	21 CFR §888.3030
	HWC	Screw, Fixation Bone	21 CFR §888.3040
Predicate Devices	Primary Predicate Device: K200108 GEO 1 st MTP Joint Arthrodesis Plating System; Additional Predicate Device: K041287 Vilex Plating System		

Device Description:

The GEO *FirstFUSE* 1st MTP Joint Arthrodesis Plating System consists of anatomically contoured, low-profile plates in a variety of configurations of length, left and right versions, and degrees of dorsiflexion and abduction. The plates accept both fixed angle locking screws and non-locking Extremity Plating System (EPS) screws. Fixed angle cortical locking screws are available in diameters of 2.7mm and 3.5. Non-locking cortical screws are available in 2.7mm, 3.5mm and 4.0mm; and non-locking cancellous screws in 4.0mm. Screws lengths are offered in 10-40mm for the 2.7mm diameter plate screws and in 10 – 60mm in 3.5mm and 4.0mm diameter plate screws. All components (implants and the associated instruments) are provided sterile and for single use only.

Indications for Use:

The GEO *FirstFUSE* 1st MTP Joint Arthrodesis Plating System is indicated for use in the stabilization and fixation of the first metatarsal-phalangeal joint in the foot for surgical fusion (arthrodesis), osteotomy, nonunion, malunion or revision surgery.

Technological Characteristics:

The GEO *FirstFUSE* 1st MTP Joint Arthrodesis Plating System subject device has the same design and technological features as the cleared, commercially distributed predicate GEO 1st MTP Joint Arthrodesis Plating System (K200108). The additional screws lengths are equivalent to the screws lengths collectively offered by the primary predicate and the additional predicate, Vilex Plating System (K041287). The implant material remains unchanged. All patient contacting components are comprised of biocompatible materials. All System components are provided sterile by gamma radiation and are disposable, for single-use only.

Substantial Equivalence:

The GEO *FirstFUSE* 1st MTP Joint Arthrodesis Plating System is substantially equivalent to the GEO 1st MTP Joint Arthrodesis Plating System cleared under K200108 and the Vilex Plating System cleared under K041287. The GEO *FirstFUSE* 1st MTP Joint Arthrodesis Plating System has the same intended use and material as the predicate(s) and the minor technological differences in plate and screw geometry (overall

plate and screw lengths, plate hole placement/configuration and angulation, and overall screw lengths) are not considered to raise new questions of safety or effectiveness.

Performance Data:

Engineering analysis and rationale was completed to provide evidence that the modifications to system do not adversely affect the safety and effectiveness of the system components. The results of this analysis demonstrate substantial equivalence of the GEO FirstFUSE 1st MTP Joint Arthrodesis Plating System to the predicate device.

Conclusion:

The GEO FirstFUSE 1st MTP Joint Arthrodesis Plating System is considered to be substantially equivalent to the predicate device. This conclusion is based on the similarities in principles of operation, technology, materials and intended use.