



May 27, 2021

Medtronic, Inc.
% Shweta Sharma
Principal Regulatory Affairs Specialist
Medtronic Sofamor Danek
1800 Pyramid Place
Memphis, Tennessee 38132

Re: K211057

Trade/Device Name: LigaPASS® Spinal System, CD Horizon™ Spinal System
Regulation Number: 21 CFR 888.3010
Regulation Name: Bone Fixation Cerclage
Regulatory Class: Class II
Product Code: OWI, NKB, KWP, KWQ
Dated: April 9, 2021
Received: April 9, 2021

Dear Shweta Sharma:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for

Colin O'Neill, M.B.E.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K211057

Device Name
LigaPASS® Spinal System

Indications for Use (Describe)

The LigaPASS is a temporary implant for use in orthopedic surgery. The system is intended to provide temporary stabilization as a bone anchor during the development of solid bony fusion and aid in the repair of bone fractures. The indications for use include the following applications:

- Spinal trauma, used in sublaminar, or facet wiring techniques
- Spinal reconstruction surgery, incorporated into constructs for the purpose of correction of spinal deformities such as idiopathic and neuromuscular scoliosis in patients 8 years of age and older, adult scoliosis, and kyphosis;
- Spinal degenerative surgery, as an adjunct to spinal fusions;

The LigaPASS system may also be used in conjunction with other medical implant grade implants made of titanium or cobalt chrome alloy whenever "wiring" may help secure the attachment of other implants.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Indications for Use

510(k) Number (if known)
K211057

Device Name

CD Horizon™ Spinal System

Indications for Use (Describe)

The CD Horizon™ Spinal System with or without Sextant™ instrumentation is intended for posterior, non-cervical fixation as an adjunct to fusion for the following indications: degenerative disc disease (DDD - defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, trauma (i.e. fracture or dislocation), spinal stenosis, curvatures (i.e. scoliosis, kyphosis, or lordosis), tumor, pseudarthrosis, and/or failed previous fusion. Except for hooks, when used as an anterolateral thoracic/lumbar system, the CD Horizon™ Spinal System titanium, cobalt chrome, and stainless steel implants may also be used for the same indications as an adjunct to fusion.

With the exception of DDD, the CD Horizon™ Legacy™ 3.5mm rods and associated components may be used for the aforementioned indications in skeletally mature patients as an adjunct to fusion. The 3.5mm rods may be used for the specific pediatric indications noted below.

When used for posterior non-cervical pedicle screw fixation in pediatric patients, the CD Horizon™ Spinal System titanium, cobalt chrome, and stainless steel implants are indicated as an adjunct to fusion to treat progressive spinal deformities (i.e. scoliosis, kyphosis, or lordosis) including idiopathic scoliosis, neuromuscular scoliosis, and congenital scoliosis. Additionally, the CD Horizon™ Spinal System is intended to treat pediatric patients diagnosed with the following conditions: spondylolisthesis/ spondylolysis, fracture caused by tumor and/or trauma, pseudarthrosis, and/or failed previous fusion. These devices are to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

The CD Horizon™ PEEK rods are intended to provide posterior supplemental fixation when used with an interbody fusion cage for patients diagnosed with DDD. These DDD patients may also have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level. This device is intended for 1-2 level use in the lumbosacral spine (L2 – S1) in skeletally mature patients. The device is intended for use with an interbody fusion cage at the instrumented level and is not intended for stand-alone use. The CD Horizon™ Spire™ plate is a posterior, single-level, non-pedicle supplemental fixation device intended for use in the non-cervical spine (T1-S1) as an adjunct to fusion in skeletally mature patients. It is intended for plate fixation/attachment to spinous processes for the purpose of achieving supplemental fixation in the following conditions: DDD (as previously defined), spondylolisthesis, trauma, and/or tumor.

In order to achieve additional levels of fixation, the CD Horizon™ Spinal System rods may be connected to the Vertex™ Reconstruction System with the Vertex™ rod connector. Refer to the Vertex™ Reconstruction System package insert for a list of the Vertex™ indications of use . Vertex™ rod connector. Refer to the Vertex™ Reconstruction System package insert for a list of the Vertex™ indications of use.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

510(k) Summary

09 April 2021

- I. Company:** Medtronic, Inc.
1800 Pyramid Place
Memphis, TN 38132
Telephone Number: 901.396.3133
- Applicant:** Taylor Gold West, MBA
Principal Regulatory Affairs Specialist
Telephone Number: 720.890.2322
Fax Number: 720.890.3500
Email: taylor.c.goldwest@medtronic.com
- Primary Contact:** Shweta Sharma
Principal Regulatory Affairs Specialist
Telephone Number: 901.399.2425
Email: Shweta.s.sharma@medtronic.com
- Alternate Contact:** Raphael McInnis
Senior Regulatory Affairs Manager
Telephone Number: 901-399-2057
Email: Raphael.McInnis@medtronic.com
- II. Proprietary Trade Name:** LigaPASS® Spinal System
CD Horizon™ Spinal System
- III. Classification Name:**
LigaPASS® Spinal System: Bone fixation cerclage
CD Horizon™ Spinal System: Thoracolumbosacral pedicle screw system
- IV. Common Name:**
LigaPASS® Spinal System: Bone fixation cerclage, sublaminar
CD Horizon™ Spinal System: Spinal Fixation Appliance, Spinal Fixation Orthosis
- V. Classification:** Class II
- VI. Product Code:**
LigaPASS® Spinal System: OWI (21 CFR 888.3010)
CD Horizon™ Spinal System: NKB, KWP, KWQ (21 CFR 888.3070, 21 CFR 888.3060, 21 CFR 888.3050)
- VII. Primary Predicate Device LigaPASS® Spinal System**
510(k): K173506
Device name: LigaPASS® Spinal System
Clearance Date: April 03, 2018
- Primary Predicate Device CD Horizon™ Spinal System**
510(k): K202771
Device name: CD Horizon™ Spinal System
Clearance Date: October 19, 2020

These predicate devices have not been subject to a design-related recall.

VIII. Product Description

The purpose of this submission is to expand use of the previously cleared LigaPASS® Spinal System with the previously cleared CD Horizon™ 5.5/6.0 Spinal System rods.

LigaPASS® Spinal System

The LigaPASS® MEDICREA® INTERNATIONAL S.A. spinal system is composed of four components: a connector, two set screws and a band.

The LigaPASS® system is designed to stabilize a vertebra in the same manner as a hook around the vertebra during development of solid bony fusion. The LigaPASS® system must be implanted via a posterior approach to complete a thoraco-lumbar fixation system as the PASS LP® MEDICREA® INTERNATIONAL S.A. spinal system.

The implants are manufactured in titanium alloy Ti-6Al-4V ELI conforming to ISO 5832-3 specifications and ASTM F136, with the exception of the band which is manufactured in polyethylene terephthalate (PET) and titanium T40 conforming to ISO 5832-2 specifications and ASTM F67.

As any orthopaedic implant, these implants must not be reused.

CD Horizon™ Spinal System

The CD Horizon™ Spinal System consists of a variety of shapes and sizes of rods, hooks, screws, Crosslink™ plates, staples and connecting components, as well as implant components from other Medtronic spinal systems, which can be rigidly locked into a variety of configurations, with each construct being tailor-made for the individual case.

IX. Indications for Use

LigaPASS® Spinal System

The LigaPASS is a temporary implant for use in orthopedic surgery. The system is intended to provide temporary stabilization as a bone anchor during the development of solid bony fusion and aid in the repair of bone fractures. The indications for use include the following applications:

- Spinal trauma, used in sublaminar, or facet wiring techniques
- Spinal reconstruction surgery, incorporated into constructs for the purpose of correction of spinal deformities such as idiopathic and neuromuscular scoliosis in patients 8 years of age and older, adult scoliosis, and kyphosis;
- Spinal degenerative surgery, as an adjunct to spinal fusions;

The LigaPASS system may also be used in conjunction with other medical implant grade implants made of titanium or cobalt chrome alloy whenever "wiring" may help secure the attachment of other implants.

CD Horizon™ Spinal System

The CD Horizon™ Spinal System with or without Sextant™ instrumentation is intended for posterior, non-cervical fixation as an adjunct to fusion for the following indications: degenerative disc disease (DDD - defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, trauma (i.e. fracture or dislocation), spinal stenosis, curvatures (i.e. scoliosis, kyphosis, or lordosis), tumor, pseudarthrosis, and/or failed previous fusion.

Except for hooks, when used as an anterolateral thoracic/lumbar system, the CD Horizon™ Spinal System titanium, cobalt chrome, and stainless steel implants may also be used for the same indications as an adjunct to fusion.

With the exception of DDD, the CD Horizon™ Legacy™ 3.5mm rods and associated components may be used for the aforementioned indications in skeletally mature patients as an adjunct to fusion. The 3.5mm rods may be used for the specific pediatric indications noted below.

When used for posterior non-cervical pedicle screw fixation in pediatric patients, the CD Horizon™ Spinal System titanium, cobalt chrome, and stainless steel implants are indicated as an adjunct to fusion to treat progressive spinal deformities (i.e. scoliosis, kyphosis, or lordosis) including idiopathic scoliosis, neuromuscular scoliosis, and congenital scoliosis. Additionally, the CD Horizon™ Spinal System is intended to treat pediatric patients diagnosed with the following conditions: spondylolisthesis/ spondylolysis, fracture caused by tumor and/or trauma, pseudarthrosis, and/or failed previous fusion. These devices are to be used with autograft and/or allograft. Pediatric pedicle screw fixation is limited to a posterior approach.

The CD Horizon™ PEEK rods are intended to provide posterior supplemental fixation when used with an interbody fusion cage for patients diagnosed with DDD. These DDD patients may also have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level. This device is intended for 1-2 level use in the lumbosacral spine (L2 – S1) in skeletally mature patients. The device is intended for use with an interbody fusion cage at the instrumented level and is not intended for stand-alone use.

The CD Horizon™ Spire™ plate is a posterior, single-level, non-pedicle supplemental fixation device intended for use in the non-cervical spine (T1-S1) as an adjunct to fusion in skeletally mature patients. It is intended for plate fixation/attachment to spinous processes for the purpose of achieving supplemental fixation in the following conditions: DDD (as previously defined), spondylolisthesis, trauma, and/or tumor.

In order to achieve additional levels of fixation, the CD Horizon™ Spinal System rods may be connected to the Vertex™ Reconstruction System with the Vertex™ rod connector. Refer to the Vertex™ Reconstruction System package insert for a list of the Vertex™ indications of use.

X. Comparison of the Technological Characteristics

When the subject devices, LigaPASS® Spinal System and the CD Horizon™ Spinal System, are used together, there is no change to the fundamental scientific technology, indications for use, intended use, materials, and levels of attachment in comparison to the predicates (K173506, LigaPASS® Spinal System, and K202771, CD Horizon™ Spinal System,

respectively). The purpose of the submission is to expand use of the previously cleared LigaPASS® Spinal System (K173506) with the previously cleared CD Horizon® 5.5/6.0 Spinal System (K202771).

XI. Discussion of Performance Testing

In accordance with the Guidance for Industry and FDA Staff – Spinal System 510(k)'s, Medtronic has evaluated the subject devices to demonstrate substantial equivalence to the predicate devices.

Mechanical testing was conducted in accordance with ASTM 1798. The performance testing demonstrates mechanical equivalence of the LigaPASS® Spinal System when used with the CD Horizon™ 5.5/6.0 Spinal System.

XII. Conclusion

Based on the test results and supporting information provided in the subject submission, Medtronic believes the LigaPASS® Spinal System devices when used with CD Horizon 5.5/6.0 Spinal System implants are as safe and effective as the legally marketed predicate devices.