



August 19, 2022

Spinal Simplicity LLC  
Mr. Adam Rogers  
Director of Regulatory  
6363 College Blvd, Ste 320  
Overland Park, Kansas 66211

Re: K221023

Trade/Device Name: Posterior Fusion Plate/HA Posterior Fusion Plate, Minuteman G3 MIS Fusion Plate, HA Minuteman G3 MIS Fusion Plate, HA Minuteman G3-R MIS Fusion Plate, Minuteman G5 MIS Fusion Plate

Regulation Number: 21 CFR 888.3050

Regulation Name: Spinal Interlaminar Fixation Orthosis

Regulatory Class: Class II

Product Code: PEK

Dated: July 25, 2022

Received: July 25, 2022

Dear Mr. Rogers:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for

Colin O'Neill, M.B.E.  
Assistant Director  
DHT6B: Division of Spinal Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K221023

Device Name  
Spinal Simplicity Posterior Fusion Plate / HA Posterior Fusion Plate

### Indications for Use (Describe)

The Spinal Simplicity Posterior Fusion Plate/HA Posterior Fusion Plate is a posterior, non-pedicle fusion device, intended for use at a single level in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving fusion in the following conditions:

- Lumbar spinal stenosis;
- degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); and/or
- spondylolisthesis.

The Posterior Fusion Plate/HA Posterior Fusion Plate is intended for use with bone graft material. The device may be implanted via a minimally invasive lateral approach (L1-S1) or a minimally invasive posterior approach (T1-S1).

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## Indications for Use

510(k) Number (*if known*)  
K221023

Device Name  
Spinal Simplicity Minuteman G3 MIS Fusion Plate

### Indications for Use (*Describe*)

The Spinal Simplicity Minuteman G3 MIS Fusion Plate is a posterior, non-pedicle fusion device, intended for use at a single level in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving fusion in the following conditions:

- Lumbar spinal stenosis;
- degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); and/or
- spondylolisthesis.

The Minuteman G3 MIS Fusion Plate is intended for use with bone graft material. The device may be implanted via a minimally invasive lateral approach (L1-S1) or a minimally invasive posterior approach (T1-S1).

Type of Use (*Select one or both, as applicable*)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## Indications for Use

510(k) Number (*if known*)  
K221023

Device Name  
Spinal Simplicity HA Minuteman G3 MIS Fusion Plate

### Indications for Use (*Describe*)

The Spinal Simplicity HA Minuteman G3 MIS Fusion Plate is a posterior, non-pedicle fusion device, intended for use at a single level in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving fusion in the following conditions:

- Lumbar spinal stenosis;
- degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); and/or
- spondylolisthesis.

The HA Minuteman G3 MIS Fusion Plate is intended for use with bone graft material. The device may be implanted via a minimally invasive lateral approach (L1-S1) or a minimally invasive posterior approach (T1-S1).

Type of Use (*Select one or both, as applicable*)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## Indications for Use

510(k) Number (*if known*)  
K221023

Device Name  
Spinal Simplicity HA Minuteman G3-R MIS Fusion Plate

### Indications for Use (*Describe*)

The Spinal Simplicity HA Minuteman G3-R MIS Fusion Plate is a posterior, non-pedicle fusion device, intended for use at a single level in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving fusion in the following conditions:

- Lumbar spinal stenosis;
- degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); and/or
- spondylolisthesis.

The HA Minuteman G3-R MIS Fusion Plate is intended for use with bone graft material. The device may be implanted via a minimally invasive lateral approach (L1-S1).

Type of Use (*Select one or both, as applicable*)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## Indications for Use

510(k) Number (if known)  
K221023

Device Name  
Spinal Simplicity Minuteman G5 MIS Fusion Plate

### Indications for Use (Describe)

The Spinal Simplicity Minuteman G5 MIS Fusion Plate is a posterior, non-pedicle fusion device, intended for use at a single level in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving fusion in the following conditions:

- Lumbar spinal stenosis;
- degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); and/or
- spondylolisthesis.

The Minuteman G5 MIS Fusion Plate is intended for use with bone graft material. The device may be implanted via a minimally invasive lateral approach (L1-S1) or a minimally invasive posterior approach (T1-S1).

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

### Submitter's Name, Address, Telephone Number, Contact Person and Date Prepared:

Spinal Simplicity LLC  
6363 College Blvd  
Suite 320  
Overland Park, KS 66211  
Phone: (913) 451-4414  
Facsimile: (913) 888-0075

Contact Person: Adam Rogers

Date Prepared: July 25, 2022

### Name of Device:

- Minuteman G3 MIS Fusion Plate
- HA Minuteman G3 MIS Fusion Plate
- HA Minuteman G3-R MIS Fusion Plate
- Posterior Fusion Plate / HA Posterior Fusion Plate
- Minuteman G5 MIS Fusion Plate

### Common / Classification Name:

Common Name: Spinous Process Plate  
Classification Name: 21 CFR 888.3050 - Spinal Interlaminar Fixation Orthosis, Class II  
Product code: PEK

### Predicate Devices:

- Spinal Simplicity's Posterior Fusion Plate / HA Posterior Fusion Plate (K200066 & K212781 (Primary Predicate))
- Spinal Simplicity's HA Minuteman G3 MIS Fusion Plate (K151741 & K212781)
- Spinal Simplicity's Minuteman G3 MIS Fusion Plate (K140046 & K212781)
- Spinal Simplicity's HA Minuteman G3-R MIS Fusion Plate (K163428 & K212781)
- Spinal Simplicity's Minuteman G5 MIS Fusion Plate (K211880)

### Device Description:

The Minuteman devices consist of bilateral Plates and a Body/Post that connects the Plates. The Plate components include several gripping features for attachment of the device to the spinous processes. The Minuteman devices are available in multiple sizes to accommodate varying patient anatomy. The Minuteman devices are made from Ti6Al4V and Ti6Al4V ELI. The HA



versions of the Minuteman devices have an additional hydroxyapatite coating on the distal regions of the device.

The purpose of this submission is to update the Indications for Use. No technological, engineering, performance, or material changes have been made to the devices.

**Intended Use / Indications for Use:**

- Spinal Simplicity Posterior Fusion Plate / HA Posterior Fusion Plate:

The Spinal Simplicity Posterior Fusion Plate/HA Posterior Fusion Plate is a posterior, non-pedicle fusion device, intended for use at a single level in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving fusion in the following conditions:

- Lumbar spinal stenosis;
- degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); and/or
- spondylolisthesis.

The Posterior Fusion Plate/HA Posterior Fusion Plate is intended for use with bone graft material. The device may be implanted via a minimally invasive lateral approach (L1-S1) or a minimally invasive posterior approach (T1-S1).

- Spinal Simplicity Minuteman G3 MIS Fusion Plate:

The Spinal Simplicity Minuteman G3 MIS Fusion Plate is a posterior, non-pedicle fusion device, intended for use at a single level in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving fusion in the following conditions:

- Lumbar spinal stenosis;
- degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); and/or
- spondylolisthesis.

The Minuteman G3 MIS Fusion Plate is intended for use with bone graft material. The device may be implanted via a minimally invasive lateral approach (L1-S1) or a minimally invasive posterior approach (T1-S1).

- Spinal Simplicity HA Minuteman G3 MIS Fusion Plate:

The Spinal Simplicity HA Minuteman G3 MIS Fusion Plate is a posterior, non-pedicle fusion device, intended for use at a single level in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving fusion in the following conditions:

- Lumbar spinal stenosis;
- degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); and/or
- spondylolisthesis.

The HA Minuteman G3 MIS Fusion Plate is intended for use with bone graft material. The device may be implanted via a minimally invasive lateral approach (L1-S1) or a minimally invasive posterior approach (T1-S1).

- Spinal Simplicity HA Minuteman G3-R MIS Fusion Plate:

The Spinal Simplicity HA Minuteman G3-R MIS Fusion Plate is a posterior, non-pedicle fusion device, intended for use at a single level in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving fusion in the following conditions:

- Lumbar spinal stenosis;
- degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); and/or
- spondylolisthesis.

The HA Minuteman G3-R MIS Fusion Plate is intended for use with bone graft material. The device may be implanted via a minimally invasive lateral approach (L1-S1).

- Spinal Simplicity Minuteman G5 MIS Fusion Plate:

The Spinal Simplicity Minuteman G5 MIS Fusion Plate is a posterior, non-pedicle fusion device, intended for use at a single level in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving fusion in the following conditions:

- Lumbar spinal stenosis;
- degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); and/or
- spondylolisthesis.

The Minuteman G5 MIS Fusion Plate is intended for use with bone graft material. The device may be implanted via a minimally invasive lateral approach (L1-S1) or a minimally invasive posterior approach (T1-S1).

**Technological Characteristics:**

There are no changes to the technological characteristics from the previously cleared versions of each Minuteman device. The only differences are changes to the Indications for Use.

**Performance Data:**

There are no changes to the functional characteristics from the previously cleared versions of each Minuteman device. The only differences are changes to the Indications for Use. Therefore, no performance bench testing data has been submitted.

**Clinical Data:**

Clinical data were provided to demonstrate satisfactory clinical and radiological outcomes to support the modified Indications for Use.

**Substantial Equivalence:**

The Spinal Simplicity Minuteman family of implants subject to this 510(k) notice are the same as the predicate devices cleared under K140046, K151741, K163428, K200066, K211880, and K212781. The only difference is updated Indications for Use. The provided clinical data and information from the previous 510(k) applications support the conclusion that the subject devices with updated Indications for Use are substantially equivalent to the previously cleared predicate devices.

**Conclusions:**

The information submitted by Spinal Simplicity in this premarket notification demonstrates that the Minuteman implants perform as intended and are substantially equivalent to the predicate devices.