



July 26, 2022

Parcus Medical, LLC
Vasavi Vora
Regulatory Affairs Specialist II
6423 Parkland Drive
Sarasota, Florida 34243

Re: K221502

Trade/Device Name: Parcus Synd-EZ SS
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener
Regulatory Class: Class II
Product Code: MBI
Dated: May 24, 2022
Received: May 25, 2022

Dear Vasavi Vora:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Laura C. Rose, Ph.D.
Assistant Director
DHT6C: Division of Restorative, Repair
and Trauma Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2023
See PRA Statement below.

Indications for Use

510(k) Number *(if known)*
K221502

Device Name
Parcus Synd-EZSS

Indications for Use *(Describe)*

The Parcus Synd-EZ SS is intended to be used as an adjunct in fracture repair involving metaphyseal and periarticular small bone fragments where screws are not indicated, and as adjuncts in external and intramedullary fixation systems involving plates and rods, with fracture braces and casting.

The Parcus Synd-EZ SS is intended to provide fixation during the healing process following syndesmotic trauma, such as fixation of the syndesmosis (syndesmosis disruptions) in connection with Weber B and Weber C ankle fractures.

Type of Use *(Select one or both, as applicable)*

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

510(k) Owner & Submitter: Parcus Medical, LLC
6423 Parkland Dr
Sarasota, FL 34243

Company Contact: Vasavi Vora
Phone: 781-457-9237
Fax: 781-305-9720

Date Prepared: July 26th 2022

Device Trade Name: Parcus Synd-EZ SS

Common Name: Fastener, Fixation, Non-Degradable, Soft Tissue

Device Class: Class II

Classification Name: 21 CFR 888.3040 - Product Code MBI

Predicate Device:

K191783 (cleared October 9, 2019) – Parcus Synd-EZ Ti
Secondary Predicate Device: K192964 – Parcus Synd-EZ SS

Device Description:

The Parcus Synd-EZ SS consists of two buttons with suture connecting them and are designed to be used for stabilizing the syndesmosis during healing. Building upon the technology of the Parcus Synd-EZ Ti, the Synd-EZ SS is easily adjusted for length and resists elongation without requiring knots to secure the fixation. The device is made from medical grade titanium, stainless steel and UHMWPE suture and is provided sterile.

Intended Use:

The Parcus Synd-EZ SS is intended to be used as an adjunct in fracture repair involving metaphyseal and periarticular small bone fragments where screws are not indicated, and as adjuncts in external and intramedullary fixation systems involving plates and rods, with fracture braces and casting.

The Parcus Synd-EZ SS is intended to provide fixation during the healing process following syndesmotic trauma, such as fixation of the syndesmosis (syndesmosis disruptions) in connection with Weber B and Weber C ankle fractures.

Substantial Equivalence Summary:

The Synd-EZ SS is very similar to the predicate Parcus Medical Synd-EZ Ti in that they are both utilize titanium and UHMWPE suture, are intended for the same indications and both utilize the same design. LAL testing was conducted on a representative device comprised of similar materials and it was concluded that the Synd-EZ SS does not raise any additional concerns regarding pyrogenicity.

As the purpose of this submission is limited to a labeling change for MR conditional parameters, there is no change to device design, materials or manufacturing processes. Because there was no change to the device, there was no resulting requirement to complete additional sterilization, shelf-life, biocompatibility or mechanical performance tests.

Summary Performance Data:

The Synd-EZ SS was evaluated for use in the MR Environment and were determined to fit the definition of MR Conditional. This implant was evaluated based on the FDA Guidance Document – *Establishing Safety and Compatibility of Passive Implants in the Magnetic Resonance (MR) Environment* and applicable ASTM standards, and worst-case device was selected for testing. The device was tested for magnetically induced force, magnetically induced torque, heating by RF fields, and image artifact. The results of this testing have been used to establish the MR system conditions suitable for safe use when this implant is present.

No new testing was completed, apart from the testing required to support MR Conditional parameters that was provided in this submission.