



May 9, 2023

Alphatec Spine, Inc.
Aditya Aiyer
Regulatory Affairs Specialist
1950 Camino Vida Roble
Carlsbad, California 92008

Re: K230721

Trade/Device Name: ATEC Lateral Plate System
Regulation Number: 21 CFR 888.3060
Regulation Name: Spinal Intervertebral Body Fixation Orthosis
Regulatory Class: Class II
Product Code: KWQ
Dated: March 15, 2023
Received: March 15, 2023

Dear Aditya Aiyer:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin O'Neill -S



Colin O'Neill, M.B.E.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K230721

Device Name
ATEC Lateral Plate System

Indications for Use (Describe)

The ATEC Lateral Plate System is intended for use via the lateral or anterolateral surgical approach above the bifurcation of the great vessels in the treatment of thoracic and thoracolumbar (T1-L5) spine instability or via the anterior surgical approach, below the bifurcation of the great vessels in the treatment of lumbar and lumbosacral (L1- S1) spine instability as a result of fracture (including dislocation and subluxation), tumor, degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), pseudoarthrosis, spondylolysis, spondylolisthesis, scoliosis, lordotic deformities of the spine, spinal stenosis, or a failed previous spine surgery.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

This 510(k) summary of safety and effectiveness is being submitted in accordance with the requirements of 21 CFR 807.92.

I. SUBMITTER: Alphatec Spine, Inc.
 1950 Camino Vida Roble
 Carlsbad, CA 92008
 Phone: (760) 431-9286
 Fax: (760) 431-0289

Contact Person: Aditya Aiyer
 Regulatory Affairs Specialist
 Contact Phone: (760) 494-6671

Date Summary Prepared: March 15, 2023

II. DEVICE

Name of Device: ATEC Lateral Plate System
 Common or Usual Name: Spinal Intervertebral Body Fixation Orthosis
 Classification Name: Spinal Intervertebral Body Fixation Orthosis
 (21 CFR 888.3060)
 Regulatory Class: Class II
 Product Code: KWQ

III. LEGALLY MARKETED PREDICATE DEVICES

510(k)	Product Code	Trade Name	Manufacturer
Primary Predicate Device			
K202624	KWQ	Z-Span Plate System	Zavation
Additional Predicate/Reference Devices			
K092108	KWQ	TRUSS™ Thoracolumbar Plate System	Globus
K180166	KWQ	Life Spine Lumbar Fixation System (SENTRY®)	Life Spine
K221926	PML, NKB, KWP, KWQ	Invictus® Bone Cement, Invictus Spinal Fixation System	Alphatec Spine
K182808	KWQ	Aspida Anterior Lumbar Plating System	Alphatec Spine

IV. DEVICE DESCRIPTION

The ATEC Lateral Plate System is implanted through a lateral or anterolateral surgical approach to the thoracolumbar spine to provide temporary stabilization until fusion has occurred. The system will be comprised of both implants manufactured from titanium alloy (Ti-6Al-4V ELI) per ASTM F136 for fixation and the required instruments to place them.

ATEC Lateral Plate System implants will be provided in a range of sizes and lengths used to create a construct to immobilize the spinal segments.

The purpose of this submission is to seek initial clearance for the ATEC Lateral Plate System.

V. INDICATIONS FOR USE

The ATEC Lateral Plate System is intended for use via the lateral or anterolateral surgical approach above the bifurcation of the great vessels in the treatment of thoracic and thoracolumbar (T1-L5) spine instability or via the anterior surgical approach, below the bifurcation of the great vessels in the treatment of lumbar and lumbosacral (L1- S1) spine instability as a result of fracture (including dislocation and subluxation), tumor, degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), pseudoarthrosis, spondylolysis, spondylolisthesis, scoliosis, lordotic deformities of the spine, spinal stenosis, or a failed previous spine surgery

VI. TECHNOLOGICAL COMPARISON TO PREDICATES

The technological design features of the subject implants were compared to the predicates in intended use, indications for use, design, function and technology and it was demonstrated that they are substantially equivalent.

VII. PERFORMANCE DATA

Nonclinical testing performed on the subject ATEC Lateral Plate System, supports substantial equivalence to the other predicate devices. The following testing was performed:

- Static Axial Compression Bending testing per ASTM F1717
- Dynamic Axial Compression Bending testing per ASTM F1717
- Static Torsion testing per ASTM F1717
- Static Screw Push-out testing

The results demonstrate that the subject ATEC Lateral Plate System is substantially equivalent to other predicate devices for nonclinical testing.

Clinical Information

Not applicable; determination of substantial equivalence is not based on an assessment of clinical performance data.

VIII. CONCLUSION

Based upon the information provided in this 510(k) submission it has been determined that the subject devices are substantially equivalent to legally marketed devices in regard to indications for use, intended use, design, technology, and performance.