



June 9, 2023

Shanxi Hongjin Plastic Technology Co., Ltd.
% Kathy Liu
Project Manager
Hongray(USA) Medical Products Inc.
3973 Schaefer Avenue
Chino, California 91710

Re: K230779

Trade/Device Name: Powder Free Nitrile Examination Gloves (Black), Tested for Use with
Chemotherapy Drugs and Fentanyl Citrate

Regulation Number: 21 CFR 880.6250

Regulation Name: Non-powdered patient examination glove

Regulatory Class: Class I, reserved

Product Code: LZA, LZC, OPJ, QDO

Dated: March 20, 2023

Received: March 21, 2023

Dear Kathy Liu:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,


Bifeng Qian -S

Bifeng Qian, M.D., Ph.D.

Assistant Director

DHT4B: Division of Infection Control
and Plastic Surgery Devices

OHT4: Office of Surgical
and Infection Control Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K230779

Device Name

Powder Free Nitrile Examination Gloves (Black), Tested for Use with Chemotherapy Drugs and Fentanyl Citrate

Indications for Use (Describe)

The glove is a disposable device intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner.

Gloves have been tested for use with chemotherapy drugs and Fentanyl Citrate using ASTM D6978-05(2019)

Chemotherapy Drug	Minimum Breakthrough Detection Time (BDT) in Minutes
Carmustine 3.3 mg/ml (3,300 ppm)	11.3
Cisplatin 1mg/ml (1,000 ppm)	>240
Cyclophosphamide 20mg/ml (20,000 ppm)	>240
Dacarbazine 10 mg/ml (10,000 ppm)	>240
Doxorubicin HCL, 2 mg/ml (2,000 ppm)	>240
Etoposide, 20 mg/ml (20,000 ppm)	>240
Fluorouracil, 50mg/ml (50,000ppm)	>240
Methotrexate, 25mg/ml (25,000ppm)	>240
Paclitaxel, 6mg/ml (6,000ppm)	>240
Thiotepa, 10mg/ml (10,000ppm)	24.7
Fentanyl Citrate Injection (100 mcg/2ml)	>240

Please note that the following drugs have extremely low permeation times:

Carmustine: 11.3 minutes, Thio Tapa: 24.7 minutes

*Warning: Do not use with Carmustine and Thio Tapa.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Shanxi Hongjin Plastic Technology Co., Ltd.

Coal Bed Gas Industrial Zone, Qu'e Town, Daning County,
Linfen City, Shanxi Province

510(K) SUMMARY

The assigned 510(K) numbers: K230779

Date Prepared: June 08, 2023

1. Owner's Identification:

Mr. Wu Zhigang
Shanxi Hongjin Plastic Technology Co., Ltd.
Coal Bed Gas Industrial Zone, Qu'e Town, Daning County,
Linfen City, Shanxi Province
Email: fdareg@hongray.com.cn

Contact: Ms. Kathy Liu, Project Manager
Address: 3973 Schaefer Avenue, Chino, CA 91710, USA
Tel: 909-590-1611
Email: kathyliu@hongrayusa.com

2. Name of the Device:

Trade / Product Name: Powder Free Nitrile Examination Gloves (Black), Tested for Use with Chemotherapy Drugs and Fentanyl Citrate
Common Name: Exam Gloves
Classification Name: Patient Examination Glove Specialty
Classification Regulation: 21 CFR 880.6250
Product Code: LZA, LZC, QDO, OPJ
Classification Panel: General Hospital
Device Class: Class I

3. Predicate Device Information:

Comfort Rubber Gloves Industries Sdn. Bhd.
Blue Colored, Powder Free Nitrile Examination Gloves Tested for Use with Chemotherapy Drugs and Fentanyl Citrate (K192954)

4. Device Description:

Powder Free Nitrile Examination Gloves (Black), Tested for Use with Chemotherapy Drugs and Fentanyl Citrate Are Class I Patient Examination Gloves and Specialty Chemotherapy Gloves. They are ambidextrous and come in different sizes – Extra Small, Small, Medium, Large, Extra Large and XXL.
Gloves meet the specification of ASTM D6319-19 and have been tested for resistance to permeation by chemotherapy drugs and Fentanyl Citrate as per ASTM D6978-05(2019). The gloves are single use, disposable, and provided non-sterile.

5. Indications for Use:

The Glove is a disposable device intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner.
Gloves have been tested for use with chemotherapy drugs and Fentanyl Citrate using ASTM D6978-05(2019)
The following chemicals have been tested with these gloves:

Shanxi Hongjin Plastic Technology Co., Ltd.

Coal Bed Gas Industrial Zone, Qu'e Town, Daning County,
Linfen City, Shanxi Province

Chemotherapy Drug	Minimum Breakthrough Detection Time (BDT) in Minutes
Carmustine 3.3 mg/ml (3,300 ppm)	11.3
Cisplatin 1mg/ml (1,000 ppm)	>240
Cyclophosphamide 20mg/ml (20,000 ppm)	>240
Dacarbazine 10 mg/ml (10,000 ppm)	>240
Doxorubicin HCL, 2 mg/ml (2,000 ppm)	>240
Etoposide, 20 mg/ml (20,000 ppm)	>240
Fluorouracil, 50mg/ml (50,000ppm)	>240
Methotrexate, 25mg/ml (25,000ppm)	>240
Paclitaxel, 6mg/ml (6,000ppm)	>240
Thiotepa, 10mg/ml (10,000ppm)	24.7
Fentanyl Citrate Injection (100 mcg/2ml)	>240

* Please note that the following drugs have extremely low permeation times:

Carmustine: 11.3 minutes, Thiotepa: 24.7minutes, Or

*Warning: Do not use with Carmustine and Thiotepa.

6. Comparison of Subject Device and Predicate Device:

General Comparison Table:

	Subject Device K230779	Predicate Device K192954	Comparison
Trade Name	Powder Free Nitrile Examination Gloves (Black), Tested for Use with Chemotherapy Drugs and Fentanyl Citrate	Blue Colored, Powder Free Nitrile Examination Gloves Tested for Use with Chemotherapy Drugs and Fentanyl Citrate	Similar
Product Code	LZA, LZC, QDO, OPJ	LZA, LZC, QDO	Same
Regulation Number	21 CFR 880.6250	21 CFR 880.6250	Same
Class	I	I	Same
Indications for Use	Powder Free Nitrile Examination Gloves (Black), Tested for Use with Chemotherapy Drugs and Fentanyl Citrate is a disposable device intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner. Gloves have been tested for use with chemotherapy drugs and Fentanyl Citrate using ASTM D6978	Blue Colored, Powder Free Nitrile Examination Gloves, Non-sterile, and Tested for Use with Chemotherapy Drugs and Fentanyl Citrate is a patient medical exam glove which is a disposable device intended for medical purpose that is worn on the examiner's hand or finger to prevent contamination between examiner and patient. Glove was tested for use with Chemotherapy Drugs and Fentanyl Citrate as per ASTM D6978	Similar

Shanxi Hongjin Plastic Technology Co., Ltd.

Coal Bed Gas Industrial Zone, Qu'e Town, Daning County,
Linfen City, Shanxi Province

Material	Nitrile	Nitrile	Same
Powder or Powder Free	Powder Free	Powder Free	Same
Color	Black	Blue	Different
Single use	Single use	Single use	Same
Chemotherapy Drugs and Fentanyl Citrate Claim	See below comparison table	See below comparison table	See below comparison table

Technological Characteristic Comparison Table:

Technological Characteristics	Subject Device K230779	Predicate Device K192954	Comparison
Length	Minimum 230mm	Minimum 240mm	Similar
Palm Width (size) (mm)			
XS	70±10	70±10	Same
S	80±10	80±10	Same
M	95±10	95±10	Same
L	110±10	110±10	Same
XL	120±10	120±10	Same
XXL	130±10	N/A	Different
Thickness(mm)			
Finger	Minimum 0.05	Minimum 0.05	Same
Palm	Minimum 0.05	Minimum 0.05	Same
Tensile Strength, Before Aging	14MPa, min	14MPa, min	Same
Ultimate Elongation, Before Aging	500%, min	500%, min	Same
Tensile Strength, After Accelerated Aging	14MPa, min	14MPa, min	Same
Ultimate Elongation, After Accelerated Aging	400%, min	400%, min	Same
Watertight (1000ml)	21 CFR 800.20 ASTM D5151 AQL 2.5	21 CFR 800.20 ASTM D5151 AQL 2.5	Same
Powder-Content	≤ 2 mg per glove	≤ 2 mg per glove	Same
10993-23:2021 Skin Irritation Study	Under the conditions of the study, not an irritant	Under the conditions of the study, the subject device is non-irritating	Same
10993-10:2021 Maximization Sensitization Study	Under the conditions of the study, not a sensitizer	Under the conditions of the study, the subject device is non sensitization	Same

Shanxi Hongjin Plastic Technology Co., Ltd.

Coal Bed Gas Industrial Zone, Qu'e Town, Daning County,
Linfen City, Shanxi Province

10993-5:2009 Cytotoxicity Test	Under the conditions of this study, the test article extract showed potential toxicity to L929 cells.	Exhibits severe cytotoxicity reactivity at 100%, and 66% extract concentrations and no cytotoxicity reactivity at 44%, 30%, 20% and 15% extract concentrations under the condition of this test.	Similar
ISO 10993-11:2017 Acute Systemic toxicity study	Under the conditions of this study, there was no evidence of systemic toxicity.	Under the conditions of the study, the subject showed no adverse biological reaction.	Same

Chemotherapy Permeation and Fentanyl Citrate Comparison Claim:

Tested Chemotherapy Drug and Concentration	Minimum BDT (Minutes)		Comparison
	Subject Device K230779	Predicate Device K192954	
Carmustine 3.3 mg/ml (3,300 ppm)	11.3	18.2	Similar
Cisplatin 1mg/ml (1,000 ppm)	>240	>240	Same
Cyclophosphamide 20mg/ml (20,000 ppm)	>240	>240	Same
Dacarbazine 10 mg/ml (10,000 ppm)	>240	>240	Same
Doxorubicin HCL, 2 mg/ml (2,000 ppm)	>240	>240	Same
Etoposide, 20 mg/ml (20,000 ppm)	>240	>240	Same
Fluorouracil, 50mg/ml (50,000ppm)	>240	>240	Same
Methotrexate, 25mg/ml (25,000ppm)	>240	N/A	Different
Paclitaxel, 6mg/ml (6,000ppm)	>240	>240	Same
Thiotepa, 10mg/ml (10,000ppm)	24.7	57.3	Similar
Fentanyl Citrate Injection (100 mcg/2ml)	>240	>240	Same

7. Summary of Non-Clinical Performance Data

Non-clinical tests were conducted to verify that the proposed device met all design specifications. The test results demonstrated that the proposed device met the performance criteria with the following standards:

Methodology	Test Performed	Acceptance Criteria	Results
ASTM D6319- 19	Physical Dimensions Length	Minimum 230mm for all sizes	Pass
ASTM D6319- 19	Physical Dimensions Palm Width	XS: 70±10mm S: 80±10mm M: 95±10mm L:110±10mm XL: 120±10mm XXL: 130±10mm	Pass
ASTM D6319- 19	Physical Dimensions Thickness	Finger: 0.05mm (min) Palm: 0.05mm (min)	Pass
ASTM D6319- 19 ASTM D412-16(2021)	Physical Properties	Tensile Strength (Min14 MPa) and Elongation (Before Aging 500% and after aging 400%) Min	Pass

Shanxi Hongjin Plastic Technology Co., Ltd.

Coal Bed Gas Industrial Zone, Qu'e Town, Daning County,
Linfen City, Shanxi Province

ASTM D6319- 19 ASTM D5151-19	Water leak test	AQL 2.5 (ISO 2859-1)	Pass
ASTM D6319- 19 ASTM D6124-06 (2017)	Powder Residue	Max 2mg/glove	Pass
ASTM D6978-05 (2019)	Permeation by Chemotherapy Drugs	Refer the above table	Pass
ISO 10993-10 &23:2021	Irritation and Skin Sensitization	Skin sensitization and Skin irritation	Is non-sensitization and Non-irritation
ISO 10993-5:2009	Cytotoxicity	Cytotoxicity reactivity	showed potential toxicity to L929 cells.
ISO 10993-11:2017	Acute systemic toxicity study	Subject showed no adverse biological reaction	no evidence of systemic toxicity.

- ASTM D6319-19, Standard Specification for Nitrile Examination Gloves for Medical Application.
- ASTM D5151-19, Standard Test Method for Detection of Holes in Medical Gloves.
- ASTM D6124-06 (Reapproved 2017), Standard Test Method for Residual Powder on Medical Gloves
- ASTM D412-16 (2021) Standard Test Methods for Vulcanized Rubber and Thermoplastic Elastomers—Tension
- ASTM D6978-05 (Reapproved 2019), Assessment of Reissuance of Medical Gloves to Permeation by Chemotherapy Drugs.
- ISO 10993-10:2021 Biological Evaluation of Medical Devices - Part 10: Tests For Skin Sensitization.
- ISO 10993-5:2009 Biological Evaluation of Medical Devices - Part 5: Tests For In Vitro Cytotoxicity
- ISO 10993-23:2021 Biological Evaluation of Medical Devices - Part 10: Tests For Skin Irritation.
- ISO 10993-11:2017 Biological evaluation of medical devices — Part 11: Tests for systemic toxicity

8. Clinical Performance Data

N/A

9. Conclusion:

The conclusions drawn from the nonclinical tests demonstrate that the proposed device is as safe, as effective, and performs as well as or better than the legally marketed predicated device.