



April 6, 2023

Philips Medical Systems Nederland B.V.
% Ioana Ulea
Senior Regulatory Affairs Specialist
Veenpluis 6
5684 PC Best
NETHERLANDS

Re: K223458

Trade/Device Name: Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700
MR Systems

Regulation Number: 21 CFR 892.1000

Regulation Name: Magnetic Resonance Diagnostic Device

Regulatory Class: Class II

Product Code: LNH, LNI

Dated: March 6, 2023

Received: March 6, 2023

Dear Ioana Ulea:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

A handwritten signature in blue ink, appearing to read 'D. Krainak', is written over a large, light blue watermark of the letters 'FDA'.

Daniel M. Krainak, Ph.D.
Assistant Director
DHT8C: Division of Radiological Imaging
and Radiation Therapy Devices
OHT8: Office of Radiological Health
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K223458

Device Name

Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems

Indications for Use (Describe)

Philips Magnetic Resonance (MR) systems are Medical Electrical Systems indicated for use as a diagnostic device.

This MR system enables trained physicians to obtain cross-sectional images, spectroscopic images and/or spectra of the internal structure of the head, body or extremities, in any orientation, representing the spatial distribution of protons or other nuclei with spin.

Image appearance is determined by many different physical properties of the tissue and the anatomy, the MR scan technique applied, and presence of contrast agents. The use of contrast agents for diagnostic imaging applications should be performed consistent with the approved labeling for the contrast agent.

The trained clinical user can adjust the MR scan parameters to customize image appearance, accelerate image acquisition, and synchronize with the patient's breathing or cardiac cycle.

The systems can use combinations of images to produce physical parameters, and related derived images. Images, spectra, and measurements of physical parameters, when interpreted by a trained physician, provide information that may assist diagnosis, and therapy planning. The accuracy of determined physical parameters depends on system and scan parameters and must be controlled and validated by the clinical user.

In addition, the Philips MR systems provide imaging capabilities, such as MR fluoroscopy, to guide and evaluate interventional and minimally invasive procedures in the head, body and extremities. MR Interventional procedures, performed inside or adjacent to the Philips MR system, must be performed with MR Conditional or MR Safe instrumentation as selected and evaluated by the clinical user for use with the specific MR system configuration in the hospital. The appropriateness and use of information from a Philips MR system for a specific interventional procedure and specific MR system configuration must be validated by the clinical user.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary of Safety and Effectiveness

This 510(k) summary of safety and effectiveness information is prepared in accordance with 21 CFR §807.92.

Date Prepared:	Apr. 05, 2023	
Manufacturer:	Philips Medical Systems Nederland B.V. Veenpluis 6, 5684 PC, Best, The Netherlands Establishment Registration Number: 3003768277	
Primary Contact Person:	Ioana Ulea Senior Regulatory Affairs Specialist Phone: +31 618345875 E-mail: ioana.ulea@philips.com	
Secondary Contact Person	Jan van de Kerkhof Associate Director Regulatory Affairs Phone: +31 613300542 E-mail: jan.van.de.kerkhof@philips.com	
Device Name:	Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems	
Classification:	Classification name:	Magnetic Resonance Diagnostic Device (MRDD)
	Classification Regulation:	21CFR 892.1000
	Classification Panel:	Radiology
	Device Class:	Class II
	Primary Product Code:	90LNH 90LNI
Primary Predicate Device:	Trade name:	Achieva, Ingenia, Ingenia CX, Ingenia Elition, and Ingenia Ambition MR Systems (R11)
	Manufacturer:	Philips Medical Systems Nederland B.V.
	510(k) Clearance:	K213583
	Classification Regulation:	21CFR 892.1000
	Classification name:	Magnetic Resonance Diagnostic Device (MRDD)
	Classification Panel:	Radiology
	Device class	Class II
	Product Code:	90LNH 90LNI
	Trade name:	MR 5300 and MR 7700 R11 MR Systems

Secondary Predicate Device:	Manufacturer:	Philips Medical Systems Nederland B.V.
	510(k) Clearance:	K223442
	Classification Regulation:	21CFR 892.1000
	Classification name:	Magnetic Resonance Diagnostic Device (MRDD)
	Classification Panel:	Radiology
	Device class	Class II
	Product Code:	90LNH 90LNI
Reference Device	Trade name:	SIGNA Premier
	Manufacturer:	GE Medical Systems, LLC
	510(k) Clearance:	K193282
	Classification Regulation:	21CFR 892.1000
	Classification name:	Magnetic Resonance Diagnostic Device (MRDD)
	Classification Panel:	Radiology
	Device class	Class II
	Product Code:	90LNH 90LNI

Device Description:	<p>The proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems R12 are 60 cm and 70 cm bore 1.5 and 3.0 Tesla (1.5T and 3.0T) Magnetic Resonance Diagnostic Devices, hereafter to be known as Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems.</p> <p>This bundled abbreviated 510(k) submission, is prompted by the introduction of a new optional software feature called Precise Image contained in software R12 for the proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems, as compared to our legally marketed primary predicate device <i>Achieva, Ingenia, Ingenia CX, Ingenia Elition, and Ingenia Ambition MR Systems (R11)</i> (K213583) and the secondary predicate device <i>MR 5300 and MR 7700 R11 MR Systems</i> (K223442).</p> <p>Precise Image is a deep learning based reconstruction technique designed to increase signal-to-noise ratio (SNR), increase sharpness and decrease ringing artefacts from MR images.</p> <p>The introduction of Precise Image only required updates to the MR System Software.</p> <p>The proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems are intended to be marketed with</p>
----------------------------	---

	<p>the following pulse sequences and coils that are previously cleared by FDA:</p> <ol style="list-style-type: none"> 1. mDIXON (K102344) 2. SWIp (K131241) 3. mDIXON-Quant (K133526) 4. MRE (K140666) 5. mDIXON XD (K143128) 6. O-MAR (K143253) 7. 3D APT (K172920) 8. Compatible System Coils <p>The accessories to be used with the proposed device Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems have not changed compared to the <i>primary predicate device</i> and <i>secondary predicate device</i> and can be found in the Instructions for Use accompanying the device:</p> <ul style="list-style-type: none"> • System coils • PPU Sensor for wireless physiology • Pediatric PPU Sensor • FlexTrak trolleys (FlexTrak / HA FlexTrak II) • Acoustic Hood • MR Elastography <p>When Philips MRI system is used in combination with the Philips MR-RT or MR-OR solutions, the user is referred to the dedicated MR-RT and MR-OR Instructions for Use for information on additional accessories that may apply:</p> <ul style="list-style-type: none"> • Flextrak OR • MR-RT CouchTop • RT CouchTop XD
<p>Indications for Use:</p>	<p>The indications for Use statement provided below for the proposed device is identical to the one of the primary and secondary predicate devices. The intended use is also not impacted by the introduction of Precise Image feature.</p> <p><i>Philips Magnetic Resonance (MR) systems are Medical Electrical Systems indicated for use as a diagnostic device.</i></p> <p><i>This MR system enables trained physicians to obtain cross-sectional images, spectroscopic images and/or spectra of the internal structure of the head, body or extremities, in any orientation, representing the spatial distribution of protons or other nuclei with spin.</i></p> <p><i>Image appearance is determined by many different physical properties of the tissue and the anatomy, the MR scan technique applied, and presence of contrast agents.</i></p>

	<p><i>The use of contrast agents for diagnostic imaging applications should be performed consistent with the approved labeling for the contrast agent.</i></p> <p><i>The trained clinical user can adjust the MR scan parameters to customize image appearance, accelerate image acquisition, and synchronize with the patient's breathing or cardiac cycle. The systems can use combinations of images to produce physical parameters, and related derived images. Images, spectra, and measurements of physical parameters, when interpreted by a trained physician, provide information that may assist diagnosis and therapy planning. The accuracy of determined physical parameters depends on system and scan parameters and must be controlled and validated by the clinical user.</i></p> <p><i>In addition, the Philips MR systems provide imaging capabilities, such as MR fluoroscopy, to guide and evaluate interventional and minimally invasive procedures in the head, body and extremities. MR Interventional procedures, performed inside or adjacent to the Philips MR system, must be performed with MR Conditional or MR Safe instrumentation as selected and evaluated by the clinical user for use with the specific MR system configuration in the hospital. The appropriateness and use of information from a Philips MR system for a specific interventional procedure and specific MR system configuration must be validated by the clinical user.</i></p>
<p>Design Features/ Fundamental Scientific Technology:</p>	<p>Just as the primary and secondary predicate devices, the proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems are based on the principle that certain atomic nuclei present in the human body will emit a weak relaxation signal when placed in a strong magnetic field and excited by a radio signal at the precession frequency. The emitted relaxation signals are analyzed by the system and a computed image reconstruction is displayed on a video screen.</p> <p>The principal technological components (magnet, transmit body coil, gradient coil, gradient amplifier, RF amplifier and patient support) of the proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems are unchanged compared to primary predicate (K213583) and the secondary predicate device (K223442).</p> <p>The reconstruction software of the proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems has been modified to include the Precise Image, offering an alternative to the current way of increasing resolution, available on the primary and secondary predicate devices. The resulting images can have higher SNR, increased sharpness and decreased ringing artefacts compared to primary and secondary predicate devices. The user interface provides operators of the system with new options for selecting Precise Image feature and adjusting the associated level of image noise reduction.</p>

<p>Summary of Non-Clinical Performance Data:</p>	<p>Non-clinical performance testing has been performed on the proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems and demonstrates compliance with following international and FDA-recognized consensus standards:</p>			
	No.	Recognition Number	Standard Number and Date	Standard Name
	1	12-295	IEC60601-2-33 Ed. 3.2:2010 + Amd 1:2013 + Amd 2:2015	<i>Medical electrical equipment - Part 2-33: Particular requirements for the basic safety and essential performance of magnetic resonance equipment for medical diagnosis</i>
	2	19-4	ANSI / AAMI ES60601-1:2005/(R)2012 and A1:2012	<i>Medical Electrical Equipment - Part 1: General Requirements For Basic Safety And Essential Performance (IEC 60601-1:2006, MOD).</i>
	3	19-8	IEC60601-1-2 Ed. 4.0:2014	<i>Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic disturbances - Requirements and tests</i>
	4	5-89	IEC 60601-1-6 Ed. 3.1:2010 + Amd 1:2013	<i>Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral Standard: Usability</i>
	5	5-76	IEC 60601-1-8 Ed. 2.1:2006 + Amd 1:2012 (Ed.2.1)	<i>Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance - Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems</i>
	6	5-125	ISO 14971 Third Edition 2019	<i>Medical devices – Application of risk management to medical devices.</i>
	7	5-114	IEC 62366-1 Ed. 1.0:2015	<i>Medical devices – Part 1: Application of usability engineering to medical devices</i>
	8	13-79	IEC 62304 Ed. 1.1:2015	<i>Medical device software – Software life cycle processes.</i>
<p>Non-Clinical verification and validation tests have been performed with regards to the intended use, the technical claims, the requirement specifications and the risk management results.</p> <p>Simulated use validation was executed to obtain evidence of Precise image with respect to its predicate by qualitative review performed by application specialists. Based on the results of the qualitative image</p>				

	<p>review, reproducibility of Precise in comparison to predicate device reconstruction technique can be considered established. Additionally, a reader evaluation by ABR board certified radiologists was performed, where following properties have been analyzed as part of the comparison: signal-to-noise ratio (SNR), artifact level, sharpness, and contrast-to-noise ratio (CNR). Furthermore, the radiologists have assessed the quality of the visualization of abnormalities and pathologies in case they were present in the images as well as if the images were of sufficient quality for diagnostic purposes. The review evaluation shows that the proposed device is assessed as equivalent for diagnosis and holds significantly better SNR and sharpness compared to the predicate reconstruction technology, also in the presence of (subtle) abnormalities and pathology.</p> <p>The verification and/or validation test results demonstrate that the proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems meet the acceptance criteria and are adequate for the intended use.</p> <p>The risk management activities show that all risks are sufficiently mitigated, that no new risks are introduced, and that the overall residual risks are acceptable.</p> <p>Therefore, the proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems are substantially equivalent to the legally marketed primary predicate device <i>Achieva, Ingenia, Ingenia CX, Ingenia Elition, and Ingenia Ambition MR Systems (R11)</i> (K213583, 16/05/2022) and the secondary predicate device <i>MR 5300 and MR 7700 R11 MR Systems</i> (K223442) in terms of safety and effectiveness.</p>
<p>Summary of Clinical Data:</p>	<p>With the proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems the indications for use remain unchanged and there were no technological characteristics relative to the primary and secondary predicate device that would require clinical testing.</p>
<p>Substantial Equivalence:</p>	<p>The proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems and the legally marketed primary predicate device <i>Achieva, Ingenia, Ingenia CX, Ingenia Elition, and Ingenia Ambition MR Systems (R11)</i> (K213583, 16/05/2022) and the secondary predicate device <i>MR 5300 and MR 7700 R11 MR Systems</i> (K223442), have the same indications for use with respect to the following:</p> <ul style="list-style-type: none"> • Providing cross-sectional images based on the magnetic resonance phenomenon • Interpretation of the images is the responsibility of trained physicians • Images can be used for interventional and treatment planning purposes
<p>Conclusion:</p>	<p>The proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems are substantially equivalent to the legally marketed primary predicate device <i>Achieva, Ingenia, Ingenia CX, Ingenia Elition, and Ingenia Ambition MR Systems (R11)</i> (K213583) and the secondary predicate device <i>MR 5300 and MR 7700 R11 MR Systems</i></p>

	<p>(K223442), in terms of design features, fundamental scientific technology, indications for use, and safety and effectiveness.</p> <p>Additionally, substantial equivalence is demonstrated with non-clinical performance (verification and validation) tests, which complied with the requirements specified in the international and FDA-recognized consensus standards and device-specific guidance.</p> <p>The results of these tests demonstrate that the proposed Ingenia, Ingenia CX, Ingenia Elition, Ingenia Ambition, MR 5300 and MR 7700 MR Systems meet the acceptance criteria and are adequate for the intended use.</p>
--	---