



January 29, 2021

Aegis Spine, Inc.
Meredith Lee May, MS, RAC
Director of Consulting
Empirical Testing Corp.
4628 Northpark Drive
Colorado Springs, Colorado 80918

Re: K203531

Trade/Device Name: XYSPAN Expandable Lumbar Cage System
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: MAX
Dated: November 24, 2020
Received: December 2, 2020

Dear Ms. May:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Brent L. Showalter, Ph.D.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K203531

Device Name

XYPAN Expandable Lumbar Cage System

Indications for Use (Describe)

XYPAN Expandable Lumbar Cage System is indicated for intervertebral body fusion procedures in skeletally mature patients with degenerative disc disease (DDD) at one or two contiguous levels from L2-S1. DDD is defined as discogenic back pain with degeneration of the disc confirmed by patient history and radiographic studies. These DDD patients may also have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level(s). This device is to be used with autogenous bone graft and/or allogeneous bone graft composed of cancellous and/or corticocancellous bone. XYPAN Expandable Lumbar Cage System is to be used with supplemental fixation. Patients should have at least six (6) months of non-operative treatment prior to treatment with an intervertebral cage.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D) Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

510(K) SUMMARY

Submitter's Name:	Aegis Spine, Inc.
Submitter's Address:	9781 S. Meridian Blvd, Ste 300 Englewood, CO 80112
Submitter's Telephone:	+1.303.741.4123
Contact Person:	Meredith Lee May MS, RAC Empirical Testing Corp. 719.337.7579 MMay@EmpiricalTech.com
Date Summary was Prepared:	24Nov2020
Trade or Proprietary Name:	XYPAN Expandable Lumbar Cage System
Common or Usual Name:	intervertebral fusion device with bone graft, lumbar
Classification:	Class II per 21 CFR §888.3080
Product Code:	MAX
Classification Panel:	Orthopedics

DESCRIPTION OF THE DEVICE SUBJECT TO PREMARKET NOTIFICATION:

The XYPAN Lumbar Expandable Cage System's implants are interbody fusion devices intended for use as an aid in spinal fixation. They are made of Titanium 6AL-4V Alloy (ASTM F136). These hollow, rectangular implants are offered in a variety of widths, lengths, heights and lordotic angles designed to adapt to a variety of patient anatomies. The implants can be expanded in height after insertion in the unexpanded state using the system instrumentation. The implants have serrations on the superior and inferior surfaces designed for fixation.

- XYPAN-XT Expandable Cage is to be implanted via transforaminal and posterior approach.
- XYPAN-XTP Expandable Cage is to be implanted via Anterior to Psoas approach.
- XYPAN-XL Expandable Cage is to be implanted via lateral approach.

INDICATIONS FOR USE

XYPAN Expandable Lumbar Cage System is indicated for intervertebral body fusion procedures in skeletally mature patients with degenerative disc disease (DDD) at one or two contiguous levels from L2-S1. DDD is defined as discogenic back pain with degeneration of the disc confirmed by patient history and radiographic studies. These DDD patients may also have up to Grade 1 spondylolisthesis or retrolisthesis at the involved level(s). This device is to be used with autogenous bone graft and/or allogeneous bone graft composed of cancellous and/or corticocancellous bone. XYPAN Expandable Lumbar Cage System is to be used with supplemental fixation. Patients should have at least six (6) months of non-operative treatment prior to treatment with an intervertebral cage.

TECHNOLOGICAL CHARACTERISTICS

XYPAN Expandable Lumbar Cage System is made from Titanium 6AL-4V alloy that conforms to ASTM F136. The subject and predicate device have identical technological characteristics. Specifically, the following characteristics are identical between the subject and predicates:

- Indications for Use
- Materials of manufacture
- Structural support mechanism

Table 5-1 Predicate device

510k Number	Trade or Proprietary or Model Name	Manufacturer	Predicate Type
K190708	AccelFix Lumbar Interbody Fusion Cage System	L&K Biomed Co., Ltd.	Primary

PERFORMANCE DATA

Engineering rationales and comparisons between the subject and predicate devices were used to leverage previously conducted performance testing.

CONCLUSION

The overall technology characteristics and mechanical performance data lead to the conclusion that the XYPAN Expandable Lumbar Cage System is substantially equivalent to the predicate device.