



March 11, 2021

Auris Health Inc., a Johnson and Johnson Family Company  
Somi Ekwealor  
Staff Regulatory Affairs Analyst  
150 Shoreline Drive  
Redwood City, California 94065

Re: K203614

Trade/Device Name: Monarch Bronchoscope  
Regulation Number: 21 CFR 874.4680  
Regulation Name: Bronchoscope (Flexible Or Rigid) And Accessories  
Regulatory Class: Class II  
Product Code: QNW  
Dated: December 8, 2020  
Received: December 14, 2020

Dear Somi Ekwealor:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shu-Chen Peng, Ph.D.  
Assistant Director  
DHT1C: Division of ENT, Sleep Disordered  
Breathing, Respiratory and  
Anesthesia Devices  
OHT1: Office of Ophthalmic, Anesthesia,  
Respiratory, ENT and Dental Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K203614

Device Name  
Monarch Bronchoscope

Indications for Use (Describe)

The reprocessed Monarch Bronchoscope, used in conjunction with the Monarch Platform, is intended to provide bronchoscopic visualization of and access to patient airways for diagnostic and therapeutic procedures.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

---

---

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

---

---

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

K203614

**510(k) Summary****General Information**

<b>510(k) Submitter</b>	Auris Health, Inc., a Johnson and Johnson Family Company 150 Shoreline Drive Redwood City, CA 94065
<b>Original Device Manufacturer</b>	Auris Health, Inc., a Johnson and Johnson Family Company 150 Shoreline Drive Redwood City, CA 94065
<b>Reprocessed Device Manufacturer</b>	Auris Health, Inc., a Johnson and Johnson Family Company 2383 Bering Dr. San Jose, CA 95131
<b>FDA Registration Number</b>	3014447948
<b>Primary Correspondent</b>	Somi Ekwealor, MSRS, RAC Staff Regulatory Affairs Specialist Johnson and Johnson, Robotics & Digital Solutions
<b>Contact Information</b>	Email: sekwealo@its.jnj.com Phone: (408) 320-5385
<b>Date Prepared</b>	08 December 2020

**Device Identification****Proposed Reprocessed Device:**

<b>Proprietary Name</b>	Monarch Bronchoscope
<b>Common Name</b>	Reprocessed Bronchoscope (Flexible Or Rigid)
<b>Classification Name</b>	Bronchoscope (flexible or rigid) and accessories
<b>Regulation Number</b>	21 CFR 874.4680
<b>Product Code</b>	QNW
<b>Regulatory Class</b>	II
<b>Model Number</b>	MBR-000211-B

**Predicate OEM Device:**

<b>Proprietary Name</b>	Monarch Platform
<b>Common Name</b>	Bronchoscope (Flexible Or Rigid)
<b>Premarket Notification</b>	K193534
<b>Classification Name</b>	Bronchoscope (flexible or rigid) and accessories
<b>Regulation Number</b>	21 CFR 874.4680
<b>Product Code</b>	EOQ
<b>Regulatory Class</b>	II
<b>Model Number</b>	MBR-000211-A

## Device Description

The reprocessed Monarch Bronchoscope, MBR-000211-B, (hereafter referred to as “Proposed Device”) is identical to the Original Monarch Bronchoscope, MBR-000211-A, (hereafter referred to as “Predicate Device”), which is a component of, and must be used with, the Monarch Platform, cleared under K193534. The Monarch Bronchoscope is connected to the robotic arms of the Monarch Platform to provide bronchoscopic visualization of and access to patient airways for diagnostic and therapeutic procedures. The Proposed Device is a reprocessed single-use device that can be reprocessed up to one (1) time.

The Monarch Bronchoscope, consisting of the Inner Scope (“scope”) and the Outer Sheath (“sheath”), has 4-way articulation controlled by continuous, direct, visual control of the physician using the Monarch Platform. The Proposed Device contains a working channel through which biopsy devices, or other working channel instruments, may be introduced. The distal tip of the Proposed Device has a camera control unit (CCU) that collects live images that are then transmitted to the physician’s display interface of the Monarch Platform. The camera transmits vision data to the Monarch Tower through the camera cable. The single-use, manually controlled, working channel instruments compatible with the Proposed Device are identical to the working channel instruments compatible with the predicate Monarch Platform bronchoscope.

## Intended Use/Indications for Use

The reprocessed Monarch Bronchoscope, used in conjunction with the Monarch Platform, is intended to provide bronchoscopic visualization of and access to patient airways for diagnostic and therapeutic procedures.

## Summary of Technological Characteristics

The Proposed Device (MBR-000211-B) is a single-use device identical to the Predicate Device (MBR-000211-A) in design, patient-contacting materials, clinical applications, patient population, performance specifications, and principles of operation. All technological characteristics including articulation, vision, and compatibility with working channel instruments are identical. Key technological characteristics are listed in the table below.

Key Attributes	Proposed Device (MBR-000211-B)
Product Code	QNW
Regulation Number	21 CFR 874.4680
Classification	II
Intended for Single Use	Yes
Field of View (FOV) in air	90 degrees
Direction of view	0 degrees
FOV depth	3-30mm
Imaging type	CMOS Imager
Illumination type	LED

Key Attributes	Proposed Device (MBR-0000211-B)
Active angulation degrees up/down or 4 directions	180/180/180/180
Pixel resolution	200 x 200
Camera lens	Aluminosilicate glass
Light source	LED (covered in cyanoacrylate adhesive)
Working Channel Instruments Compatibility	Auris working channel instruments and third party instruments that meet working channel length and diameter requirements (e.g., REBUS probe)

This premarket notification is submitted to demonstrate the ability to reprocess the Monarch Bronchoscope one (1) time without impacting safety or effectiveness. Auris Health's reprocessing of the device includes removal of adherent soil and decontamination. Each individual device is tested for appropriate function of its components prior to packaging, labeling, and sterilization operations.

### Performance Testing

The Reprocessed Monarch Bronchoscope was tested for performance in accordance with internal design specification and with the applicable performance standards to demonstrate safety and effectiveness. This includes the following tests:

Test Name	Description	Results
Cleaning Validation	Worst case device conditioning was taken into consideration while conducting this testing.	Pass
Functional Performance and System Compatibility	Functional performance and system compatibility were performed to verify the performance of the Proposed Device was not negatively impacted by reprocessing. Worst case device conditioning was taken into consideration while conducting this testing.	Pass
Electromagnetic Compatibility and Electrical Safety	The Proposed Device has been fully evaluated for electrical safety and EMC compliance to the following standards: AAMI/ANSI ES60601-1, IEC 60601-1-2, and IEC 60601-2-18. Worst case device conditioning was taken into consideration while conducting this testing.	Pass
Biocompatibility	Evaluates the end of life biocompatibility of the limited contact device in accordance with ISO 10993-1:2018. Worst case device conditioning was taken into consideration while conducting this testing.	Pass
Sterilization	Sterilization was assessed in accordance with ISO 11135:2018. Sterilization residuals were assessed in accordance with ISO 10993-7:2008. The Proposed Device uses a validated Ethylene oxide sterilization process using a half-cycle overkill approach to achieve a minimum sterility assurance level of $10^{-6}$ .	Pass

The performance testing demonstrates that the reprocessed Monarch Bronchoscope is as safe and effective as the legally marketed Original Monarch Bronchoscope and operates as originally intended.

### **Conclusion**

The indications for use/intended use, patient-contacting materials, clinical applications, patient population, performance specifications, technological characteristics, and principles of operation of the Proposed Device are identical to the Predicate Device.

Performance testing and process validation results demonstrated substantial equivalence of the Proposed Device (reprocessed Monarch Bronchoscope, MBR-000211-B) to the Predicate Device (Original Monarch Bronchoscope, MBR-000211-A) with respect to safety and effectiveness. We therefore conclude that the reprocessed Monarch Bronchoscope is as safe, effective, and substantially equivalent to the Predicate Device, Monarch Bronchoscope.

The following device is included in this submission:

<b>Description</b>	<b>OEM Model Number</b>	<b>Reprocessed Model Number</b>
Monarch Bronchoscope	MBR-000211-A	MBR-000211-B