



April 18, 2023

Sofwave Medical Ltd
% Janice Hogan
Partner
Hogan Lovells LLP
1735 Market Street, Suite 2300
Philadelphia, Pennsylvania 19103

Re: K230820

Trade/Device Name: SofWave System
Regulation Number: 21 CFR 878.4590
Regulation Name: Focused Ultrasound Stimulator System For Aesthetic Use
Regulatory Class: Class II
Product Code: OHV
Dated: March 24, 2023
Received: March 24, 2023

Dear Janice Hogan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Mark

Trumbore -S

Digitally signed by Mark
Trumbore -S

Date: 2023.04.18 12:55:32
-04'00'

Mark Trumbore, Ph.D.

Assistant Director

DHT4A: Division of General Surgery Devices

OHT4: Office of Surgical

and Infection Control Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

510(k) Number (if known)

K230820

Device Name

SofWave System

Indications for Use (Describe)

The SofWave System is indicated for use as a non-invasive dermatological aesthetic treatment to improve facial lines and wrinkles, lift the eyebrow, and lift lax submental (beneath the chin) and neck tissue; which can also affect the appearance of lax tissue in the submental and neck regions for subjects aged 22 and older. The SofWave System is also intended for short-term improvement in the appearance of cellulite.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human
Services Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA)
Staff PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY**Sofwave Medical's SofWave System****Submitter's Name, Address, Telephone Number, Contact Person, and Date Prepared**

Sofwave Medical Ltd.
1 Ha-Otsma St.
Yokneam Ilit,
Israel 2069200

Submission Correspondent:

Janice M. Hogan
Hogan Lovells US LLP
janice.hogan@hoganlovells.com
(267) 675-4611

Date Prepared: March 24, 2023

Name of Device:

SofWave System

Common or Usual Name:

Focused Ultrasound Stimulator System for Aesthetic Use

Classification Name:

21 CFR 878.4590 (Ultrasound for Tissue Heat or Mechanical Cellular Disruption), Class II,
product code OHV

Predicate Device

Sofwave Medical's SofWave System (K230019) (Predicate Device)

Intended Use / Indications for Use

The SofWave System is indicated for use as a non-invasive dermatological aesthetic treatment to improve facial lines and wrinkles, lift the eyebrow, and lift lax submental (beneath the chin) and neck tissue; which can also affect the appearance of lax tissue in the submental and neck regions for subjects aged 22 and older. The SofWave System is also intended for short-term improvement in the appearance of cellulite.

Technological Characteristics

The SofWave System is an ultrasound system intended for aesthetic purposes. The system generates high frequency ultrasonic pulses that elevate the temperature in the dermis layer and cause controlled isolated areas of thermal damage.

The SofWave System consists of two main functional components: 1) the console and 2) the applicator. The console includes the power sources, cooling unit, electrical components, the IoT and the user interface. The applicator is comprised of an array of ultrasonic transducers that emit continuous acoustic waves at 10-12 MHz and an active cooling element that is used to cool the skin area in contact with the applicator. The applicator is connected by a flexible cable to the console.

The SofWave System is a modification to the company's SofWave System that has already been cleared for the same indications for use (K230019). Only minor hardware and software changes have been made, including the addition of Internet of Things (IoT) capability via Wi-Fi. These changes do not alter the fundamental scientific technology of the modified device. Specifically, the energy parameters for the console or the applicators are not changed.

Performance Data

In accordance with SofWave's design control procedures, risk analysis was conducted to assess the impact of the modifications on the SofWave device. Verification testing demonstrates that the modified device functions as specified and is as safe and effective as the cleared predicate device.

Electrical safety testing and Electromagnetic compatibility (EMC) testing was repeated for the SofWave device by an independent test laboratory in accordance with IEC 60601-1-2, IEC 60601-1, IEC 60601-1-6, and IEC 60601-2-62.

Biocompatibility of the patient-contacting components of the device was established based on the evaluation of the predicate device. No new materials have been introduced on the surface of the device that contacts the patient's skin.

Software verification and validation testing was performed. The results were found acceptable for the software changes and demonstrated that the software performs as intended.

Conclusion

The subject SofWave System and its predicate have the same indications and principles of operation, and similar technological characteristics. The minor differences in the newer device version do not present different questions of safety or effectiveness as compared to the predicate device. Verification testing according to the company's design control procedures demonstrates that the subject device is as safe and effective as its predicate device. Thus, the subject SofWave System is substantially equivalent to its predicate device.