



September 11, 2023

MiRus, LLC  
Jordan Bauman  
Vice President, Regulatory Affairs  
1755 West Oak Parkway, Suite 100  
Marietta, Georgia 30062

Re: K232154

Trade/Device Name: MiRus 3DR Lateral Lumbar Interbody Fusion System with Integrated Plate  
Fixation

Regulation Number: 21 CFR 888.3080

Regulation Name: Intervertebral Body Fusion Device

Regulatory Class: Class II

Product Code: OVD, MAX

Dated: July 19, 2023

Received: July 20, 2023

Dear Jordan Bauman:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

  
**Brent Showalter -S**

Brent Showalter, Ph.D.

Assistant Director

DHT6B: Division of Spinal Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K232154

Device Name  
MiRus 3DR Lateral Lumbar Interbody Fusion System with Integrated Plate Fixation

### Indications for Use (Describe)

The MiRus™ 3DR Lumbar Interbody Fusion System is indicated for intervertebral body fusion procedures in skeletally mature patients with degenerative disc disease (DDD) of the lumbar spine at one or two contiguous levels from L1-L2 to L5-S1. DDD is defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies. These DDD patients may also have up to Grade I spondylolisthesis or retrolisthesis at the involved level(s). Devices are to be used with autogenous or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft and supplemental fixation. Patients should have at least six (6) months of non-operative treatment prior to treatment with an intervertebral body fusion device.

When used with or without the MiRus MoRe Lateral Lumbar Plating System, the system is indicated for use with supplemental fixation cleared by the FDA for use in the lumbar spine.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**510(k) Summary**

This 510(k) summary is being submitted in accordance with the requirements of 21 CFR 807.92(c).

<b>I. SUBMITTER</b>	MiRus™, LLC 1755 West Oak Parkway Suite 100 Marietta, Georgia 30062 Tel: (678) 324-6272 Fax: (678) 401-5607
<b>II. OFFICIAL CORRESPONDENT</b>	Jordan Bauman  Vice President, Regulatory Affairs MiRus™, LLC 1755 West Oak Parkway Suite 100 Marietta, Georgia 30062 Tel: (678) 324-6272 Fax: (678) 401-5607
<b>III. DATE PREPARED</b>	September 7, 2023
<b>IV. DEVICE</b>	
<b>Name of Device</b>	MiRus 3DR Lateral Lumbar Interbody Fusion System with Integrated Plate Fixation
<b>Common Name</b>	Intervertebral Body Fusion Device
<b>Classification Name</b>	21 CFR 888.3080
<b>Regulatory Class</b>	Class II
<b>Product Codes</b>	OVD, MAX
<b>Submission Type</b>	Traditional 510(k)
<b>V. PREDICATE DEVICE</b>	<u>Primary Predicate</u> CASCADIA Interbody System – K2M (K172941) <u>Additional Predicate</u> MiRus 3D Printed Lumbar Interbody Fusion Systems – MiRus, LLC (K191906) <u>Reference Device</u> CYGNUS™ MoRe Anterior Cervical Plate System – MiRus, LLC (K220441)

**VI. DEVICE DESCRIPTION**

The MiRus 3DR Lateral Lumbar Interbody Fusion System with Integrated Plate Fixation is a spinal intervertebral body fusion device used to provide structural stability following lateral lumbar interbody fusion for lumbar degenerative disorders.

The MiRus 3DR LLIF is additively manufactured from Titanium-6 Aluminum-4 Vanadium ELI per ASTM F3001.

The MiRus MoRe Lateral Lumbar Plating System consists of a lateral plate manufactured from Molybdenum-47.5 Rhenium Alloy (MoRe) per ASTM F3273, bone screws, locking cam mechanism, and a connecting screw manufactured from Titanium-6 Aluminum-4 Vanadium ELI per ASTM F136.

Instrumentation is manufactured from Stainless Steel per ASTM F899. The system is offered in several configurations of various sizes to accommodate different patient anatomy and surgical approaches.

## **VII. INDICATIONS FOR USE**

The MiRus™ 3DR Lumbar Interbody Fusion System is indicated for intervertebral body fusion procedures in skeletally mature patients with degenerative disc disease (DDD) of the lumbar spine at one or two contiguous levels from L1-L2 to L5-S1. DDD is defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies. These DDD patients may also have up to Grade I spondylolisthesis or retrolisthesis at the involved level(s). Devices are to be used with autogenous or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft and supplemental fixation. Patients should have at least six (6) months of non-operative treatment prior to treatment with an intervertebral body fusion device.

When used with or without the MiRus MoRe Lateral Lumbar Plating System, the system is indicated for use with supplemental fixation cleared by the FDA for use in the lumbar spine.

## **VIII. PREDICATE DEVICE COMPARISON**

The MiRus 3DR Lateral Lumbar Interbody Fusion System with Integrated Plate Fixation has the same intended use, indications for use, labeling, and technological characteristics as the predicate systems, including the same design features, geometries, sizes, and materials.

The MiRus™ 3DR LLIF device has been previously cleared (additional predicate - K191906). The MiRus MoRe Lateral Lumbar Plating System device configuration consisting of MoRe alloy plates and titanium alloy locking cams and bone screws is similar to the CYGNUS™ MoRe Anterior Cervical Plate System (reference device - K220441) in terms of materials and design features.

## **IX. PERFORMANCE DATA**

The mechanical performance profile of MiRus 3DR Lateral Lumbar Interbody Fusion System with Integrated Plate Fixation was assessed through dynamic construct testing and screw dissociation testing in accordance with the following test methods:

- Dynamic Compression (ASTM F2077-18)
- Plate/Cage Dissociation Testing

## **X. CONCLUSIONS**

The MiRus 3DR Lateral Lumbar Interbody Fusion System with Integrated Plate Fixation has the same intended use, indications for use, labeling, and technological characteristics as the predicate systems, including the same design features, geometries, sizes, and materials. Performance data demonstrate that the MiRus 3DR Lateral Lumbar Interbody Fusion System with Integrated Plate Fixation is substantially equivalent to legally marketed predicate systems.