



May 6, 2022

BioFilm, Inc.
Kevin Jamil
Regulatory Affairs
3225 Executive Ridge
Vista, CA 92081

Re: K220653
Trade/Device Name: Astroglide® Glycerin & Paraben Free
Regulation Number: 21 CFR§ 884.5300
Regulation Name: Condom
Regulatory Class: II
Product Code: NUC
Dated: February 27, 2022
Received: March 7, 2022

Dear Kevin Jamil:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Monica D. Garcia, Ph.D.
Assistant Director
DHT3B: Division of Reproductive,
Gynecology and Urology Devices
OHT3: Office of GastroRenal, ObGyn,
General Hospital and Urology Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K220653

Device Name
Astroglide® Glycerin & Paraben Free

Indications for Use (Describe)

Astroglide® Glycerin & Paraben Free is a personal lubricant for penile, vaginal, and/or anal application intended to moisturize and lubricate, to enhance the ease and comfort of intimate sexual activity and supplement the body's natural lubrication. This product is compatible with natural rubber latex and polyisoprene condoms. This product is not compatible with polyurethane condoms.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) Summary – K220653
Astroglide® Glycerin & Paraben Free

I. General Information on Submitter

Applicant: BioFilm, Inc.
Address: 3225 Executive Ridge
Vista, CA 92081 USA
Telephone: 760-727-9030
Fax: 760-727-8080
Contact Person: Kevin Jamil
Contact Title: Regulatory Affairs

Email: kevin@biofilm.com
Date Prepared: May 6, 2022

II. General Information on Device

Proprietary Name: Astroglide® Glycerin & Paraben Free
Common Name: Personal Lubricant
Regulation Name: Condom
Regulation Number: 21 CFR 884.5300
Regulatory Class: II
Product Code: NUC (Lubricant, Personal)

III. Predicate Device

Predicate Device	510(k) Number
Astroglide® Glycerin & Paraben Free	K072647

This predicate device has not been subject to a design-related recall.

IV. Description of Device

Astroglide® Glycerin & Paraben Free is a non-sterile, water-based liquid personal lubricant for penile, anal, and/or vaginal application. This product is not a spermicide or contraceptive. It is compatible with natural rubber latex and polyisoprene condoms. It is not compatible with polyurethane condoms. The subject device is sold as an over-the-counter (OTC) product in a 2.5 oz. clear PETE bottle with a screw on polypropylene flip-top cap. The bottle is labeled front and back and then packaged in a cardboard carton which constitutes the final packaging.

This device is composed of water, butylene glycol, propylene glycol, xylitol, methyl gluceth-20, polyquarternium-7, hydroxyethyl cellulose, potassium sorbate, sodium benzoate, citric acid.

The specifications for Astroglide® Glycerin & Paraben Free are described in **Table 1**.

Table 1. Device Specifications

Parameter	Specification (Test Method)
Color	Clear to slight golden
Clarity	Clear
Odor	Slight sweet odor
Absence of particulate matter	Absent
Viscosity	1000-2800 cP
pH	3.5-5.5
Osmolality	230-300 mOsm/kg, 1:5 dilution factor
Total yeast/mold count (TYMC)	<10 cfu/mL (USP <61>)
Total aerobic microbial count (TAMC)	<100 cfu/mL (USP <61>)
Presence of Pathogenic Organisms (<i>Staphylococcus aureus</i> , <i>Pseudomonas aeruginosa</i> , and <i>Candida albicans</i>)	Absent (USP <62>)
Antimicrobial effectiveness	Meets USP <51> acceptance criteria for Category 2 products.

V. Indications for Use

Astroglide® Glycerin & Paraben Free is a personal lubricant for penile, vaginal, and/or anal application intended to moisturize and lubricate, to enhance the ease and comfort of intimate sexual activity and supplement the body's natural lubrication. This product is compatible with natural rubber latex and polyisoprene condoms. This product is not compatible with polyurethane condoms.

VI. Substantial Equivalence Discussion

The following table compares the intended use and key technological characteristics of the subject and predicate device:

Characteristic / Feature	Astroglide® Glycerin & Paraben Free (subject device)	Astroglide® Glycerin & Paraben Free Personal Lubricant (predicate device) – K072647	Comparison

Indication for use	Astroglide® Glycerin & Paraben Free is a personal lubricant for penile, vaginal, and/or anal application intended to moisturize and lubricate, to enhance the ease and comfort of intimate sexual activity and supplement the body's natural lubrication. This product is compatible with natural rubber latex and polyisoprene condoms. This product is not compatible with polyurethane condoms.	Glycerin & Paraben Free Astroglide® is a personal lubricant for penile, anal or vaginal application, intended to moisturize and lubricate, to enhance the ease and comfort of intimate sexual activity and supplement the body's natural lubrication. This product is compatible with latex condoms.	Similar: The subject and predicate devices have similar indications for use, with the exception of the condom compatibility statement, and they have the same intended use.
Water-Based Lubricant	Yes	Yes	Same
Over the Counter	Yes	Yes	Same
Odorless	No	No	same
Not a contraceptive or Spermicide	Yes	Yes	Same
Non-sterile	Yes	Yes	Same
Primary Ingredients	Purified Water, Xylitol, Butylene Glycol, Methyl Gluceth-20, Propylene Glycol, Polyquaternium-7, Hydroxyethylcellulose, Potassium Sorbate, Sodium Benzoate, Citric Acid	Not publicly available	Different: The ingredients of the predicate device are different; the ingredients do not raise different questions of Safety & Effectiveness (S & E)
Microbial Limits	Total mold/yeast count <10 cfu/mL Total aerobic microbial count <100 cfu/mL Absence of pathogens (<i>Candida albicans</i> , <i>Pseudomonas aeruginosa</i> , <i>Staphylococcus aureus</i>)	Not publicly available	N/A

Viscosity	1000-2800 cP	Not publicly available	N/A
Osmolality	230-300 mOsm/kg, 1:5 dilution factor	Not publicly available	N/A
pH	3.5-5.5	Not publicly available	N/A

The subject and predicate devices have similar indications for use and have the same intended use – to provide lubrication during intimate sexual activity. The subject and predicate devices have different technological characteristics, including different

formulations and device specifications. The different technological characteristics do not raise different types of safety and effectiveness questions.

VII. Summary of Non-Clinical Performance Testing

Biocompatibility

Biocompatibility testing on the subject lubricant was performed in accordance with the 2020 FDA guidance document *Use of International Standard ISO 10993-1, "Biological Evaluation of Medical Devices – Part 1: Evaluation and testing within a risk management process"* and ISO 10993-1:2009 as follows:

- Cytotoxicity (per ISO 10993-5:2009/(R)2014)
- Human Repeat Insult Patch Testing (HRIPT) per ASTM D6355 to address sensitization and irritation
- Systemic Toxicity (per ISO 10993-11:2017)

The results of this testing demonstrate that the subject lubricant is non-cytotoxic, non-irritating, non-sensitizing, and not systemically toxic.

Shelf Life

The subject device is a non-sterile personal lubricant packaged in a 2.5 oz. bottle with a 22-month shelf-life in accordance with the results of an accelerated aging study, conducted for 8 months at 40°C per ASTM F1980-16, *Standard Guide for Accelerated Aging of Sterile Barrier Systems for Medical Devices*. The device specifications listed in **Table 1** were tested across the device shelf-life and the subject device met the specifications at all time points.

Condom Compatibility

Astroglide® Glycerin & Paraben Free was tested for compatibility with natural rubber latex, polyisoprene, and polyurethane condoms using ASTM D7661-18, *Standard Test Method for Determining Compatibility of Personal Lubricants with Natural Rubber Latex Condoms*. The results show that Astroglide® Glycerin & Paraben Free is compatible with natural rubber latex and polyisoprene condoms. Astroglide® Glycerin & Paraben Free is not compatible with polyurethane condoms.

VIII. Conclusion

The results of the testing described above demonstrate that the Astroglide® Glycerin & Paraben Free is as safe and effective as the predicate device and supports a determination of substantial equivalence.