



February 16, 2023

MorLaser Shenzhen Co., Ltd.  
% Riley Chen  
Registration engineer  
Feiyang Drug & Medical Consulting Technical Service Group  
Rm 2401 Zhenye International Business Center, No. 3101-90  
Qianhai Road  
Shenzhen, Guangdong 518052  
China

Re: K223524

Trade/Device Name: IPL Hair Removal, Model(s): SB01, SB01A, SB01B, SB01C, SN02, SN03

Regulation Number: 21 CFR 878.4810

Regulation Name: Laser Surgical Instrument For Use In General And Plastic Surgery And In  
Dermatology

Regulatory Class: Class II

Product Code: OHT

Dated: November 21, 2022

Received: November 23, 2022

Dear Riley Chen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

**Shlomit  
Halachmi -S** Digitally signed by  
Shlomit Halachmi -S  
Date: 2023.02.16  
10:21:53 -05'00'

For Jianting Wang  
Acting Assistant Director  
DHT4A: Division of General Surgery Devices  
OHT4: Office of Surgical  
and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K223524

Device Name

IPL Hair Removal

Model(s): SB01, SB01A, SB01B, SB01C, SN02, SN03

Indications for Use (Describe)

The IPL Hair Removal is an over-the-counter device intended for removal of unwanted hair. The device is also indicated for the permanent reduction in hair regrowth, defined as the long-term, stable reduction in the number of hairs re-growing when measured at 6, 9 and 12 months after the completion of a treatment regime.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

# 510(k) Summary

"510(k) Summary" as required by 21 CFR Part 807.92.

## I. Submitter

MorLaser Shenzhen Co., Ltd.  
313, 314, 315 & 4th Floor, Building E, Tangtou No. 1 Industry District, Shiyan Town, Bao'an District, Shenzhen, China  
Post code: 518108  
Tel.: (+86) 0755-26738887

Janny Shao  
Project Manager  
Tel: (+86) 13632750965  
Email: jannyshao@ibe.group

## II. Device

Name of Device: IPL Hair Removal  
Model(s): SB01, SB01A, SB01B, SB01C, SN02, SN03  
Common Name: Light Based Over-The-Counter Hair Removal  
Regulation Name: Laser surgical instrument for use in general and plastic surgery and in dermatology  
Regulatory Class: II  
Product Code: OHT  
Regulation Number: 21 CFR 878.4810

## III. Predicate Device

Predicate devices:

<u>Manufacturer</u>	<u>Predicate Device</u>	<u>510(k) Number</u>	<u>Approval Date</u>
Glan Electronics Co., Ltd.	IPL Hair Removal, Model: OBT-02	K213041	November 18, 2021
Shenzhen Ulike Smart Electronics Co.,Ltd	IPL Hair Removal Device, Model(s): UI04A, UI04B, UI04C	K221002	June 01, 2022

Reference device:

<u>Manufacturer</u>	<u>Reference Device</u>	<u>510(k) Number</u>	<u>Approval Date</u>
Dongguan Define Beauty Electronic Technology Co. Ltd	IPL Hair Removal SG-8025	K212318	January 14, 2022

## IV. Device Description

IPL Hair Removal (Models: SB01, SB01A, SB01B, SB01C, SN02, SN03), is an over-the-counter, single-person-use device for the permanent reduction of hair growth based on Intense Pulsed Light (IPL). It works below the skin's surface and does not involve any cutting or pulling, reducing hair growth with minimal pain.

Model SB01, SB01A, SB01B and SB01C are powered by lithium battery and can be charged by power adapter, while model SN02 and SN03 are only powered by power adapter. Their IPL emission activation is by finger switch. All models of IPL Hair Removal adopt irreplaceable flash window and are suitable for multiple hair removal areas, including small areas (e.g. underarms, bikini lines) and large areas (e.g. arms, legs). IPL Hair Removal contains a Xenon Arc Flashlamp and a skin sensor to detect appropriate skin contact, if the device is not in full contact with the skin, the device cannot emit the treatment light pulses.

## V. Indications for Use

The IPL Hair Removal is an over-the-counter device intended for removal of unwanted hair. The device is also indicated for the permanent reduction in hair regrowth, defined as the long-term, stable reduction in the number of hairs re-growing when measured at 6, 9 and 12 months after the completion of a treatment regime.

## VI. Materials

Component name	Material of Component	Body Contact Category	Contact Duration
IPL Hair Removal (Enclosure and Flash window)	ABS+PC, Sapphire	Surface-contacting device: Intact skin	Less than 24 hours

We have tested the device for biocompatibility by a reliable third-party lab. For details, please refer to “Biocompatibility Discussion”.

## VII. Comparison of Technological Characteristics With the Predicate Device

The IPL Hair Removal has the same intended use, mode of action and similar operational characteristics as the predicate devices and reference device. Any minor differences between the subject device and the listed predicate devices and reference device do not raise any issues of safety or efficacy. Performance data supports that the device is safe and as effective as the predicate devices and reference device for its intended use. Therefore, the IPL Hair Removal may be found substantially equivalent to its predicate devices and reference device.

IPL Hair Removal is compared with the following Predicate Devices and Reference Device in terms of intended use, design, material, specifications, and performance:

<u>Comparison Elements</u>	<u>Subject Device</u>	<u>Primary Predicate Device</u>	<u>Predicate Device</u>	<u>Reference Device</u>	<u>Remark</u>
510(k)	K223524	K213041	K221002	K212318	/

<u>Comparison Elements</u>	<u>Subject Device</u>	<u>Primary Predicate Device</u>	<u>Predicate Device</u>	<u>Reference Device</u>	<u>Remark</u>
Number					
Trade name	IPL Hair Removal (Models: SB01, SB01A, SB01B, SB01C, SN02, SN03)	IPL Hair Removal, Model: OBT-02	IPL Hair Removal Device (UI04A, UI04B, UI04C)	IPL Hair Removal SG-8025	/
Manufacturer	MorLaser Shenzhen Co., Ltd.	Glan Electronics Co., Ltd.	Shenzhen Ulike Smart Electronics Co.,Ltd	Dongguan Define Beauty Electronic Technology Co. Ltd	/
Regulation number	21 CFR 878.4810	21 CFR 878.4810	21 CFR 878.4810	21 CFR 878.4810	Same
Product code	OHT	OHT	OHT	OHT	Same
Device classification	Class II	Class II	Class II	Class II	Same
Indication for use/ Intended use	The IPL Hair Removal is an over-the-counter device intended for removal of unwanted hair. The device is also indicated for the permanent reduction in hair regrowth, defined as the long-term, stable reduction in the number of hairs re-growing when measured at 6, 9 and 12 months after the completion of a treatment regime.	The IPL Hair Removal Device OBT-02 Version is indicated for the removal of unwanted hair. The device is also indicated for the permanent reduction in hair regrowth, defined as the long-term, stable reduction in the number of hairs re-growing when measured at 6, 9 and 12 months after the completion of a treatment regime. The device is used for adults.	IPL Hair Removal Device is indicated for the removal of unwanted hair. The device is also indicated for the permanent reduction in hair regrowth, defined as the long term, stable reduction in the number of hairs regrowing when measured at 6, 9 and 12 months after the completion of a treatment regime.	The IPL Hair Removal (Model: SG-8025) is an over the Counter device intended for the removal of unwanted body and/or facial hair in adults. It is also intended for permanent reduction in unwanted hair. Permanent hair reduction is defined as the long-term stable reduction in the number of hairs re-growing when measured at 6. 9. and 12 months after the completion of a treatment regimen.	Same
Prescription or OTC	OTC	OTC	OTC	OTC	Same
<b>Device design</b>					
Source energy	Supplied by lithium battery (applicable for SB01, SB01A, SB01B, SB01C) Supplied by external adapter (applicable for SN02, SN03)	Supplied by external adapter	Supplied by external adapter	Supplied by external adapter	<u>Different</u>
Power supply	AC100~240V,	100-240 V AC	100-240V~, 50/60Hz	Unknown	<u>Different</u>

<u>Comparison Elements</u>	<u>Subject Device</u>	<u>Primary Predicate Device</u>	<u>Predicate Device</u>	<u>Reference Device</u>	<u>Remark</u>
	50/60Hz Li battery: 3.6V 5000mAh				
Dimension	118.7*187.7*42.8mm (applicable for SB01, SB01A, SB01B, SB01C) 125*165*78mm (applicable for SN02, SN03)	150*75*45mm (H*W*D)	60.5(W)x38(H)x169. 7(L)mm	205*76*56mm (H*W*D)	<u>Different</u>
Sterilization	Not required	Not required	Not required	Not required	Same
<b>Output specification</b>					
Light source	Intense Pulsed Light	Intense Pulsed Light	Intense Pulsed Light	Intense Pulsed Light	Same
Energy medium	Xenon Arc Flashlamp	Xenon Arc Flashlamp	Xenon Arc Flashlamp	Xenon lamp	Same
Wavelength range	510-1100nm	510-1100nm	550-1200nm	530nm	Same
Energy density	SB01, SB01C: 1.8- 3.5J/cm <sup>2</sup> SB01A, SB01B: 1.8- 3.1J/cm <sup>2</sup> SN02, SN03: 3.5- 5J/cm <sup>2</sup>	1.5-4.0J/cm <sup>2</sup>	3.03-5.3J/cm <sup>2</sup>	2.5~4J/cm <sup>2</sup>	Similar
Output energy	<b>SB01, SB01C:</b> 5.4 - 10.5J <b>SB01A, SB01B:</b> 5.4 - 9.5J <b>SN02, SN03:</b> 10.5 - 15J	Level 1: 4.5J Level 2: 5.7J Level 3: 7.8J Level 4: 11.1J Level 5: 12J	10-17.5J	Level 1: 7.5J Level 2: 8.5J Level 3: 9.5J Level 4: 11J Level 5: 12J	Similar
Spot size	3.0cm <sup>2</sup>	3.0cm <sup>2</sup>	3.3cm <sup>2</sup>	3.0cm <sup>2</sup>	Same
Pulse duration	1ms	3ms	7-10ms	1ms	Same
Pulsing control	Finger switch	Finger switch	Finger switch	Finger switch	Same
Delivery device	Direct illumination to tissue	Direct illumination to tissue	Direct illumination to tissue	Direct illumination to tissue	Same
Output intensity level	SB01, SB01C, SN02, SN03: 6 Levels SB01A, SB01B: 5 Levels	5 Levels	5 Levels	5 Levels	Similar
Software/ Firmware/ Microprocess or Control?	Yes	Yes	Yes	Yes	Same

<u>Comparison Elements</u>	<u>Subject Device</u>	<u>Primary Predicate Device</u>	<u>Predicate Device</u>	<u>Reference Device</u>	<u>Remark</u>
Additional features					
Electrical safety	IEC 60601-1 IEC 60601-1-2 IEC 60601-1-11 IEC 60601-2-83	IEC 60601-1 IEC 60601-1-2 IEC 60601-1-11 IEC 60601-2-57	ANSI AAMI ES60601-1 IEC 60601-1-2 IEC 60601-1-11 IEC 60601-2-57 IEC 60601-2-83	IEC 60601-1 IEC 60601-1-2 IEC 60601-1-11 IEC 60601-2-57	Similar
Eye safety	IEC 62471	Unknown	IEC 62471	IEC 62471	Same
Biocompatibility	ISO 10993-5 ISO 10993-10	ISO 10993-5 ISO 10993-10	ISO 10993-5 ISO 10993-10	ISO 10993-5 ISO 10993-10	Same

### VIII. Performance Data

The following performance data were provided in support of the substantial equivalence determination.

#### **1) Biocompatibility Testing**

The biocompatibility evaluation for the body-contacting components of the IPL Hair Removal was conducted in accordance with the “Use of International Standard ISO 10993-1, 'Biological Evaluation of Medical Devices –Part 1: Evaluation and Testing Within a Risk Management Process, Document Issued on September 4, 2020”, as recognized by FDA. The following testing was performed to, and passed, including:

- ISO 10993-5:2009, Biological Evaluation of Medical Devices –Part 5: Tests for In Vitro Cytotoxicity
- ISO 10993-10:2010, Biological Evaluation of Medical Devices –Part 10: Tests for Irritation and Skin Sensitization

#### **2) Electrical Safety and EMC**

Electrical safety and EMC testing was performed to, and passed, as per the following standards:

- ANSI AAMI ES60601-1 Medical electrical equipment –Part 1: General requirements for basic safety and essential performance
- IEC 60601-1-2 Medical electrical equipment –Part 1-2: General requirements for basic safety and essential performance –Collateral standard: electromagnetic compatibility
- IEC 60601-1-11 Medical Electrical Equipment –Part 1: General Requirements for Basic Safety and Essential Performance –Collateral Standard: Requirements for Medical Electrical Equipment and Medical Electrical Systems Used in the Home Healthcare Environment
- IEC 60601-2-83 Medical electrical equipment - Part 2-83: Particular requirements for the basic safety and essential performance of home light therapy equipment

#### **3) Eye Safety**

- IEC 62471 Photobiological safety of lamps and lamp systems



#### **4) Software Verification and Validation**

Software documentation consistent with *moderate level* of concern was submitted in this 510(k). System validation testing presented in this 510(k) demonstrated that all software requirement specifications are met and all software hazards have been mitigated to acceptable risk levels.

#### **IX. Conclusions**

Based on the above analysis and non-clinical tests performed, it can be concluded that the subject device IPL Hair Removal is as safe, as effective, and performs as well as the legally marketed predicate device and reference devices.