



January 19, 2022

CoreLink, LLC
% Nathan Wright
Engineer & Regulatory Specialist
Empirical Testing Corp.
4628 Northpark Drive
Colorado Springs, Colorado 80918

Re: K214059
Trade/Device Name: CoreLink Navigation Instruments
Regulation Number: 21 CFR 882.4560
Regulation Name: Stereotaxic Instrument
Regulatory Class: Class II
Product Code: OLO
Dated: December 23, 2021
Received: December 27, 2021

Dear Nathan Wright:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, M.P.H.
Assistant Director
DHT6C: Division of Restorative, Repair
and Trauma Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K214059

Device Name
CoreLink Navigation Instruments

Indications for Use (Describe)

CoreLink Navigation Instruments are intended to be used in the preparation and placement of Tiger OCT screw implants during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. These instruments are designed for use with the Medtronic StealthStation® System S8 (V1.2.20), which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as vertebra, can be identified relative to a CT or MR based model, fluoroscopy images, or digitized landmarks of the anatomy.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

K214059 510(K) SUMMARY

| | |
|----------------------------|---|
| Submitter's Name: | CoreLink, LLC |
| Submitter's Address: | 2072 Fenton Logistics Park St. Louis, Missouri 63026 |
| Submitter's Telephone: | 888-349-7808 |
| Contact Person: | Nathan Wright MS Empirical Testing Corp. 719-351-0248 nwright@empiricaltech.com |
| Date Summary was Prepared: | December 23, 2021 |
| Trade or Proprietary Name: | CoreLink Navigation Instruments |
| Common or Usual Name: | Orthopedic Stereotaxic Instruments |
| Classification: | Class II per 21 CFR §882.4560 |
| Product Code: | OLO |
| Classification Panel: | Orthopedic – Spinal Devices (DHT6B) |



EMPIRICAL TESTING CORP.

DESCRIPTION OF THE DEVICE SUBJECT TO PREMARKET NOTIFICATION:

CoreLink Navigation Instruments are non-sterile, re-usable manual surgical instruments made from stainless steel. These instruments are designed to interface with the Medtronic StealthStation® Navigation System to assist surgeons in precisely locating anatomical structures.

The purpose of this submission is to add instruments to the previously cleared CoreLink Navigation Instruments set to offer compatibility with CoreLink OCT pedicle screws.

INDICATIONS FOR USE

CoreLink Navigation Instruments are intended to be used in the preparation and placement of Tiger OCT Screw implants during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. These instruments are designed for use with the Medtronic StealthStation® System S8 (V1.2.0), which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as vertebra, can be identified relative to a CT or MR based model, fluoroscopy images, or digitized landmarks of the anatomy.

TECHNOLOGICAL CHARACTERISTICS

The subject and predicate devices have nearly identical technological characteristics and the minor differences do not raise any new issues of safety and effectiveness. Specifically, the following characteristics are identical between the subject and predicates:

- Indications for use
- Materials of manufacture
- Principles of operation

Table 5-1 Predicate Devices

| 510k Number | Trade or Proprietary or Model Name | Manufacturer | Predicate Type |
|--------------------|--|-----------------------------------|-----------------------|
| K212825 | CoreLink Navigation Instruments | CoreLink, LLC | Primary |
| K181111 | Navigated INFINITY™ Instruments | Medtronic Sofamor Danek USA, Inc. | Additional |
| K200863 | Tiger Occipital-Cervical-Thoracic Spinal Fixation System | CoreLink, LLC | Reference |

PERFORMANCE DATA

In support of this Special 510(k) submission, CoreLink, LLC has conducted confirmatory positional accuracy testing per ASTM F2554 with Medtronic StealthStation® System S8 Version 1.2.0 (1.2.0-20) as well as engineering analysis including tolerance stack analysis to demonstrate that the system modifications provide adequate and substantially equivalent accuracy for their intended use.

CONCLUSION

The overall technology characteristics and performance data lead to the conclusion that the CoreLink Navigation Instruments are substantially equivalent to the predicate device.