



July 28, 2023

Shanghai Sanyou Medical Co, LTD
% Christine Scifert
Partner
MRC Global, LLC
9085 E. Mineral Cir., Suite 110
Centennial, Colorado 80112

Re: K231931

Trade/Device Name: Adena-Zina Spinal System
Regulation Number: 21 CFR 888.3070
Regulation Name: Thoracolumbosacral Pedicle Screw System
Regulatory Class: Class II
Product Code: NKB
Dated: June 30, 2023
Received: June 30, 2023

Dear Christine Scifert:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin
O'Neill -S 

Colin O'Neill, M.B.E.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K231931

Device Name

Adena-Zina Spinal System

Indications for Use (Describe)

The Adena-Zina Spinal System is intended for posterior, non - cervical fixation as an adjunct to fusion for skeletally mature patients for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion. In addition, when used as a pedicle screw fixation system, the Adena-Zina Spinal System is intended for skeletally mature patients with severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral, L5-S1 vertebra, who are receiving fusion by autogenous bone graft only, who are having the device attached to the lumbar and sacral spine (levels may be from L3 to the sacrum/ilium), who are having the device removed after the attainment of a solid fusion.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary
Adena-Zina Spinal System
30 June 2023

Company: Manufacturing Facility and Headquarters:
Shanghai Sanyou Medical Co, LTD
385 Huirong Road, Jiading District,
Shanghai 201815
Phone: 021-58389980

Company Contact: David Fan
VP Marketing
david.fan@sanyou-medical.com
Phone: +86 21 58389980
Fax: +86 21 59990826

Official Correspondent: Christine Scifert - MRC Global, LLC
Christine.Scifert@askmrcglobal.com
901-831-8053

Secondary Correspondent: Jen McBride - MRC Global, LLC
Jen.McBride@askmrcglobal.com
901-831-8053

Trade Name: Adena-Zina Spinal System

Common Name: Thoracolumbosacral pedicle screw system

Classification: Class II

Regulation Number: 21 CFR 888.3070 (Thoracolumbosacral pedicle screw system)

Panel: Orthopedic

Product Code: NKB

Device Description:

The Adena-Zina Spinal System consists of a variety of shapes and sizes of screws and Ø5.5mm and Ø6.35mm rods that can be rigidly locked in a variety of configurations, with each construct being tailor-made for the individual case. Fixation is provided via a posterior approach. The components are made from titanium alloy or cobalt chrome alloy. Components of the system include the originally cleared (K152781) straight and pre-bent rods, reduction screws, fixed angle screws, fixed angle reduction screws, multi-axial screws, multi-axial reduction screws, T-links, domino connectors, sacro-iliac connectors, as well as the Duetto dual headed screws and connectors, along with associated set screws (K212066).

The subject submission seeks to add ZINA II screws to the Adena-Zina Spinal System, along with associated class I instruments.

Indications for Use:

The Adena-Zina Spinal System is intended for posterior, non - cervical fixation as an adjunct to fusion for skeletally mature patients for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis and/or lordosis); tumor; pseudarthrosis; and/or failed previous fusion. In addition, when used as a pedicle screw fixation system, the Adena-Zina Spinal System is intended for skeletally mature patients with severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral, L5-S1 vertebra, who are receiving fusion by autogenous bone graft only, who are having the device attached to the lumbar and sacral spine (levels may be from L3 to the sacrum/ilium), who are having the device removed after the attainment of a solid fusion.

Substantial Equivalence:

The subject Adena-Zina System is substantially equivalent to the following predicate devices:

Primary Predicate:

- Shanghai Sanyou Adena-Zina System – K152781

Secondary Predicates:

- Medtronic CD Horizon Fenestrated Screw Set– K152604; K193011

There are insignificant differences between the subject ZINA II screws and the predicate devices. The Indications for Use and Materials for predicate devices are all inclusive of the subject device, with the subject device materials and indications being identical to the primary predicate Adena-Zina System (K152781). The device design is similar to the secondary predicate Medtronic CD Horizon Fenestrated Screw Set (K152604). Performance bench testing has been performed and shows that the subject ZINA II screws perform equivalent to or better than the primary predicate device. Thus, it can be concluded that the subject does not raise new questions about safety and effectiveness.

Performance Testing:

Bench performance testing was performed on the subject ZINA II screws including static compression bending, dynamic compression bending, and static torsion per ASTM F1717 *Standard Test Methods for Spinal Implant Constructs in a Vertebroctomy Model*.

Conclusion:

Based on the test results and the comparison to the predicate devices, the subject device is determined to be substantially equivalent to the predicate devices.