



March 19, 2021

Spineart  
Franck Pennesi  
Chief Technical Officer  
3 chemin du pré Fleuri  
Plan Les Ouates, Geneva 1228  
Switzerland

Re: K210472

Trade/Device Name: SPINEART Navigation Instrument System  
Regulation Number: 21 CFR 882.4560  
Regulation Name: Stereotaxic Instrument  
Regulatory Class: Class II  
Product Code: OLO  
Dated: February 11, 2021  
Received: February 17, 2021

Dear Franck Pennesi:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For; Shumaya Ali, MPH  
Assistant Director  
DHT6C: Division of Restorative, Repair  
and Trauma Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

**Indications for Use**

510(k) Number (if known)

K210472

Device Name

SPINEART® Navigation Instrument System

Indications for Use (Describe)

The SPINEART® Navigation Instrument System reusable instruments are intended to be used during the preparation and placement of Spineart screws during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. The SPINEART® Navigation Instrument System reusable instruments are specifically designed for use with the Medtronic StealthStation System, which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a skull, a long bone, or vertebra, can be identified relative to a CT or MR based model, fluoroscopy images, or digitized landmarks for the anatomy.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRAStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**Special 510k - Device Modification**  
**SPINEART NAVIGATION Instrument System**



**510(k) SUMMARY**

|                                      |  |
|--------------------------------------|--|
| 510k                                 | Special 510(k): Device Modification  |
| Basis for submission                 | Product Line Extension of the SPINEART Navigation Instrument System (K183630)  |
| Submitted by                         | SPINEART<br>3 Chemin du Pré Fleuri<br>1228 PLAN LES OUATES<br>GENEVA SWITZERLAND   |
| Contacts                             | Franck PENNESI Chief Technical Officer<br>Phone: +41 22 570 1200 Fax: +41 22 594 8306<br>Mail: <a href="mailto:fpennesi@spineart.com">fpennesi@spineart.com</a><br>Regulatory contact: Dr Isabelle DRUBAIX (PhD) <a href="mailto:idee-consulting@bbox.fr">idee-consulting@bbox.fr</a>  |
| Date Prepared                        | February 8, 2021   |
| Common Name                          | Orthopedic Stereotaxic Instrument  |
| Trade Name                           | SPINEART Navigation Instrument System  |
| Classification Name                  | Orthopedic Stereotaxic Instrument  |
| Class                                | II   |
| Product Code                         | OLO  |
| CFR section                          | 882.4560   |
| Device panel                         | ORTHOPEDIC   |
| Legally marketed predicate device(s) | Primary predicate: SPINEART Navigation Instrument System manufactured by SPINEART (K183630)  |
| Indications for use                  | The SPINEART® Navigation Instrument System reusable instruments are intended to be used during the preparation and placement of Spineart screws during spinal surgery to assist the surgeon in precisely locating anatomical structures in either open or minimally invasive procedures. The SPINEART® Navigation Instrument System reusable instruments are specifically designed for use with the Medtronic StealthStation System, which is indicated for any medical condition in which the use of stereotactic surgery may be appropriate, and where reference to a rigid anatomical structure, such as a skull, a long bone, or vertebra, can be identified relative to a CT or MR based model, fluoroscopy images, or digitized landmarks for the anatomy. |

|   |  |
|---|--|
| Description of the device                                       | <p>The SPINEART® Navigation Instrument System reusable instruments are surgical instruments for use with the Medtronic StealthStation® Navigation System to assist surgeons in precisely locating anatomical structures in either open, minimally invasive, or percutaneous procedures for preparation and placement of pedicle screw system implants.</p> <p>The SPINEART® Navigation Instrument System includes the following: Screwdrivers, Taps, Drills, and Drill Guides.</p> <p>The SPINEART® Navigation Instrument System are to be used with the following Spineart Systems:</p> <ul style="list-style-type: none"> <li>- Romeo® 2</li> <li>- Romeo® 2 MIS</li> <li>- Perla® Cervico-thoracic Fixation System</li> </ul> <p>All instruments are made of stainless steel per ASTM F899. All instruments are reusable instruments provided not sterile.</p> <p>The SPINEART® Navigation Instrument System instruments are not compatible with implants from other manufacturers.</p> <p>The SPINEART® Navigation Instrument System are designed for use only with Medtronic StealthStation® System (V2.1.0) and the Medtronic NavLock® Tracker System.</p> |
| Technological characteristics compared to the predicate devices | <p>The subject product line extension of the SPINEART Navigation Instrument System (K183630) consists of addition of taps, drill and screwdrivers for the Perla TL and Perla TL MIS Posterior Osteosynthesis Systems.</p> <p>As was established in this submission, the SPINEART Navigation added instruments are substantially equivalent and have the same technological characteristics to predicate devices in areas including indications for use, function, material composition, design, range of sizes and accuracy performance. Verification and validation activities conducted on the added navigated instruments demonstrate that these navigated instruments are suitable to be used with the Medtronic® StealthStation® System (V2.1.0) and the Medtronic® NavLock Tracker System when implanting screws part of the Perla TL and Perla TL MIS Posterior Osteosynthesis Systems.</p>   |
| Discussion of Testing   | <p>Addition of navigated instruments to SPINEART Navigation Instrument System (K183630) does not require testing.</p>  |
| Conclusion  | <p>Based on the design features, technological characteristics, feature comparisons, indications for use, and dimensional and tolerance stack-up analysis, the added SPINEART Navigation Instruments have demonstrated substantial equivalence to the identified predicate devices.</p>  |