



December 23, 2021

Shandong Langtai International Trade Co., Ltd.  
% Boyle Wang  
Official Correspondent  
Shanghai Truthful Information Technology Co., Ltd.  
RM.1801, No.161, East Lujiazui Rd., Pudong  
Shanghai, Shanghai 200120  
China

Re: K212436

Trade/Device Name: NBR Synthetic Examination Gloves  
Regulation Number: 21 CFR 880.6250  
Regulation Name: Non-Powdered Patient Examination Glove  
Regulatory Class: Class I, reserved  
Product Code: LYZ,  
Dated: October 18, 2021  
Received: October 29, 2021

Dear Boyle Wang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Clarence W. Murray, III, PhD  
Assistant Director  
DHT4B: Division of Infection Control  
and Plastic Surgery Devices  
OHT4: Office of Surgical  
and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K212436

Device Name  
NBR Synthetic Examination Gloves

### Indications for Use (Describe)

A patient examination glove is a disposable device intended for medical purposes that is worn upon the examiner's hands to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

# 510(k) Summary

## (K212436)

This summary of 510(k) is being submitted in accordance with 21 CFR 807.92.

### **1.0 Submitter's Information**

Name: SHANDONG LANGTAI INTERNATIONAL TRADE CO.,LTD.  
Address: 8/F,office Building,Weiqiao Aluminum Deep,Processing Industrial  
Park ,ChanshanTown,Zouping,Binzhou City, Shandong, China  
Tel: +86-15550323002  
Contact: Ping Wang

### **Designated Submission Correspondent**

Name: Shanghai Truthful Information Technology Co., Ltd.  
Address: Room 1801, No. 161 East Lujiazui Rd., Pudong,Shanghai 200120 ,China  
Tel: +86-21-50313932  
Contact: Mr. Boyle Wang  
Email: [Info@truthful.com.cn](mailto:Info@truthful.com.cn)

Date of Preparation: Dec.21<sup>st</sup>,2021

### **2.0 Device Information**

Trade name: NBR Synthetic Examination Gloves  
Common name: Patient Examination Gloves  
Classification name: Non-powdered patient examination glove  
Model(s): S, M, L, XL  
Product code: LYZ  
Regulation number: 21CFR880.6250  
Classification: Class I  
Panel: General Hospital

### **3.0 Predicate Device Information**

Manufacturer: Zibo Huiying Medical Products, Co. Ltd.  
Device: Synmax Synthetic Patient Examination Vinyl Gloves,Powder Free,Blue  
510(k) number: K153028

### **4.0 Indication for Use**

A patient examination glove is a disposable device intended for medical purposes that is worn upon the examiner's hands to prevent contamination between patient and examiner.

## 5.0 Device Description

The subject device is powder free vinyl examination gloves, adding 1%-3% nitrile to improve the tensile strength and ultimate elongation. The subject device is blue color. It can be available in four specifications: S,M,L and XL.

The subject device is non-sterile.

## 6.0 Technological Characteristic Comparison Table

**Table1-General Comparison**

<b>Item</b>	<b>Subject Device (K212436)</b>	<b>Predicate Device (K153028)</b>
Product Code	LYZ	LYZ
Regulation No.	21CFR880.6250	21CFR880.6250
Class	I	I
Intended Use	A patient examination glove is a disposable device intended for medical purposes that is worn upon the examiner's hands to prevent contamination between patient and examiner.	A patient examination glove is a disposable device intended for medical purposes that is worn upon the examiner's hands to prevent contamination between patient and examiner.
Material	Poly Vinyl Chloride Polyurethane Di-(2-ethylhexyl) Terephthalate(DOTP)	Poly Vinyl Chloride Polyurethane Diisononyl Phthalate (DINP)
Powdered or Powered free	Powdered free	Powdered free
Design Feature	Ambidextrous	Ambidextrous
Colorant	Blue	Blue
Labeling Information	Single-use indication, powder free, device color, device name, glove size and quantity, Non-Sterile	Single-use indication, powder free, device color, device name, glove size and quantity, Non-Sterile
Dimensions(mm)	Length: ≥230; Width:	Length: Average over 234 on M size; Width:

		S: 85±5; M: 95±5; L: 105±5; XL: 115±5		Average over 96 on M size	
Thickness(mm)		Finger: ≥0.08; Palm: ≥0.08		Finger: Average 0.98; Palm: Average 0.096	
Physical Properties	Before Aging	Tensile Strength	11MPa, min	Tensile Strength	Average 16.9MPa
		Ultimate Elongation	300% min	Ultimate Elongation	Average 550%
	After Aging	Tensile Strength	11MPa, min	Tensile Strength	Average 14.4MPa
		Ultimate Elongation	300%min	Ultimate Elongation	Average 500%
Freedom from Holes		Be free from holes when tested in accordance with ASTM D5151 AQL=2.5		Be free from holes when tested in accordance with ASTM D5151 AQL=2.5	
Powder Content		Meet the requirements of ASTM D6124		Meet the requirements of ASTM D6124	
Biocompatibility		ISO 10993-10; Under the conditions of the study, not an irritant or a sensitizer		ISO 10993-10; Under the conditions of the study, not an irritant or a sensitizer	
		ISO 10993-11; Under the condition of acute systemic toxicity test, the test article did not show acute systemic toxicity in vivo.		/	
		ISO 10993-5; Under conditions of the study, device extract is cytotoxic		/	

## 7.0 Summary of Non-clinical Testing

Non-clinical tests were conducted to verify that the proposed device met all design specifications. The test results demonstrated that the proposed device complies with the following standards:

ISO 10993-10:2010 Biological Evaluation of Medical Devices - Part 10: Tests For Irritation And Skin Sensitization.

ISO 10993-11:2017, Biological evaluation of medical devices - Part 11: Tests

for systemic toxicity.

ASTM D6124-06 (Reapproved 2017), Standard Test Method for Residual Powder on Medical Gloves

ASTM D5151-19, Standard Test Method for Detection of Holes in Medical Gloves.

ASTM D5250-19, Standard Specification for Poly (vinyl chloride) Gloves for Medical Application.

**Table 2 - Summary of non-clinical performance testing**

Test Method	Purpose	Acceptance Criteria			Results
ASTM D5250	Physical Dimensions Test	Length(mm): ≥230; Width(mm): S: 85±5; M: 95±5; L: 105±5; XL: 115±5			Length: > 240/Pass; Width: S: 86-89 /Pass M: 95-97/ Pass L: 106-108/ Pass XL:115-117/ Pass
		Thickness (mm): Finger: ≥0.08 Palm: ≥0.08			Finger: 0.08-0.16/Pass Palm: 0.08-0.13/Pass
ASTM D5151	Watertightness Test for Detection of Holes	Meet the requirements of ASTM D5151 AQL 2.5			0/125/Pass
ASTM D6124	Powder Content	Meet the requirements of ASTM D6124 < 2.0mg			0.12mg/Pass;
ASTM D412	Physical properties	Before Aging	Tensile Strength	≥11MPa	11 -18/Pass;
			Ultimate Elongation	≥300%	340-518/Pass;
		After Aging	Tensile Strength	≥11MPa	11-16/Pass;
			Ultimate Elongation	≥300%	315-438/Pass;
ISO 10993-11	Acute Systemic Toxicity	Non- acute systemic toxicity			Under conditions of the study, did not show acute systemic toxicity in vivo / Pass

ISO 10993-10	Irritation	Non-irritating	Under the conditions of the study, not an irritant/ Pass
ISO 10993-10	Sensitization	Non-sensitizing	Under conditions of the study, not a sensitizer./ Pass

## **8.0 Summary of Clinical Testing**

Clinical testing is not needed for this device.

## **9.0 Conclusion**

The conclusions drawn from the nonclinical tests demonstrate that the subject device NBR Synthetic Examination Gloves, is as safe, as effective, and performs as well as or better than the legally marketed predicated device.