



Nov 5, 2021

Guangzhou ZhengKang Medical Equipment Co.,Ltd  
Tracy Che  
Registration engineer  
Feiyong Drug & Medical Consulting Technical Service Group  
Rm 2401 Zhenye International Business Center, No. 3101-90  
Qianhai Road  
Shenzhen, Guangdong 518052  
China

Re: K212460

Trade/Device Name: Medical Surgical Mask (Model: ZKM-U01(non-sterilized type 17.5cm×9.5cm))  
Regulation Number: 21 CFR 878.4040  
Regulation Name: Surgical apparel  
Regulatory Class: Class II  
Product Code: FXX  
Dated: July 27, 2021  
Received: August 5, 2021

Dear Tracy Che:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For Clarence W. Murray, III, PhD  
Assistant Director  
DHT4B: Division of Infection Control  
and Plastic Surgery Devices  
OHT4: Office of Surgical  
and Infection Control Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)  
K212460

Device Name  
Medical Surgical Mask (Model: ZKM-U01(non-sterilized type 17.5cm×9.5cm))

Indications for Use (Describe)

Medical Surgical Mask is intended to be worn to protect both patients and healthcare workers against transfer of microorganisms, body fluids and particulate materials. This is a single use device, provided non-sterile.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

---

---

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

---

---

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510 (k) Summary

This “510(k) Summary” of 510(k) safety and effectiveness information is submitted in accordance with requirements of Title 21, CFR Section 807.92.

### (1) Applicant information

510 (k) owner’s name: Guangzhou ZhengKang Medical Equipment Co.,Ltd  
Address: Room101,No.9Nanling Zhongxin Kuang Road,Taihe Town,Baiyun District,GuangZhou,China  
Contact person: Mike Wu  
Phone number: 86-13600073056  
Fax number: /  
Email: info@zkmedical.com  
Date of summary prepared: 2021.08.02

### (2) Reason for the submission

New device, there were no prior submissions for the device.

### (3) Proprietary name of the device

Trade name: Medical Surgical Mask  
Regulation Name: Surgical apparel  
Regulation number: 21 CFR 878.4040  
Product code: FXX  
Review panel: General & Plastic Surgery  
Regulation class: Class II

### (4) Predicate device

<b>Sponsor</b>	Mexpo International Inc.
<b>Device Name</b>	Avianz® Surgical Face Mask
<b>510(k) Number</b>	K200847
<b>Product Code</b>	FXX
<b>Regulation Number</b>	21 CFR 878.4040
<b>Regulation Class</b>	II

### (5) Description/ Design of device

Medical Surgical Mask is a non-sterile, single use, three-layer mask with earloops and nose piece. The Medical Surgical Mask is manufactured with three layers, the outer and inner layers are made

of non-woven fabric, and the middle layer is made of melt-blown fabric. The earloops are made of polyester textured yarn. The nose piece is made of polyethylene, which allow the users to adjust the nose piece according to the shape of the bridge of the nose.

### (6) Indications for use

Medical Surgical Mask is intended to be worn to protect both patients and healthcare workers against transfer of microorganisms, body fluids and particulate materials. This is a single use device, provided non-sterile.

### (7) Materials

Component of Device Requiring Biocompatibility	Material of Component	Body Contact Category (ISO 10993-1)	Contact Duration (ISO 10993-1)
Medical Surgical Mask ( ZKM-U01 (non-sterilized type 17.5cm×9.5cm))	Non-woven fabric, Melt-blown fabric, Polyester textured yarn, Polyethylene	Surface-contacting device: skin	> 24h to 30 d

The body-contacting material used in the Medical Surgical Mask have all passed biocompatibility test. Details can be seen in “Biocompatibility Discussion”.

### (8) Comparison to Predicate Device

Item	Proposed device	Predicate device	Remark
Trade name	Medical Surgical Mask	Avianz® Surgical Face Mask	/
510 (k) number	K212460	K200847	/
Regulation number	21 CFR 878.4040	21 CFR 878.4040	Same
Regulation description	Surgical apparel	Surgical apparel	Same
Product code	FXX	FXX	Same
Class	II	II	Same
Indications for use/ Intended use	Medical Surgical Mask is intended to be worn to protect both patients and healthcare workers against transfer of microorganisms, body fluids and particulate materials. This is a single use device, provided non-sterile.	When properly worn, the surgical face masks are intended to protect both patient and healthcare workers from transfer of microorganisms, body fluids and airborne particles. This device is non-sterile and for single use only.	Similar, only wording difference

Materials	Inner layer	Non-woven fabric	Spunbond Polypropylene	Similar
	Middle layer	Melt-blown fabric	Melt Blown Polypropylene Filter	Similar
	Outer layer	Non-woven fabric	Spunbond Polypropylene	Similar
	Nose piece	Polyethylene	Single Galvanize Wire, Coated By PE	Differences resolved by biocompatibility testing
	Headband	Polyester textured yarn	Not made with natural rubber latex	Differences resolved by biocompatibility testing
Mask style	Flat pleated	Flat pleated	Same	
Design feature	Earloop	Earloop	Same	
Dimensions	17.5cm×9.5cm	(17.5cm±0.5cm)×(9.0cm±0.5cm)	Similar	
Latex	Not made with natural rubber latex	Not made with natural rubber latex	Same	
Color	Blue	White	Differences resolved by biocompatibility testing	
Sterility	Non-sterile	Non-sterile	Same	
Use	Single use	Single use	Same	
Prescription or OTC	OTC	OTC	Same	
ASTM F2100 Level	Level 2	Level 2	Same	
<b>Performance test result</b>				
Fluid resistance	No penetration pass at 120 mmHg	30 out of 32 pass at 120 mmHg	Similar	
Particle Filtration Efficiency	Average 99.99%	99.9%	Same	
Bacterial Filtration Efficiency	Average > 99.8%	> 99.9%	Similar	
Flammability Class	Class 1	Class 1	Same	
Delta – P	Average 5.5 mmH <sub>2</sub> O/cm <sup>2</sup>	3.0 mmH <sub>2</sub> O/cm <sup>2</sup>	Similar, both masks met requirements of <6.0 mmH <sub>2</sub> O/cm <sup>2</sup>	
Biocompatibility	No cytotoxicity (ISO 10993- 5) No sensitization (ISO 10993- 10)	Non-Cytotoxic, Non-Sensitizing, Non-Irritating	Same	

	No irritation (ISO 10993-10)		
--	------------------------------	--	--

### (9) Non-clinical studies and tests performed

The following performance tests of Medical Surgical Mask were conducted:

Test Methodology	Purpose	Acceptance criteria	Results
Fluid Resistance Performance ASTM F1862-17	To evaluate the effectiveness of the test article in protecting the user from possible exposure to body fluids.	No penetration pass at 120mmHg	Pass, no penetration pass at 120mmHg
Particulate Filtration Efficiency ASTM F2299-17	To evaluate the effectiveness of the test article in protecting the user from possible exposure to particulates.	$\geq 98\%$	Pass, average 99.99%
Bacterial Filtration Efficiency ASTM F2101-19	To evaluate the bacterial filtration efficiency (BFE) of mask.	$\geq 98\%$	Pass, average > 99.8%
Differential Pressure (Delta P) EN 14683:2019, Annex C and ASTM F2100-19	To measure the differential pressure of mask which is related to breathability.	$< 6.0 \text{ mmH}_2\text{O}/\text{cm}^2$	Pass, average $5.5 \text{ mmH}_2\text{O}/\text{cm}^2$
Flammability 16 CFR 1610	To evaluate the flammability of mask.	Class 1	Pass, Class I
In vitro cytotoxicity ISO 10993-5	To evaluate the biological safety of the product which has direct contact with intact skin.	The test article should not have potential toxicity to L-929 in the MTT method.	Pass, the test article Surgical face mask has no potential toxicity to L-929 in the MTT method.
Skin sensitization ISO 10993-10	To evaluate the biological safety of the product which has direct contact with intact skin.	The test article should not cause delayed dermal contact sensitization in the guinea pig.	Pass, the test article showed no evidence of causing delayed dermal contact sensitization in the guinea pig. The test article Surgical face

			mask has no potential skin sensitization on guinea pigs in the extraction method.
Skin irritation ISO 10993-10	To evaluate the biological safety of the product which has direct contact with intact skin.	The irritation response category in the rabbit should be negligible.	Pass, the response of the test article extract was categorized as negligible under the test condition. The test article Surgical Face Mask has no potential skin irritation on rabbit in the extraction method.

**(10) Conclusion**

Based on the nonclinical tests performed, the subject device, Medical Surgical Mask, is as safe, as effective, and performs as well as the legally marketed predicate device, K200847, Avianz® Surgical Face Mask.