



November 14, 2021

Hubei Huanfu Plastic Products Co., Ltd
Ivy Wang
Technical Manager
Shanghai Sungo Management Consulting Company Limited
14th Floor, 1500# Central Avenue
Shanghai, Shanghai 200122
China

Re: K212575

Trade/Device Name: Surgical Face Mask
Regulation Number: 21 CFR 878.4040
Regulation Name: Surgical Apparel
Regulatory Class: Class II
Product Code: FXX
Dated: August 16, 2021
Received: August 16, 2021

Dear Ivy Wang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For Clarence W. Murray III, PhD
Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K212575

Device Name
Surgical Face Mask

Indications for Use (Describe)

The Surgical Face Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. This face mask is intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single use, disposable device, provided non-sterile.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) Summary

This summary of 510(K) safety and effectiveness information is being submitted in accordance with the requirement of 21 CFR 807.92.

Date prepared: November 8, 2021

A. Applicant Information:

HUBEI HUANFU PLASTIC PRODUCTS CO., LTD

Address: NO.188 WANFU STREET, CHENHU TOWN, HANCHUAN CTIYI, HUBEI, 431608 CHINA

Contact person: Ms. Yan Zhao

Title: Sales Manager

Tel: +86-13707291713

Fax: +86-712-8742198

Submission Correspondent:

Primary contact: Ms. Ivy Wang

Shanghai SUNGO Management Consulting Co., Ltd.

Room 1401, Dongfang Building, 1500# Century Ave., Shanghai 200122, China

Tel: +86-21-58817802

Email: haiyu.wang@sungoglobal.com

Secondary contact: Mr. Raymond Luo

Room 1401, Dongfang Building, 1500# Century Ave., Shanghai 200122, China

Tel: +86-21-68828050

Email: fda.sungo@gmail.com

B. Subject Device Information:

Trade Name: Surgical Face Mask

Common Name: SURGICAL MASK

Model: Ear Loops

Regulatory Information

Classification Name: Surgical Face Mask

Classification: Class II

Product code: FXX

Regulation Number: 878.4040

Review Panel: Surgical Apparel

C. Predicate device Information:

K182515

Surgical Face Mask

Wuhan Dymex Healthcare Co., Ltd.

Regulatory Information

Classification Name: Surgical Face Mask

Classification: Class II

Product code: FXX

Regulation Number: 878.4040

Review Panel: Surgical Apparel

D. Indications for use of the device:

The Surgical Face Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. This face mask is intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single use, disposable device, provided non-sterile.

E. Device Description:

The Surgical Face Mask is blue color, single use, three-layer, flat-folded masks with nose piece and ear loops. The blue colorant is polypropylene (PP) master batch.

The Surgical Face Mask is manufactured with three layers, the inner and outer layers are made of spun-bond polypropylene, and the middle layer is made of melt blown polypropylene filter.

The ear loops are held in place over the users’ mouth and nose by two elastic ear loops welded to the facemask. The elastic ear loops are made of Nylon and spandex, not made with natural rubber latex or fiberglass.

The nose piece in the layers of facemask is to allow the user to fit the facemask around their nose, which is made of Iron core coated with PE.

The surgical face masks are sold non-sterile and are intended to be single use, disposable devices.

F. Comparison with predicate device

Device	Proposed Device	Predicate Device	Result
510K #	K212575	K182515	-
Manufacturer	HUBEI HUANFU PLASTIC PRODUCTS CO., LTD	Wuhan Dymex Healthcare Co., Ltd.	-
Model Name	SURGICAL FACE MASK Ear loops	SURGICAL FACE MASK Ear loops	Similar
Classification	Class II Device, FXX (21 CFR878.4040)	Class II Device, FXX (21 CFR878.4040)	Same
Indication for use	The Surgical Face Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. This face mask is intended for use in infection	The Surgical Face Masks are intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These face masks are intended for use in infection control practices to reduce	Same

	control practices to reduce the potential exposure to blood and body fluids. This is a single use, disposable device, provided non-sterile.	the potential exposure to blood and body fluids. This is a single use, disposable device(s), provided non-sterile.		
Design Features	Ear Loops, Flat-pleated, 3 layers	Ear Loops, Flat-pleated, 3 layers	Same	
Materials	Outer layer	Spunbond Polypropylene	Spunbond Polypropylene	Same
	Inner layer	Spunbond Polypropylene	Spunbond Polypropylene	Same
	Filter layer	Melt-blown Polypropylene	Melt-blown Polypropylene	Same
	Nose wire	PE + Iron core	Malleable polyethylene wire	Different
	Ear loops	Nylon and spandex	Spandex	Different
Color	Blue	Yellow	Different	
Dimension (Length)	17.5cm±0.5cm	17.5cm±0.2cm	Same	
Dimension (Width)	9.5cm±0.5cm	9.5cm±0.2cm	Same	
OTC use	Yes	Yes	Same	
Sterility	Non-Sterile	Non-Sterile	Same	
Use	Single Use, Disposable	Single Use, Disposable	Same	
ASTM F2100 Level	Level 2	Level 2	Same	
Biocompatibility				
Cytotoxicity	Under the conditions of the study, the subject device extract was determined to be non-cytotoxic	Under the conditions of the study, the subject device extract was determined to be non-cytotoxic	Same	
Irritation	Under the conditions of the study, the subject device extract was determined to be non-irritant	Under the conditions of the study, the subject device extract was determined to be non-irritant	Same	
Sensitization	Under the conditions of the study, the subject device extract was determined to be non-sensitizer	Under the conditions of the study, the subject device extract was determined to be non-sensitizer	Same	

The proposed device has different material of nose piece, ear loops and different color to the predicate device, while the proposed device has been tested and the test results shown that the material and color differences do not affect the safety of the proposed device.

The proposed device is similar in design, intended use, technological characteristics, and is composed of the same or similar components as the predicate device. The product proposed under this premarket notification submission has the same or similar performance characteristics and conform to the same or similar standards.

G. Summary of Non-Clinical Test

Non-clinical tests were conducted to verify that the proposed device met all design specifications as was same to the predicate device. The test results demonstrated that the proposed device complies with the following standards and the requirements stated in the Guidance for Industry and FDA Staff: *Surgical Masks – Premarket Notification [510(k)] Submission* issued on March 5, 2004:

- ISO 10993-5: 2009 Biological Evaluation of Medical Devices -- Part 5: Tests For In Vitro Cytotoxicity
- ISO 10993-10: 2010 Biological Evaluation of Medical Devices - Part 10: Tests For Irritation And Skin Sensitization
- ASTM F2100, Standard Specification for Performance of Materials Used In Medical Face Masks
- ASTM F1862, Standard Test Method for Resistance of Medical Face Masks To Penetration by Synthetic Blood (Horizontal Projection of Fixed Volume At A Known Velocity);
- EN 14683, Medical Face Masks—Requirements and Test Methods;
- ASTM F2101, Standard Test Method for Evaluating the Bacterial Filtration Efficiency (BFE) Of Medical Face Mask Materials, Using A Biological Aerosol of Staphylococcus Aureus;
- ASTM F2299, Standard test method for determining the initial efficiency of materials used in medical face masks to penetration by particulates using latex spheres;
- 16 CFR 1610, Standard for the Flammability of clothing textiles;

Item	Proposed device	Acceptance Criteria: ASTM F2100 Level 2	Result
Fluid Resistance	32 out of 32 pass at 120 mmHg, 3 lots	29 out of 32 pass at 120 mmHg for level 3	PASS
Particulate Filtration Efficiency	>98%, 3 lots	≥ 98%	PASS
Bacterial Filtration Efficiency	>99%, 3 lots	≥ 98%	PASS
Differential Pressure	<3.2mmH ₂ O/cm ² , 3 lots	< 6.0mmH ₂ O/cm ²	PASS
Flammability	Class 1, 3 lots	Class 1	PASS
Cytotoxicity	Under the conditions of the study, the device is non-cytotoxic.	Non-cytotoxic	PASS
Irritation	Under the conditions of the study, the device is non-irritating.	Non-irritating	PASS
Sensitization	Under the conditions of the study, the device is non-sensitizing	Non-sensitizing	PASS

H. Summary of Clinical Test Conclusion

No clinical study is included in this submission.

I. Conclusion

The conclusion drawn from the nonclinical tests demonstrates that the subject device in 510(K) submission, the Surgical Face Mask is as safe, as effective, and performs as well as or better than the legally marketed predicate device cleared under K182515.