



September 27, 2022

Yangzhou Fartley Medical Instrument Technology Co., Ltd.
% Ethan Liu
RA Specialist
Shanghai Thinkwell Consulting Co., Ltd
Room 211/6F, Xinling Road, Minhang District
Shanghai, Shanghai 201100
China

Re: K220785
Trade/Device Name: Clean Connecting Tube
Regulation Number: 21 CFR§ 876.1500
Regulation Name: Endoscope and Accessories
Regulatory Class: II
Product Code: OCX
Dated: August 23, 2022
Received: August 24, 2022

Dear Ethan Liu:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shanil P. Haugen, Ph.D.
Assistant Director
DHT3A: Division of Renal, Gastrointestinal,
Obesity and Transplant Devices
OHT3: Office of GastroRenal, ObGyn,
General Hospital and Urology Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K220785

Device Name
Clean Connecting Tube

Indications for Use (Describe)

The Clean Connecting Tube (tubing and accessories to accommodate various GI endoscopes and irrigation pumps) is intended to provide irrigation via sterile water and to supply air(via an air pump) or CO2(Via a CO2 pump) along with sterile water during GI endoscopic procedures when used in conjunction with an irrigation pump or electrosurgical unit.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

This summary of 510(k) safety and effectiveness information is being submitted in accordance with requirements of 21 CFR Part 807.92.

5.1 Submitter

Submitted by:	Yangzhou Fartley Medical Instrument Technology Co., Ltd. Address: Beizhou Road, Lidian Town, Guangling District, Yangzhou 225106 Jiangsu, China
Contact Person:	Ethan Liu RA Specialist Shanghai Thinkwell Consulting Co., Ltd Address: Room 211/6F, Xinling Road, Minhang District, Shanghai, China. Phone: 0086-15216699240 Email: xtdeepwater@126.com
Date Prepared:	August 22, 2022

5.2 Device

Device Name:	Clean Connecting Tube
Classification Name:	Endoscopic Irrigation/Suction System
Regulatory Class:	II
Regulation Number:	21 CFR 876.1500
Regulation Name:	Endoscope and accessories
Product Code:	OCX

5.3 Predicate Device

Device Name:	Universal Irrigation Solution Hybird™ K102855
Manufacturer:	BYRNE MEDICAL, INC.
Classification Name:	Endoscopic Irrigation/Suction System
Regulatory Class:	II
Regulation Number:	21 CFR 876.1500
Regulation Name:	Endoscope and accessories
Product Code:	OCX

5.4 Device Description

The Clean Connecting Tube is intended for 24 hour multiple-patient use and then discarded. The Clean Connecting Tube design minimizes infection control risks that are associated with manual cleaning and sterilization. The maximum number of patient

is up to 15 in 24 hours , and maximum duration of us shall not exceed 15h. The check valve attached to the irrigation tubing shall be replaced after every patient use.

The Clean Connecting Tube is manufactured for use in conjunction with sterile water bottle, and together with irrigation pump. The Clean Connecting Tube is designed to be attached to the auxiliary water connector or biopsy irrigation accessory and to be inserted into pump head of the irrigation pump to provide irrigation pump to provide irrigation through the auxiliary water channel to the distal end of endoscope.

5.5 Indication for Use:

The Clean Connecting Tube (tubing and accessories to accommodate various GI endoscopes and irrigation pumps) is intended to provide irrigation via sterile water and to supply air(via an air pump) or CO₂(Via a CO₂ pump) along with sterile water during GI endoscopic procedures when used in conjunction with an irrigation pump or electrosurgical unit.

5.6 Comparison of Technological Characteristics

The Clean Connecting Tube has substantially equivalent device design, configuration, packaging fundamental technology, sterilization process and intended use as those featured in the predicate device BYRNE MEDICAL, INC’s Universal Irrigation Solution Hybird™, K102855. The differences between the proposed device and the predicate devices do not raise any questions regarding its safety and effectiveness. The differences are listed in the table below.

Item	Clean Connecting Tube(Proposed Device)	Universal Irrigation Solution Hybird™, K102855	Discussion
Indication for Use	The Clean Connecting Tube (tubing and accessories to accommodate various GI endoscopes and irrigation pumps) is intended to provide irrigation via sterile water and to supply air(via an air pump) or CO ₂ (Via a CO ₂ pump) along with sterile water during GI endoscopic procedures when used in conjunction with an irrigation pump or electrosurgical unit.	The Universal Irrigation Solution Hybrid™ (tubing and accessories to accommodate various GI endoscopes and irrigation pumps) is intended to provide irrigation via sterile water and to supply air(via an air pump) or CO ₂ (Via a CO ₂ pump) along with sterile water during GI endoscopic procedures when used in conjunction with an	Substantially equivalent

Item	Clean Connecting Tube(Proposed Device)	Universal Irrigation Solution Hybird™, K102855	Discussion
		irrigation pump or Cautery unit.	
Product Code	OCX	OCX	Same
Classification	2	2	Same
Regulation Number	876.1500	876.1500	Same
Single Use	Yes(24 hr.)	Yes(24 hr.)	Same
Patient Population	Male/Female, Pediatric to Adult	Male/Female, Pediatric to Adult	Same
Reusable or Disposable	Disposable	Disposable	Same
SAL	10 ⁻⁶	10 ⁻⁶	Same
Biocompatibility	Comply with ISO 10993-1	Comply with ISO 10993-1	Same
Sterilization Method	EO Sterilization	EO Sterilization	Same

5.7 Non-clinical Performance Data

The proposed device meets the requirements of ISO 10993 “Biological Evaluation of Medical Devices – Part 1: Evaluation and Testing”, ISO 11135-1 “Sterilization of Health Care products Ethylene Oxide - Part 1: Requirements for Development, Validation, and Routine Control of Sterilization processes for Medical Devices”, and ISO 10993-7 “Biological evaluation of medical devices - Part 7: ethylene oxide sterilization residuals” .

The following bench tests were performed on Clean Connecting Tube: Appearance, Physical properties. The results of all testing were passing.

5.8 Clinical Test Data

No Clinical Study is included in this submission.

5.9 Conclusion

In accordance with the Federal Food, Drug and Cosmetic Act, 21 CFR Part 807, Based on the information provided in this premarket notification, Yangzhou Fartley Medical Instrument Technology Co., Ltd. has demonstrated that proposed device Clean Connecting Tube is substantially equivalent to BYRNE MEDICAL, INC’s currently marketed Universal Irrigation Solution Hybird™, K102855.