



September 30, 2022

Arthrex Inc.  
Stacy Valdez  
Senior Regulatory Affairs Specialist  
1370 Creekside Boulevard  
Naples, Florida 34108-1945

Re: K222244

Trade/Device Name: Arthrex 3.5 mm Locking Compression Plates

Regulation Number: 21 CFR 888.3030

Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And Accessories

Regulatory Class: Class II

Product Code: HRS, HWC

Dated: July 25, 2022

Received: July 26, 2022

Dear Stacy Valdez:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531 - 542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali, M.P.H.  
Assistant Director  
DHT6C: Division of Restorative, Repair  
and Trauma Devices  
OHT6: Office of Orthopedic Devices  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K222244

Device Name

Arthrex 3.5 mm Locking Compression Plates

Indications for Use (Describe)

The Arthrex 3.5 mm Locking Compression Plates are intended to be used for internal bone fixation for bone fractures, fusions, osteotomies and non-unions in the ankle, foot, wrist, clavicle, scapula, olecranon, humerus, radius, ulna, tibia, calcaneus, fibula.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

## 510(k) Summary

<b>Date Prepared</b>	September 26, 2022
<b>Submitter</b>	Arthrex Inc. 1370 Creekside Boulevard Naples, FL 34108-1945
<b>Contact Person</b>	Stacy Valdez Senior Regulatory Affairs Specialist 1-239-643-5553, ext. 72010 Stacy.valdez@arthrex.com
<b>Name of Device</b>	Arthrex 3.5 mm Locking Compression Plates
<b>Common Name</b>	Plate, fixation, bone
<b>Product Code</b>	HRS, HWC
<b>Classification Name</b>	21 CFR 888.3030: Single/multiple component metallic bone fixation appliances and accessories (Primary) 21 CFR 888.3040: Smooth or threaded metallic bone fixation fastener
<b>Regulatory Class</b>	II
<b>Predicate Device</b>	K123241: Arthrex Fracture Plates
<b>Reference Device(s)</b>	K203294: Arthrex Pilon Fusion System K151732: Arthrex Fracture Plates
<b>Purpose of Submission</b>	This Traditional 510(k) premarket notification is submitted to obtain clearance for the Arthrex 3.5 mm Locking Compression Plates.
<b>Device Description</b>	The Arthrex 3.5 mm Locking Compression Plates consists of a series of plates of varying lengths. The Arthrex 3.5 mm Locking Compression Plates range from 70 mm to 278 mm in length and 3.5 mm in thickness. Each plate provides locking screw fixation. The proposed plates are manufactured from Titanium Alloy (ASTM F136) and Stainless Steel (ASTM F138). The proposed plates are sold as sterile (Gamma), single-use, and non-sterile, single-use.
<b>Indications for Use</b>	The Arthrex 3.5 mm Locking Compression Plates are intended to be used for internal bone fixation for bone fractures, fusions, osteotomies and non-unions in the ankle, foot, wrist, clavicle, scapula, olecranon, humerus, radius, ulna, tibia, calcaneous, fibula.
<b>Performance Data</b>	<p>Arthrex conducted 4-Point Bend (ASTM F382-17) testing to demonstrate that the Arthrex 3.5 mm Locking Compression Plates perform statistically equivalent to the predicate devices cleared under Arthrex Fracture Plates (K123241).</p> <p>MRI force, torque, and image artifact testing were conducted in accordance with FDA guidance <i>Testing and Labeling Medical Devices for Safety in the Magnetic Resonance (MR) Environment</i>, ASTM F2052 <i>Standard Test Method for Measurement of Magnetically Induced Displacement Force on Medical Devices in the Magnetic Resonance Environment</i>, ASTM F2119 <i>Standard Test Method for Evaluation of MR Image Artifacts from Passive Implants</i>, ASTM F2182 <i>Standard Test Method for Measurement of Measurement of Radio Frequency Induced Heating Near Passive Implants During Magnetic Resonance Imaging</i> and ASTM F2213 <i>Standard Test Method for Measurement of Magnetically Induced Torque on Medical Devices in the Magnetic Resonance Environment</i>.</p> <p>Arthrex Plates are tested for Bacterial Endotoxins Test (BET) utilizing the Kinetic Chromogenic Method in accordance with ANSI/AAMI ST72:2011/(R)2016, USP &lt;161&gt;, USP &lt;85&gt;, EP 2.6.14. The testing conducted demonstrates that the sterile devices within the Arthrex Plates meet pyrogen limit specifications.</p>

	<p>Assessment of physical product attributes including product, design, size, and materials has determined that the Arthrex 3.5 mm Locking Compression Plates do not introduce additional risks or concerns regarding sterilization and shelf-life.</p>
<p><b>Technological Comparison</b></p>	<p>The Arthrex 3.5 mm Locking Compression Plates are substantially equivalent to the predicate devices cleared under K123241 in which the basic design features, intended use, fundamental scientific technology, materials (stainless steel only), shelf-life, and sterility are identical.</p> <p>The Arthrex 3.5 mm Locking Compression Plates are manufactured from Titanium Alloy conforming to ASTM F136 and Stainless Steel conforming to ASTM F138. The primary predicate plates cleared under the Arthrex Fracture Plates, K123241 are manufactured from CP Grade 4 Titanium conforming to ASTM F67 and Stainless Steel conforming to ASTM F138. The proposed Arthrex 3.5 mm Locking Compression Plates are 3.5 mm in thickness and range from 70 mm to 278 mm in length. The primary predicate plates cleared under the Arthrex Fracture Plates, K123241 are offered in thicknesses ranging from 1.0 mm to 2.0/3.6 mm.</p> <p>The sterile Arthrex 3.5 mm Locking Compression Plates are packaged in a double Nylon/Nylon pouch which is equivalent to the packaging configuration cleared under reference device Arthrex Pilon Fusion System, K203294. The non-sterile Arthrex 3.5 mm Locking Compression Plates are packaged in a Zip-Lock Polyethylene Bag or a single Polyethylene Bag. The primary predicate plates cleared under Arthrex Fracture Plates, K123241 are packaged in a polyethylene pouch.</p> <p>The Arthrex 3.5 mm Locking Compression Plates were evaluated for MR Conditional labeling as were the reference predicate devices cleared under K203294.</p> <p>The 3.5 mm Locking Compression Plates are substantially equivalent to the predicate devices cleared under K123241, with minor modifications with no change to intended use or function. Any differences between the Arthrex 3.5 mm Locking Compression Plates and the predicate devices are considered minor and do not raise different questions of safety or effectiveness.</p>
<p><b>Conclusion</b></p>	<p>The Arthrex 3.5 mm Locking Compression Plates are substantially equivalent to the predicate devices cleared under K123241 in which the basic design features and intended use are the same. Any differences between the Arthrex 3.5 mm Locking Compression Plates and the predicate devices are considered minor and do not raise different questions of safety or effectiveness.</p> <p>The submitted mechanical testing data demonstrates that the 4-Point Bend strength of the Arthrex 3.5 mm Locking Compression Plates are substantially equivalent to that of the predicate devices for the desired indications.</p> <p>Based on the indications for use, technological characteristics, and the summary of data submitted, Arthrex Inc. has determined that the proposed device is substantially equivalent to the currently marketed predicate devices.</p>