



September 29, 2022

Canon Inc.
% Ms. Saori Sawaki
Business Manager, Regulatory Consultant
Ken Block Consulting LLC
800 E Campbell Road,
Suite 202
RICHARDSON TX 75081

Re: K222687

Trade/Device Name: Digital Radiography CXDI-Elite
Digital Radiography E1
Regulation Number: 21 CFR 892.1680
Regulation Name: Stationary X-Ray System
Regulatory Class: Class II
Product Code: MQB
Dated: September 6, 2022
Received: September 6, 2022

Dear Ms. Saori Sawaki:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Laurel Burk, Ph.D.
Assistant Director
Diagnostic X-Ray Systems Team
DHT8B: Division of Imaging Devices
and Electronic Products
OHT8: Office of Radiological Health
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K222687

Device Name

DIGITAL RADIOGRAPHY CXDI-Elite
DIGITAL RADIOGRAPHY E1

Indications for Use (Describe)

The DIGITAL RADIOGRAPHY CXDI-Elite / E1 provides digital image capture for conventional film/screen radiographic examinations. This device is intended to capture, for display, radiographic images of human anatomy, and to replace radiographic film/screen systems in all general purpose diagnostic procedures. This device is not intended for mammography applications.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) SUMMARY

Applicant/ Sponsor: Canon Inc.
9-1 Imaikami-cho
Nakahara-ku, Kanagawa 211-8501 JAPAN

Contact Person: Mr. Akira Hirai
General Manager
TEL: 81-3-3758-2111
FAX: 044-739-6695
hirai.akira@mail.canon

Date Prepared: September 15, 2022

Proposed Device Manufacturer: Canon Inc.
Trade Name: DIGITAL RADIOGRAPHY CXDI-Elite
DIGITAL RADIOGRAPHY E1
Common Name: Solid State X-Ray Imager (Flat Panel/Digital Imager)
Classification Name: Stationary X-ray System
Product Code: MQB
Regulation: 892.1680, Stationary X-ray System

Predicate Device: Clearance: K221620
Manufacturer: Canon Inc.
Trade Name: DIGITAL RADIOGRAPHY CXDI-Elite
DIGITAL RADIOGRAPHY E1
Common Name: Solid State X-ray Imager (Flat Panel/Digital Imager)
Classification Name: Stationary X-ray System
Product Code: MQB
Regulation: 892.1680, Stationary X-ray System

Device Description: The DIGITAL RADIOGRAPHY CXDI-Elite, also called the DIGITAL RADIOGRAPHY E1, (hereinafter referred to as CXDI-Elite) is a solid-state x-ray imager. The CXDI-Elite is a series of detectors, and in the predicate device consists of the CXDI-720C Wireless detector (also called the AR-E3543W detector), the CXDI-420C Wireless detector (also called the AR-E4343W detector), and the CXDI-420C Fixed detector (also called the AR-E4343S detector). The detector intercepts x-ray photons, and the scintillator emits visible spectrum photons that illuminate an array of photodetectors that create electrical signals. After the electrical signals are generated, the signals are converted to digital values. The digital values are sent to the PC via a wired or wireless connection, converted to images with the CXDI Control Software, and then displayed on the PC/monitors. The PC/monitors used with the CXDI-Elite are not a part of this submission.

The proposed changes to the predicate device, DIGITAL RADIOGRAPHY CXDI-Elite (K221620), include the addition of a new detector, the CXDI-820C Wireless (also called the AR-E2735W), to the CXDI-Elite series; the standardization of the thickness of CXDI-720C Wireless and CXDI-420C Wireless; the firmware update from F2.00.03.00 to 02.01.04.00; the CXDI Control Software version update from 3.10.2.2 to 3.10.2.6; and the addition of a 25 m of the CXDI-420C Fixed detector cable and a 10 m of Wiring Cable (WC-01). Together, these changes make up the CXDI-Elite.



510(k) SUMMARY

Indications for Use: The DIGITAL RADIOGRAPHY CXDI-Elite / E1 provides digital image capture for conventional film/screen radiographic examinations. This device is intended to capture, for display, radiographic images of human anatomy, and to replace radiographic film/screen systems in all general purpose diagnostic procedures. This device is not intended for mammography applications.

Summary of Technological Characteristics: Comparison with the predicate device shows the technological characteristics of the CXDI-Elite is substantially equivalent to the predicate device. The flat panel detector units are functionally same, using the same components.

	Proposed Device	Predicate Device	
Trade Name	DIGITAL RADIOGRAPHY CXDI-Elite / E1	DIGITAL RADIOGRAPHY CXDI-Elite / E1	
510(k) Submitter [Number]	Canon Inc. [K222687]	Canon Inc. [K221620]	IDENTICAL
Indication for Use	The DIGITAL RADIOGRAPHY CXDI-Elite / E1 provides digital image capture for conventional film/screen radiographic examinations. This device is intended to capture, for display, radiographic images of human anatomy, and to replace radiographic film/screen systems in all general purpose diagnostic procedures. This device is not intended for mammography applications.	The DIGITAL RADIOGRAPHY CXDI-Elite / E1 provides digital image capture for conventional film/screen radiographic examinations. This device is intended to capture, for display, radiographic images of human anatomy, and to replace radiographic film/screen systems in all general purpose diagnostic procedures. This device is not intended for mammography applications.	IDENTICAL
Application	General Radiography	General Radiography	IDENTICAL
Detector	CXDI-720C Wireless (AR-E3543W) CXDI-420C Wireless (AR-E4343W) CXDI-820C Wireless (AR-E2735W) CXDI-420C Fixed (AR-E4343S)	CXDI-720C Wireless (AR-E3543W) CXDI-420C Wireless (AR-E4343W) CXDI-420C Fixed (AR-E4343S)	MODIFIED
Components	<u>CXDI-720C Wireless,</u> <u>CXDI-420C Wireless, and</u> <u>CXDI-820C Wireless:</u> Detector (Sensor) Battery Pack (LB-4A) Battery Charger (BC-01/BC-1A) Multi Box (MB-02) Status Indicator (SI-01/SI-4A) Wiring Cable, 1.5/7.5/10/15/25m (WC-01) PC Connection Cable (CP-01) Ready Indicator (RI-3A) X-ray Interface Box (XB-1A) Detector Stand (DS-01) <u>CXDI-420C Fixed:</u> Detector (Sensor) Multi Box (MB-02) Status Indicator (SI-01/SI-4A) Ready Indicator (RI-3A) X-ray Interface Box (XB-1A)	<u>CXDI-720C Wireless and</u> <u>CXDI-420C Wireless:</u> Detector (Sensor) Battery Pack (LB-4A) Battery Charger (BC-01/BC-1A) Multi Box (MB-02) Status Indicator (SI-01/SI-4A) Wiring Cable, 1.5/7.5/15/25m (WC-01) PC Connection Cable (CP-01) Ready Indicator (RI-3A) X-ray Interface Box (XB-1A) Detector Stand (DS-01) <u>CXDI-420C Fixed:</u> Detector (Sensor) Multi Box (MB-02) Status Indicator (SI-01/SI-4A) Ready Indicator (RI-3A) X-ray Interface Box (XB-1A)	MODIFIED
Other Compatible Components	Power Box (PB-01)	Power Box (PB-01)	IDENTICAL
Software	CXDI Control Software V3.10.2.6	CXDI Control Software V3.10.2.2	MODIFIED
Detector Firmware	02.01.04.00	F2.00.03.00	MODIFIED
Pixel Pitch	125µm	125µm	IDENTICAL
Scintillator	CsI(Tl) [Cesium Iodide doped with Thallium]	CsI(Tl) [Cesium Iodide doped with Thallium]	IDENTICAL
Spatial Resolution	45% [MTF@2lp/mm]	45% [MTF@2lp/mm]	IDENTICAL
DQE	67% [@0.5 lp/mm, 3.5 uGy]	67% [@0.5 lp/mm, 3.5 uGy]	IDENTICAL

510(k) SUMMARY

IP Level	CXDI-720C Wireless: IP57 CXDI-420C Wireless: IP57 CXDI-820C Wireless: IP57 CXDI-420C Fixed: N/A	CXDI-720C Wireless: IP57 CXDI-420C Wireless: IP57 CXDI-420C Fixed: N/A	MODIFIED
Wireless Communication	IEEE 802.11a/b/g/n/ac 2.4GHz/5GHz Bluetooth Low Energy (only for CXDI-720C Wireless, CXDI-420C Wireless, and CXDI-820C Wireless)	IEEE 802.11a/b/g/n/ac 2.4GHz/5GHz Bluetooth Low Energy (only for CXDI-720C Wireless and CXDI-420C Wireless)	IDENTICAL
External Dimensions	CXDI-720C Wireless: 384 x 460 x 15.7 mm CXDI-420C Wireless: 460 x 460 x 15.7 mm CXDI-820C Wireless: 384 x 307.5 x 15.7mm CXDI-420C Fixed: 460 x 460 x 15.3 mm	CXDI-720C Wireless: 384 x 460 x 15.5 mm CXDI-420C Wireless: 460 x 460 x 15.5 mm CXDI-420C Fixed: 460 x 460 x 15.3 mm	MODIFIED
Photographing Mode	Standard Synchronization Mode, Standard Synchronization Mode with Built in AEC Assistance, Non Generator Connection Mode, Standalone Mode (only for CXDI-720C Wireless, CXDI-420C Wireless, and CXDI-820C Wireless)	Standard Synchronization Mode, Standard Synchronization Mode with Built in AEC Assistance, Non Generator Connection Mode, Standalone Mode (only for CXDI-720C Wireless and CXDI-420C Wireless)	IDENTICAL

Summary of Non-Clinical / Test Data:

The fundamental scientific technology of the CXDI-Elite has not been modified. The change is the addition of new detector CXDI-820C Wireless (AR-E2735W).

Evaluation of the addition of the new detector confirmed that the change did not impact CXDI-Elite conformance with the U.S. Performance Standard for radiographic equipment and with relevant voluntary safety standards for Electrical safety and Electromagnetic Compatibility testing, specifically IEC standards 60601-1, 60601-1-2, 60601-1-3, 60601-2-54, and 60601-1-6.

These verification/validation activities successfully demonstrated that the device continues to meet the standards for the areas impacted by the modifications to the predicate device and raises no new questions regarding either safety or effectiveness when compared to the predicate device. Clinical testing is not necessary for the current submission, based on the minor differences from the predicate device. Adequate detector bench testing should be sufficient to demonstrate that the subject detector CXDI-Elite / E1 works as intended. Therefore, the verification/validation activities conducted support a determination of substantial equivalence for the CXDI-Elite.

The proposed device follows the applicable elements of the following FDA guidance documents: *Guidance for the Submission of 510(k)s for Solid State X-ray Imaging Devices*, *Guidance for the Content of Premarket Submissions for Software contained in Medical Devices*, *Content of Premarket Submissions for Management of Cybersecurity in Medical Devices*, *Use of International Standard ISO 10993-1, "Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process"*, *Radio Frequency Wireless Technology in Medical Devices*, and *Pediatric Information for X-ray Imaging Device Premarket Notifications*.

Conclusion:

Canon Inc. considers the DIGITAL RADIOGRAPHY CXDI-Elite / E1 device to be substantially equivalent to the predicate device listed above. This conclusion is based on the similarities in primary intended use, principles of operation, functional design, and established medical use.