



January 15, 2020

Arthrex Inc.
Jessica Singelais
Regulatory Affairs Specialist
1370 Creekside Boulevard
Naples, Florida 34108-1945

Re: K193503

Trade/Device Name: Arthrex SwiveLock Suture Anchor
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And Accessories
Regulatory Class: Class II
Product Code: MAI, MBI
Dated: December 12, 2019
Received: December 18, 2019

Dear Ms. Singelais:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

for

Laura Rose, Ph.D.
Acting Assistant Director
DHT6C: Division of Restorative, Repair
and Trauma Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug AdministrationForm Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.**Indications for Use**

510(k) Number (if known)

K193503

Device Name

Arthrex SwiveLock Suture Anchor

Indications for Use (Describe)

The Arthrex SwiveLock Suture Anchors are intended for fixation of suture (soft tissue) to bone in the shoulder, foot/ankle, knee, hand/wrist, elbow, and hip in the following procedures:

- Shoulder: Rotator Cuff Repair, Bankart Repair, SLAP Lesion Repair, Biceps Tenodesis, Acromio-Clavicular Separation Repair, Deltoid Repair, Capsular Shift or Capsulolabral Reconstruction.
- Foot/Ankle: Lateral Stabilization, Medial Stabilization, Achilles Tendon Repair, Hallux Valgus Reconstruction, Mid-foot Reconstruction, Metatarsal Ligament Repair/Tendon Repair and Bunionectomy.
- Knee: Anterior Cruciate Ligament Repair (4.75-5.5 SwiveLock Only), Medial Collateral Ligament Repair, Lateral Collateral Ligament Repair, Patellar Tendon Repair, Posterior Oblique Ligament Repair, Iliotibial Band Tenodesis, and Quadriceps Tendon Repair. Secondary or adjunct fixation for ACL/PCL reconstruction or repair (4.75-5.5 SwiveLock only).
- Hand/Wrist: Scapholunate Ligament Reconstruction, Ulnar or Radial Collateral Ligament Reconstruction.
- Elbow: Biceps Tendon Reattachment, Tennis Elbow Repair, Ulnar or Radial Collateral Ligament Reconstruction, Lateral Epicondylitis repair.
- Hip: Capsular repair, acetabular labral repair.

Type of Use (Select one or both, as applicable)

 Prescription Use (Part 21 CFR 801 Subpart D) Over-The-Counter Use (21 CFR 801 Subpart C)**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

<i>Date Prepared</i>	January 15, 2020
<i>Submitter</i>	Arthrex Inc. 1370 Creekside Boulevard Naples, FL 34108-1945
<i>Contact Person</i>	Jessica L. Singelais Regulatory Affairs Specialist 1-239-598-4302, ext. 73091 Jessica.singelais@arthrex.com
<i>Name of Device</i>	Arthrex SwiveLock Suture Anchor
<i>Common Name</i>	Smooth or threaded metallic bone fixation fastener
<i>Product Code</i>	MAI, MBI
<i>Classification Name</i>	21 CFR 888.3040: Smooth or threaded metallic bone fixation fastener 21 CFR 888.3030: Single/multiple component metallic bone fixation appliances and accessories
<i>Regulatory Class</i>	II
<i>Predicate Device</i>	K191226: Arthrex SwiveLock Suture Anchor
<i>Reference Device</i>	K173240: Arthrex PushLock Suture Anchor
<i>Purpose of Submission</i>	This Special 510(k) premarket notification is submitted to add a line extension to the Arthrex SwiveLock Suture Anchors cleared under predicate K191226.
<i>Device Description</i>	The Arthrex SwiveLock Anchor is a sterile two-component suture anchor comprised of an eyelet and a hollow anchor body. The Arthrex SwiveLock Anchor is pre-mounted on a driver with the anchor body and eyelet physically separated on the driver shaft. Arthrex 510(k) cleared suture may also be provided with the device.
<i>Indications for Use</i>	The Arthrex SwiveLock Suture Anchors are intended for fixation of suture (soft tissue) to bone in the shoulder, foot/ankle, knee, hand/wrist, elbow, and hip in the following procedures: <ul style="list-style-type: none"> • Shoulder: Rotator Cuff Repair, Bankart Repair, SLAP Lesion Repair, Biceps Tenodesis, Acromio-Clavicular Separation Repair, Deltoid Repair, Capsular Shift or Capsulolabral Reconstruction. • Foot/Ankle: Lateral Stabilization, Medial Stabilization, Achilles Tendon Repair, Hallux Valgus Reconstruction, Mid-foot Reconstruction, Metatarsal Ligament Repair/Tendon Repair and Bunionectomy. • Knee: Anterior Cruciate Ligament Repair (4.75-5.5 SwiveLock Only), Medial Collateral Ligament Repair, Lateral Collateral Ligament Repair, Patellar Tendon Repair, Posterior Oblique Ligament Repair and Iliotibial Band Tenodesis and Quadriceps Tendon Repair, Secondary fixation for ACL/PCL reconstruction or repair (4.75 – 5.5 SwiveLock only). • Hand/Wrist: Scapholunate Ligament Reconstruction, Ulnar or Radial Collateral Ligament Reconstruction. • Elbow: Biceps Tendon Reattachment, Tennis Elbow Repair, Ulnar or Radial Collateral Ligament Reconstruction. • Hip: Capsular repair, acetabular labral repair.
<i>Comparison Summary of Technological Characteristics and Modifications Proposed</i>	The proposed device is a line extension to the predicate device. The proposed and predicate devices (K191226) have the same basic design, intended use, indications for use, packaging and sterilization method. In comparison to the predicate device, the proposed device has an extended shelf life. Proposed modifications consist of a modified eyelet, knotless suture construct and suture manufacturing process changes. Any differences between the proposed device and the predicate device are considered minor and do not raise questions concerning safety or effectiveness.

Performance Data

Cyclic pull-out, biocompatibility testing per ISO 10993-1:2018 and pyrogenicity testing was conducted on a representative device and compared to the predicate device to demonstrate that the proposed devices are substantially equivalent to the predicate.

Conclusion

The proposed Arthrex SwiveLock Suture Anchors are substantially equivalent to the predicate device in which the basic design features and intended uses are the same. Any differences between the proposed device and the predicate device are considered minor and do not raise questions concerning safety or effectiveness.

Based on the indications for use, technological characteristics, and the summary of data submitted, Arthrex Inc. has determined that the Arthrex SwiveLock is substantially equivalent to the currently marketed predicate device.