



August 17, 2021

Jiangsu Health-Peace Medical Technology Co. Ltd
% Johnson Lui
Consultant
CNMED Consultant
31 Archer St
Upper MT Gravatt, QLD 4122
Australia

Re: K211708

Trade/Device Name: Disposable Medical Nitrile Gloves
Regulation Number: 21 CFR 880.6250
Regulation Name: Non-Powdered Patient Examination Glove
Regulatory Class: Class I, reserved
Product Code: LZA
Dated: July 30, 2021
Received: August 3, 2021

Dear Johnson Lui:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

For Clarence W. Murray III, Ph.D.
Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K211708

Device Name

Disposable Medical Nitrile Gloves

Indications for Use (Describe)

Disposable Medical Nitrile Gloves are disposable device intended for medical purpose that is worn on the examiner's hand to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) Summary

K211708

Disposable Medical Nitrile Gloves

Preparation Date: Aug 18th, 2021

1. SUBMITTER

Company Name: JIANGSU HEALTH-PEACE MEDICAL TECHNOLOGY CO. LTD
Company Address: DongAn Youdian Street, Huangli Town, Wujin, Changzhou , CN 213155

Contact Person: Johnson Liu
Telephone Number: +614-0158-9995
Email: Johnson@cnmed.com.au

2. NAME OF THE DEVICE

Trade Name / Proprietary Name: Disposable Medical Nitrile Gloves
Device Name: Disposable Medical Nitrile Gloves
Device Classification Name: Patient Examination Gloves
Device Class: Class I
Device Classification Number: 21 CFR 880.6250
Product Code: LZA

3. IDENTIFICATION OF THE LEGALLY MARKETED DEVICE

Predicate Device: K203191
Applicant: Nathan Trading Co., Ltd
Device Name: LYDUS NITRILE EXAMINATION GLOVES, POWDER FREE
Device Classification Name: Patient Examination Gloves Device
Classification Number: 21 CFR 880.6250
Device Class: Class I
Product Code: LZA
Review Panel: General Hospital

4. DEVICE DESCRIPTION

The subject device in this 510(k) Notification is Disposable Medical Nitrile Gloves. The subject device is a patient examination glove made from nitrile compound, blue color, powder free and non-sterile (Per 21 CFR 880.6250, Class I). The device meets the specifications in ASTM D6319-19 *Standard Specification for Nitrile Examination Gloves for Medical Application*.

5. INTENDED USE/Indications for Use

Disposable Medical Nitrile Gloves is a disposable device intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner.

6. TECHNOLOGICAL CHARACTERISTIC COMPARISON FOR THE PROPOSED AND PREDICATE DEVICES

CHARACTERISTICS	DEVICE PERFORMANCE		REMARKS
	SUBJECT	PREDICATE	
510(k) Number	K211708	K203191	-
Device Name	Disposable Medical Nitrile Gloves	LYDUS Nitrile Examination Gloves, Powder Free	-
Product Code	LZA	LZA	Same
Intended Use	Disposable Medical Nitrile Gloves is disposable devices intended for medical purpose that are worn on the examiner's hand to prevent contamination between patient and examiner. The device is for over the-counter use.	LYDUS Nitrile Examination Gloves, Powder Free is a disposable device intended for medical purposes that is worn on the examiner's hand or fingers to prevent contamination between patient and examiner.	Same
Materials of Use (ASTM D6910/D6910M-19)	Nitrile compound	Nitrile compound	Same
Color	Blue	Blue	Same
Texture	Finger Textured	Finger Textured	Same
Size (ASTM D6319-19)	Small, Medium, Large,	Small, Medium, Large, Extra Large	Same
Sterilization	Non-sterile	Non-sterile	Same
Usage	Single usage	Single usage	Same
Dimensions (ASTM D6319-19)	Length Min. 220 mm Width Min 80+/-10 mm (for small size) Length Min. 230 mm Width Min 95+/-10 mm (for medium size) Length Min. 230 mm Width Min 110+/-10 mm (for large size)	Length Min. 220 mm Width Min 80+/-10 mm (for small size) Length Min. 230 mm Width Min 95+/-10 mm (for medium size) Length Min. 230 mm Width Min 110+/-10 mm (for large size)	Same
Physical Properties (ASTM D6319-19)	<u>Before Aging</u> Tensile Strength Min 14 Mpa Ultimate Elongation Min 500% <u>After Aging</u> Tensile Strength Min 14 Mpa Ultimate Elongation Min 400%	<u>Before Aging</u> Tensile Strength Min 14 Mpa Ultimate Elongation Min 500% <u>After Aging</u> Tensile Strength Min 14 Mpa Ultimate Elongation Min 400%	Same
Thickness (ASTM D6319-19)	Palm min 0.05 mm Finger min 0.05 mm	Palm min 0.05 mm Finger min 0.05 mm	Same
Powder Free (ASTM D6319-19)	≤2 mg/glove	≤2 mg/glove	Same

Freedom from Holes (Water Tight -1000 ml) – ASTM D6319-19 (Cross Reference D5151)	Passed	Passed	Same
Biocompatibility - SKIN SENSITIZATION - ISO 10993-10: 2010 (E)	Under the conditions of study not a sensitizer	Under the conditions of study not a sensitizer	Same
Biocompatibility - SKIN IRRITATION - ISO 10993-10: 2010 (E)	Under the conditions of the study not an irritant	Under the conditions of the study not an irritant	Same
Biocompatibility - <i>IN VITRO</i> CYTOTOXICITY - ISO 10993-5: 2009(E)	Under the conditions of the cytotoxicity study, mild Systemic cytotoxicity (Grade 2) observed. It complied with the criteria in ISO 10993-5:2009.	Exhibit cytotoxic reactivity at 100% extract concentration (Grade 4 with neat extract). Non-cytotoxic reactivity at 50%, 25%, 12.5% and 6.25% extract concentration.	Similar
Biocompatibility - ACUTE SYSTEMIC TOXICITY - ISO 10993- 11: 2017(E)	Under the conditions of the study no systemic toxicity	No systemic toxicity under the experimental conditions employed	Similar
Manufacturer(s)	JIANGSU HEALTH-PEACE MEDICAL TECHNOLOGY CO. LTD	Nathan Trading Co., Ltd	-----

There are no significant differences between the two products and are identical in terms of intended use, materials, design, manufacturing methods. Both devices meet the ASTM standard D6319-19.

7. NON-CLINICAL TESTING SUMMARY

PERFORMANCE DATA

Test Method	Purpose	Acceptance Criteria	Result
ASMT D6319-19 Standard Specification for Nitrile Examination Gloves for Medical Application - Physical Dimensions Test	To determine the width, length, and thickness of the gloves	Width: 80±10 mm (for small size) 95±10 mm (for medium size) 110 ± 10 mm (for large size) Length: (Minimum) 220 mm (for small size) 230 mm (for medium, large size) Thickness: (Minimum) <i>Finger</i> -0.05 mm (for all sizes) <i>Palm</i> -0.05 mm (for all sizes)	Passed
ASMT D6319-19 Standard Specification for Nitrile Examination Gloves for Medical Application - Physical Requirements Test	To determine the tensile strength and ultimate elongation before and after acceleration aging	Before Acceleration Aging: Tensile Strength (MPa): 14 (Minimum) Ultimate Elongation (%): 500 (Minimum) After Acceleration Aging: Tensile Strength (MPa): 14 (Minimum) Ultimate Elongation (%): 400 (Minimum)	Passed
ASTM D6319-19 (ASTM D5151-11) Standard Test Method for Detection of Holes in Medical Gloves	To determine the holes in the gloves	(Mean) AQL 2.5	Passed
ASMT D6319-19 (ASTM D6124-11) Standard Test Method for Residual Powder on Medical Gloves	To determine the residual powder in the gloves	≤ 2.0 mg/glove	0.1 mg/glove, Passed

BIO-COMPATIBILITY DATA

Test Method	Purpose	Acceptance Criteria	Result
ISO 10993-10 Biological evaluation of medical devices — Part 10: Tests for skin irritation and skin sensitization	To determine the potential of the material under test to produce skin irritation in rabbits	Under the condition of study not an irritant	Under the condition of study not an irritant

ISO 10993-10 Biological evaluation of medical devices — Part 10: Tests for skin irritation and skin sensitization	To determine the skin sensitization potential of the material both in terms of induction and elicitation in guinea pigs.	Under the conditions of the study not a sensitizer.	Under the conditions of the study not a sensitizer.
ISO 10993-5 Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity	To evaluate the in vitro cytotoxic potential of the test item (both inner and outer surface) Extracts in L-929 mouse fibroblasts cells using elution method	Under the conditions of study, no more than grade 2 cytotoxic reaction	Mild (Grade 2) cytotoxicity reaction observed.
ISO 10993-11:2017 Biological evaluation of medical devices— Part 11: Tests for acute systemic toxicity	The test item was evaluated for acute systemic toxicity in ICR Mice	Under the conditions of the study no systemic toxicity	Under the conditions of the study no systemic toxicity

8. CLINICAL TESTING SUMMARY

Not applicable - Clinical data is not needed for gloves or for most devices cleared by the 510(k) process.

9. CONCLUSION

The conclusions drawn from the non-clinical test demonstrate that the subject device in 510(k) submission, the Disposable Medical Nitrile Gloves is as safe, as effective, and performs as well as or better than the legally marketed predicate device K203191- LYDUS Nitrile Examination Gloves, Powder Free.